

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY013023\
 Data File : VY012408.D
 Acq On : 30 Jan 2023 19:18
 Operator : KP/MD
 Sample : 01333-04
 Misc : 5.32g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 MH-11

Quant Time: Jan 31 01:49:44 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 04:31:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	167354	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	279939	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	276648	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	135303	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	101385	75.058	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	150.120%
35) Dibromofluoromethane	7.716	113	82168	55.434	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	110.860%
50) Toluene-d8	10.179	98	332676	66.177	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	132.360%
62) 4-Bromofluorobenzene	12.477	95	114850	54.964	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	109.920%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY013023\
Data File : VY012408.D
Acq On : 30 Jan 2023 19:18
Operator : KP/MD
Sample : 01333-04
Misc : 5.32g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
MH-11

Quant Time: Jan 31 01:49:44 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 04:31:47 2023
Response via : Initial Calibration

