

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY013023\
 Data File : VY012409.D
 Acq On : 30 Jan 2023 19:41
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 31 01:49:52 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 04:31:47 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	82	0.00
2 T	Dichlorodifluoromethane	0.437	0.366	16.2	81	0.00
3 P	Chloromethane	0.403	0.336	16.6	75	0.00
4 C	Vinyl Chloride	0.471	0.426	9.6#	80	0.00
5 T	Bromomethane	0.364	0.357	1.9	88	0.00
6 T	Chloroethane	0.307	0.300	2.3	86	0.00
7 T	Trichlorofluoromethane	0.919	0.920	-0.1	90	0.00
8 T	Diethyl Ether	0.273	0.266	2.6	84	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.447	0.451	-0.9	91	0.00
10 T	Methyl Iodide	0.680	0.563	17.2	67	0.00
11 T	Tert butyl alcohol	0.057	0.049	14.0	87	0.00
12 CM	1,1-Dichloroethene	0.459	0.454	1.1#	85	0.00
13 T	Acrolein	0.023	0.010	56.5#	50#	0.00
14 T	Allyl chloride	0.568	0.535	5.8	81	0.00
15 T	Acrylonitrile	0.119	0.122	-2.5	86	0.00
16 T	Acetone	0.139	0.118	15.1	63	0.00
17 T	Carbon Disulfide	1.420	1.259	11.3	77	0.00
18 T	Methyl Acetate	0.312	0.312	0.0	85	0.00
19 T	Methyl tert-butyl Ether	1.280	1.338	-4.5	86	0.00
20 T	Methylene Chloride	0.762	0.540	29.1#	78	0.00
21 T	trans-1,2-Dichloroethene	0.512	0.512	0.0	85	0.00
22 T	Diisopropyl ether	1.132	1.127	0.4	84	0.00
23 T	Vinyl Acetate	0.705	0.708	-0.4	82	0.00
24 P	1,1-Dichloroethane	0.785	0.795	-1.3	86	0.00
25 T	2-Butanone	0.154	0.143	7.1	74	0.00
26 T	2,2-Dichloropropane	0.852	0.855	-0.4	86	0.00
27 T	cis-1,2-Dichloroethene	0.563	0.575	-2.1	86	0.00
28 T	Bromochloromethane	0.171	0.151	11.7	78	0.00
29 T	Tetrahydrofuran	0.084	0.085	-1.2	82	0.00
30 C	Chloroform	0.927	0.968	-4.4#	88	0.00
31 T	Cyclohexane	0.729	0.664	8.9	86	0.00
32 T	1,1,1-Trichloroethane	0.918	0.969	-5.6	90	0.00
33 S	1,2-Dichloroethane-d4	0.465	0.405	12.9	81	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	80	0.00
35 S	Dibromofluoromethane	0.265	0.245	7.5	82	0.00
36 T	1,1-Dichloropropene	0.450	0.456	-1.3	86	0.00
37 T	Ethyl Acetate	0.198	0.197	0.5	80	0.00
38 T	Carbon Tetrachloride	0.581	0.626	-7.7	91	0.00
39 T	Methylcyclohexane	0.550	0.567	-3.1	88	0.00
40 TM	Benzene	1.299	1.334	-2.7	86	0.00
41 T	Methacrylonitrile	0.117	0.113	3.4	75	-0.01
42 TM	1,2-Dichloroethane	0.413	0.447	-8.2	88	0.00
43 T	Isopropyl Acetate	0.369	0.381	-3.3	83	0.00
44 TM	Trichloroethene	0.388	0.409	-5.4	88	0.00
45 C	1,2-Dichloropropane	0.271	0.283	-4.4#	87	0.00
46 T	Dibromomethane	0.181	0.191	-5.5	86	0.00
47 T	Bromodichloromethane	0.457	0.496	-8.5	89	0.00
48 T	Methyl methacrylate	0.170	0.178	-4.7	81	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY013023\
 Data File : VY012409.D
 Acq On : 30 Jan 2023 19:41
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 31 01:49:52 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 04:31:47 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.003	-50.0#	85	0.00
50 S	Toluene-d8	1.044	0.887	15.0	80	0.00
51 T	4-Methyl-2-Pentanone	0.197	0.209	-6.1	84	0.00
52 CM	Toluene	0.865	0.921	-6.5#	88	0.00
53 T	t-1,3-Dichloropropene	0.478	0.514	-7.5	87	0.00
54 T	cis-1,3-Dichloropropene	0.514	0.540	-5.1	85	0.00
55 T	1,1,2-Trichloroethane	0.258	0.283	-9.7	89	0.00
56 T	Ethyl methacrylate	0.338	0.373	-10.4	87	0.00
57 T	1,3-Dichloropropane	0.421	0.454	-7.8	88	0.00
58 T	2-Chloroethyl Vinyl ether	0.112	0.110	1.8	82	0.00
59 T	2-Hexanone	0.150	0.153	-2.0	78	0.00
60 T	Dibromochloromethane	0.342	0.377	-10.2	88	0.00
61 T	1,2-Dibromoethane	0.249	0.269	-8.0	87	0.00
62 S	4-Bromofluorobenzene	0.373	0.352	5.6	85	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	84	0.00
64 T	Tetrachloroethene	0.456	0.506	-11.0	94	0.00
65 PM	Chlorobenzene	1.023	1.056	-3.2	89	0.00
66 T	1,1,1,2-Tetrachloroethane	0.386	0.411	-6.5	89	0.00
67 C	Ethyl Benzene	1.828	1.902	-4.0#	89	0.00
68 T	m/p-Xylenes	0.723	0.758	-4.8	89	0.00
69 T	o-Xylene	0.690	0.725	-5.1	89	0.00
70 T	Styrene	1.154	1.234	-6.9	90	0.00
71 P	Bromoform	0.247	0.271	-9.7	91	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	86	0.00
73 T	Isopropylbenzene	3.605	3.722	-3.2	92	0.00
74 T	N-amyl acetate	0.707	0.692	2.1	84	0.00
75 P	1,1,2,2-Tetrachloroethane	0.568	0.568	0.0	87	0.00
76 T	1,2,3-Trichloropropane	0.460	0.482	-4.8	82	0.00
77 T	Bromobenzene	0.868	0.878	-1.2	89	0.00
78 T	n-propylbenzene	4.219	4.262	-1.0	92	0.00
79 T	2-Chlorotoluene	2.385	2.390	-0.2	90	0.00
80 T	1,3,5-Trimethylbenzene	3.134	3.227	-3.0	92	0.00
81 T	trans-1,4-Dichloro-2-butene	0.223	0.226	-1.3	87	0.00
82 T	4-Chlorotoluene	2.506	2.533	-1.1	91	0.00
83 T	tert-Butylbenzene	2.732	2.806	-2.7	94	0.00
84 T	1,2,4-Trimethylbenzene	3.091	3.201	-3.6	92	0.00
85 T	sec-Butylbenzene	3.889	3.974	-2.2	93	0.00
86 T	p-Isopropyltoluene	3.398	3.543	-4.3	95	0.00
87 T	1,3-Dichlorobenzene	1.725	1.757	-1.9	91	0.00
88 T	1,4-Dichlorobenzene	1.721	1.724	-0.2	91	0.00
89 T	n-Butylbenzene	2.945	3.019	-2.5	93	0.00
90 T	Hexachloroethane	0.613	0.632	-3.1	94	0.00
91 T	1,2-Dichlorobenzene	1.540	1.567	-1.8	91	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.117	0.122	-4.3	91	0.00
93 T	1,2,4-Trichlorobenzene	0.987	1.000	-1.3	93	0.00
94 T	Hexachlorobutadiene	0.594	0.605	-1.9	96	0.00
95 T	Naphthalene	1.947	2.054	-5.5	92	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY013023\
Data File : VY012409.D
Acq On : 30 Jan 2023 19:41
Operator : KP/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Jan 31 01:49:52 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 04:31:47 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.849	0.877	-3.3	95	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6