



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY013023\
Data File : VY012386.D
Acq On : 30 Jan 2023 09:09
Operator : KP/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
Title : SW846 8260
Last Update : Tue Jan 17 04:31:47 2023



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.4	13045	PASS
75	95	30	60	54.2	40640	PASS
95	95	100	100	100.0	74960	PASS
96	95	5	9	6.5	4888	PASS
173	174	0.00	2	1.2	841	PASS
174	95	50	100	94.5	70856	PASS
175	174	5	9	7.5	5319	PASS
176	174	95	101	95.8	67888	PASS
177	176	5	9	6.8	4634	PASS