

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020322\
 Data File : VY007367.D
 Acq On : 03 Feb 2022 10:35
 Operator : SY/MD
 Sample : VY0203SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0203SBL01

Quant Time: Feb 04 08:38:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020222S.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 03 04:36:48 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	105956	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	201698	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	186702	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	78460	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	62119	50.969	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	101.940%
35) Dibromofluoromethane	7.722	113	64729	49.977	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.960%
50) Toluene-d8	10.179	98	246937	51.606	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.220%
62) 4-Bromofluorobenzene	12.483	95	83533	51.361	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	102.720%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020322\
Data File : VY007367.D
Acq On : 03 Feb 2022 10:35
Operator : SY/MD
Sample : VY0203SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0203SBL01

Quant Time: Feb 04 08:38:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020222S.M
Quant Title : SW846 8260
QLast Update : Thu Feb 03 04:36:48 2022
Response via : Initial Calibration

