

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020322\
 Data File : VY007370.D
 Acq On : 03 Feb 2022 11:48
 Operator : SY/MD
 Sample : N1381-02
 Misc : 6.26g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 DB-05(33-35)-02-01-22

Quant Time: Feb 04 08:39:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020222S.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 03 04:36:48 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	109791	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	203816	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	191036	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	81096	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	68597	54.319	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 108.640%	
35) Dibromofluoromethane	7.722	113	67430	51.521	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 103.040%	
50) Toluene-d8	10.179	98	250738	51.856	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 103.720%	
62) 4-Bromofluorobenzene	12.483	95	88304	53.730	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 107.460%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020322\
Data File : VY007370.D
Acq On : 03 Feb 2022 11:48
Operator : SY/MD
Sample : N1381-02
Misc : 6.26g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
DB-05(33-35)-02-01-22

Quant Time: Feb 04 08:39:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020222S.M
Quant Title : SW846 8260
QLast Update : Thu Feb 03 04:36:48 2022
Response via : Initial Calibration

