

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY021122\
 Data File : VY007481.D
 Acq On : 11 Feb 2022 16:56
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VIBLK

Quant Time: Feb 12 02:54:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020222S.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 03 04:36:48 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	115956	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	207483	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	197829	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	87159	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	67298	50.457	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 100.920%	
35) Dibromofluoromethane	7.716	113	66308	49.769	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 99.540%	
50) Toluene-d8	10.179	98	252386	51.274	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 102.540%	
62) 4-Bromofluorobenzene	12.483	95	92640	55.372	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 110.740%	
Target Compounds						
					Qvalue	
39) Methylcyclohexane	9.179	83	3570	1.220	ug/l	88
67) Ethyl Benzene	11.593	91	57838	6.926	ug/l	98
78) n-propylbenzene	12.672	91	25952	2.690	ug/l	98
80) 1,3,5-Trimethylbenzene	12.812	105	12182	1.875	ug/l	97
95) Naphthalene	15.239	128	60552	17.182	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY021122\
Data File : VY007481.D
Acq On : 11 Feb 2022 16:56
Operator : SY/MD
Sample : VIBLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Quant Time: Feb 12 02:54:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020222S.M
Quant Title : SW846 8260
QLast Update : Thu Feb 03 04:36:48 2022
Response via : Initial Calibration

