

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY021122\
 Data File : VY007485.D
 Acq On : 11 Feb 2022 18:30
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VIBLK

Quant Time: Feb 12 02:56:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020222S.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 03 04:36:48 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	109627	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	199261	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	196203	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	88318	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	68916	54.653	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 109.300%	
35) Dibromofluoromethane	7.722	113	65989	51.573	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 103.140%	
50) Toluene-d8	10.179	98	245662	51.967	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 103.940%	
62) 4-Bromofluorobenzene	12.483	95	90736	56.472	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 112.940%	
Target Compounds						
95) Naphthalene	15.239	128	9067	2.539	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY021122\
Data File : VY007485.D
Acq On : 11 Feb 2022 18:30
Operator : SY/MD
Sample : VIBLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Quant Time: Feb 12 02:56:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020222S.M
Quant Title : SW846 8260
QLast Update : Thu Feb 03 04:36:48 2022
Response via : Initial Calibration

