

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041223\
 Data File : VY013274.D
 Acq On : 12 Apr 2023 18:09
 Operator : KP/MD
 Sample : 02292-09
 Misc : 7.57g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 IDW-SO-G3-041123

Quant Time: Apr 13 06:18:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041223S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 13 05:55:16 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	164769	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	281374	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	261404	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	123505	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	123619	62.429	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	124.860%
35) Dibromofluoromethane	7.710	113	100523	55.184	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	110.360%
50) Toluene-d8	10.173	98	347994	51.324	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	102.640%
62) 4-Bromofluorobenzene	12.477	95	107754	48.912	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	97.820%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041223\
Data File : VY013274.D
Acq On : 12 Apr 2023 18:09
Operator : KP/MD
Sample : 02292-09
Misc : 7.57g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
IDW-SO-G3-041123

Quant Time: Apr 13 06:18:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041223S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 13 05:55:16 2023
Response via : Initial Calibration

