

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041522\
 Data File : VY008348.D
 Acq On : 15 Apr 2022 14:05
 Operator : KP/MD
 Sample : N2401-11
 Misc : 4.81g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 DISP-1-6

Quant Time: Apr 16 04:58:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y040822S.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 08 16:02:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	26130	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	41534	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	34072	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	12395	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	11414	46.503	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	93.000%
35) Dibromofluoromethane	7.716	113	12334	51.415	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.820%
50) Toluene-d8	10.179	98	47801	58.232	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	116.460%
62) 4-Bromofluorobenzene	12.483	95	12945	41.927	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	83.860%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041522\
Data File : VY008348.D
Acq On : 15 Apr 2022 14:05
Operator : KP/MD
Sample : N2401-11
Misc : 4.81g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
DISP-1-6

Quant Time: Apr 16 04:58:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y040822S.M
Quant Title : SW846 8260
QLast Update : Fri Apr 08 16:02:50 2022
Response via : Initial Calibration

