

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041522\
 Data File : VY008351.D
 Acq On : 15 Apr 2022 15:15
 Operator : KP/MD
 Sample : N2390-01
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CL-1-041422

Quant Time: Apr 16 04:58:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y040822S.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 08 16:02:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	564	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	1159	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	814	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	146	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	504	95.134	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	190.260%#
35) Dibromofluoromethane	7.716	113	260	38.840	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	77.680%
50) Toluene-d8	10.185	98	1198	52.300	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.600%
62) 4-Bromofluorobenzene	12.489	95	113	13.116	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	26.240%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041522\
Data File : VY008351.D
Acq On : 15 Apr 2022 15:15
Operator : KP/MD
Sample : N2390-01
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
CL-1-041422

Quant Time: Apr 16 04:58:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y040822S.M
Quant Title : SW846 8260
QLast Update : Fri Apr 08 16:02:50 2022
Response via : Initial Calibration

