

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041522\
 Data File : VY008352.D
 Acq On : 15 Apr 2022 15:39
 Operator : KP/MD
 Sample : N2424-01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OR-02-041422

Quant Time: Apr 16 04:58:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y040822S.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 08 16:02:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	58787	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	107266	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	106559	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	47290	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	36325	65.782	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	131.560%
35) Dibromofluoromethane	7.722	113	35794	57.774	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	115.540%
50) Toluene-d8	10.178	98	129880	61.265	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	122.520%
62) 4-Bromofluorobenzene	12.483	95	45098	56.558	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	113.120%
Target Compounds						
16) Acetone	3.942	43	2408	18.130	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041522\
Data File : VY008352.D
Acq On : 15 Apr 2022 15:39
Operator : KP/MD
Sample : N2424-01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OR-02-041422

Quant Time: Apr 16 04:58:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y040822S.M
Quant Title : SW846 8260
QLast Update : Fri Apr 08 16:02:50 2022
Response via : Initial Calibration

