

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041522\
 Data File : VY008353.D
 Acq On : 15 Apr 2022 16:02
 Operator : KP/MD
 Sample : N2432-01
 Misc : 7.45g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-6075

Quant Time: Apr 16 04:59:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y040822S.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 08 16:02:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	96539	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	175515	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	169150	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	78004	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	56683	62.508	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 125.020%	
35) Dibromofluoromethane	7.722	113	55794	55.038	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 110.080%	
50) Toluene-d8	10.179	98	212482	61.255	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 122.500%	
62) 4-Bromofluorobenzene	12.483	95	78588	60.234	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 120.460%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041522\
Data File : VY008353.D
Acq On : 15 Apr 2022 16:02
Operator : KP/MD
Sample : N2432-01
Misc : 7.45g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-6075

Quant Time: Apr 16 04:59:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y040822S.M
Quant Title : SW846 8260
QLast Update : Fri Apr 08 16:02:50 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.789 min Scan# 91
Delta R.T. -0.000 min
Lab File: VY008353.D
Acq: 15 Apr 2022 16:02




Instrument :
MSVOA_Y
ClientSampleId :
SB-6075


Tgt Ion:168 Resp: 96539
Ion Ratio Lower Upper
168 100
99 55.9 46.9 70.3



#33
1,2-Dichloroethane-d4
Concen: 62.508 ug/l
RT: 8.142 min Scan# 1037
Delta R.T. 0.000 min
Lab File: VY008353.D
Acq: 15 Apr 2022 16:02

Tgt Ion: 65 Resp: 56683
Ion Ratio Lower Upper
65 100
67 54.7 0.0 99.4





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.691 min Scan# 1127

Delta R.T. 0.000 min

Lab File: VY008353.D

Acq: 15 Apr 2022 16:02

Instrument :

MSVOA_Y

ClientSampleId :

SB-6075





Tgt Ion:114 Resp: 175515

Ion Ratio Lower Upper

114 100

63 17.6 0.0 43.2

88 14.4 0.0 31.8



#35

Dibromofluoromethane

Concen: 55.038 ug/l

RT: 7.722 min Scan# 968

Delta R.T. 0.007 min

Lab File: VY008353.D

Acq: 15 Apr 2022 16:02




Tgt Ion:113 Resp: 55794


Ion Ratio Lower Upper

113 100

111 102.9 82.6 124.0

192 18.2 17.6 26.4









#50
Toluene-d8
Concen: 61.255 ug/l
RT: 10.179 min Scan# 11
Delta R.T. 0.001 min
Lab File: VY008353.D
Acq: 15 Apr 2022 16:02




Instrument :
MSVOA_Y
ClientSampleId :
SB-6075


Tgt Ion: 98 Resp: 212482
Ion Ratio Lower Upper
98 100
100 65.9 50.2 75.2



#62
4-Bromofluorobenzene
Concen: 60.234 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VY008353.D
Acq: 15 Apr 2022 16:02

Tgt Ion: 95 Resp: 78588
Ion Ratio Lower Upper
95 100
174 87.1 0.0 177.0
176 87.5 0.0 175.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.489 min Scan# 11
Delta R.T. 0.000 min
Lab File: VY008353.D
Acq: 15 Apr 2022 16:02

Instrument :
MSVOA_Y
ClientSampleId :
SB-6075

Tgt Ion:117 Resp: 169150
Ion Ratio Lower Upper
117 100
82 53.5 43.0 64.4
119 32.7 25.4 38.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.428 min Scan# 1904
Delta R.T. 0.000 min
Lab File: VY008353.D
Acq: 15 Apr 2022 16:02

Tgt Ion:152 Resp: 78004
Ion Ratio Lower Upper
152 100
115 58.0 28.2 84.7
150 156.7 0.0 347.6

