



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041522\
 Data File : VY008341.D
 Acq On : 15 Apr 2022 08:49
 Operator : KP/MD
 Sample : BFB
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y040822S.M
 Title : SW846 8260
 Last Update : Fri Apr 08 16:02:50 2022



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1741

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.2	6272	PASS
75	95	30	60	52.5	18109	PASS
95	95	100	100	100.0	34467	PASS
96	95	5	9	6.5	2240	PASS
173	174	0.00	2	1.1	306	PASS
174	95	50	100	81.4	28053	PASS
175	174	5	9	7.6	2136	PASS
176	174	95	101	98.1	27528	PASS
177	176	5	9	6.9	1896	PASS