

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041923\
 Data File : VY013345.D
 Acq On : 19 Apr 2023 15:08
 Operator : KP/MD
 Sample : 02371-09
 Misc : 5.08g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-20-8.0-8.5

Quant Time: Apr 20 06:56:19 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041223S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 13 05:55:16 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	12055	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	17452	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	17248	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	5970	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	9877	68.177	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	136.360%
35) Dibromofluoromethane	7.710	113	8136	72.011	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	144.020%
50) Toluene-d8	10.179	98	20717	49.263	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	98.520%
62) 4-Bromofluorobenzene	12.483	95	5998	43.896	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	87.800%
Target Compounds						Qvalue
16) Acetone	3.936	43	2331	46.796	ug/l	# 79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041923\
Data File : VY013345.D
Acq On : 19 Apr 2023 15:08
Operator : KP/MD
Sample : 02371-09
Misc : 5.08g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-20-8.0-8.5

Quant Time: Apr 20 06:56:19 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041223S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 13 05:55:16 2023
Response via : Initial Calibration

