

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041923\
 Data File : VY013337.D
 Acq On : 19 Apr 2023 10:42
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 20 06:54:32 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041223S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 13 05:55:16 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	108	0.00
2 T	Dichlorodifluoromethane	0.450	0.388	13.8	96	0.00
3 P	Chloromethane	0.564	0.484	14.2	97	0.00
4 C	Vinyl Chloride	0.535	0.519	3.0#	109	0.00
5 T	Bromomethane	0.410	0.322	21.5	91	0.00
6 T	Chloroethane	0.360	0.295	18.1	94	0.00
7 T	Trichlorofluoromethane	0.832	0.872	-4.8	119	0.00
8 T	Diethyl Ether	0.302	0.319	-5.6	119	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.489	0.520	-6.3	119	0.00
10 T	Methyl Iodide	0.600	0.692	-15.3	120	0.00
11 T	Tert butyl alcohol	0.096	0.088	8.3	128	0.01
12 CM	1,1-Dichloroethene	0.470	0.503	-7.0#	118	0.00
13 T	Acrolein	0.074	0.107	-44.6#	158#	0.00
14 T	Allyl chloride	0.899	0.889	1.1	109	0.00
15 T	Acrylonitrile	0.174	0.194	-11.5	118	0.00
16 T	Acetone	0.207	0.246	-18.8	125	0.00
17 T	Carbon Disulfide	1.549	1.450	6.4	104	0.00
18 T	Methyl Acetate	0.668	0.764	-14.4	121	0.01
19 T	Methyl tert-butyl Ether	1.409	1.508	-7.0	114	0.00
20 T	Methylene Chloride	0.641	0.633	1.2	129	0.00
21 T	trans-1,2-Dichloroethene	0.537	0.534	0.6	109	0.00
22 T	Diisopropyl ether	1.771	1.605	9.4	97	0.00
23 T	Vinyl Acetate	1.120	1.069	4.6	99	0.01
24 P	1,1-Dichloroethane	0.989	0.994	-0.5	112	0.00
25 T	2-Butanone	0.265	0.292	-10.2	115	0.00
26 T	2,2-Dichloropropane	0.888	0.867	2.4	111	0.00
27 T	cis-1,2-Dichloroethene	0.584	0.608	-4.1	113	0.01
28 T	Bromochloromethane	0.394	0.352	10.7	97	0.00
29 T	Tetrahydrofuran	0.161	0.168	-4.3	105	0.00
30 C	Chloroform	1.001	1.016	-1.5#	115	0.00
31 T	Cyclohexane	0.909	0.804	11.6	101	0.00
32 T	1,1,1-Trichloroethane	0.906	0.934	-3.1	117	0.00
33 S	1,2-Dichloroethane-d4	0.601	0.559	7.0	102	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	99	0.00
35 S	Dibromofluoromethane	0.324	0.360	-11.1	113	0.00
36 T	1,1-Dichloropropene	0.490	0.516	-5.3	108	0.00
37 T	Ethyl Acetate	0.395	0.384	2.8	94	0.00
38 T	Carbon Tetrachloride	0.542	0.597	-10.1	115	0.00
39 T	Methylcyclohexane	0.570	0.598	-4.9	104	0.00
40 TM	Benzene	1.419	1.466	-3.3	106	0.00
41 T	Methacrylonitrile	0.164	0.213	-29.9#	129	0.00
42 TM	1,2-Dichloroethane	0.481	0.495	-2.9	105	0.00
43 T	Isopropyl Acetate	0.657	0.679	-3.3	101	0.00
44 TM	Trichloroethene	0.366	0.410	-12.0	117	0.00
45 C	1,2-Dichloropropane	0.368	0.374	-1.6#	104	0.01
46 T	Dibromomethane	0.222	0.238	-7.2	108	0.00
47 T	Bromodichloromethane	0.507	0.545	-7.5	112	0.00
48 T	Methyl methacrylate	0.302	0.311	-3.0	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041923\
 Data File : VY013337.D
 Acq On : 19 Apr 2023 10:42
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 20 06:54:32 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041223S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 13 05:55:16 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.003	0.004	-33.3#	129	0.00
50 S	Toluene-d8	1.205	1.213	-0.7	99	0.00
51 T	4-Methyl-2-Pentanone	0.357	0.386	-8.1	103	0.00
52 CM	Toluene	0.857	0.907	-5.8#	108	0.00
53 T	t-1,3-Dichloropropene	0.531	0.574	-8.1	108	0.00
54 T	cis-1,3-Dichloropropene	0.582	0.623	-7.0	108	0.00
55 T	1,1,2-Trichloroethane	0.292	0.326	-11.6	115	0.00
56 T	Ethyl methacrylate	0.400	0.463	-15.8	109	0.00
57 T	1,3-Dichloropropane	0.497	0.535	-7.6	108	0.00
58 T	2-Chloroethyl Vinyl ether	0.167	0.169	-1.2	102	0.00
59 T	2-Hexanone	0.260	0.296	-13.8	106	0.00
60 T	Dibromochloromethane	0.367	0.414	-12.8	115	0.00
61 T	1,2-Dibromoethane	0.285	0.324	-13.7	113	0.00
62 S	4-Bromofluorobenzene	0.391	0.415	-6.1	104	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	98	0.00
64 T	Tetrachloroethene	0.353	0.405	-14.7	117	0.00
65 PM	Chlorobenzene	1.046	1.119	-7.0	110	0.00
66 T	1,1,1,2-Tetrachloroethane	0.393	0.446	-13.5	114	0.00
67 C	Ethyl Benzene	1.822	1.967	-8.0#	107	0.00
68 T	m/p-Xylenes	0.697	0.777	-11.5	111	0.00
69 T	o-Xylene	0.646	0.726	-12.4	111	0.00
70 T	Styrene	1.090	1.231	-12.9	109	0.00
71 P	Bromoform	0.272	0.325	-19.5	117	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
73 T	Isopropylbenzene	3.430	3.748	-9.3	111	0.00
74 T	N-amyl acetate	1.224	1.236	-1.0	100	0.00
75 P	1,1,2,2-Tetrachloroethane	0.847	0.909	-7.3	110	0.00
76 T	1,2,3-Trichloropropane	0.623	0.582	6.6	96	0.00
77 T	Bromobenzene	0.858	0.953	-11.1	114	0.00
78 T	n-propylbenzene	4.236	4.519	-6.7	108	0.00
79 T	2-Chlorotoluene	2.384	2.531	-6.2	108	0.00
80 T	1,3,5-Trimethylbenzene	2.873	3.166	-10.2	111	0.00
81 T	trans-1,4-Dichloro-2-butene	0.300	0.338	-12.7	112	0.00
82 T	4-Chlorotoluene	2.514	2.621	-4.3	107	0.00
83 T	tert-Butylbenzene	2.465	2.716	-10.2	111	0.00
84 T	1,2,4-Trimethylbenzene	2.796	3.123	-11.7	112	0.00
85 T	sec-Butylbenzene	3.686	4.023	-9.1	111	0.00
86 T	p-Isopropyltoluene	3.025	3.365	-11.2	112	0.00
87 T	1,3-Dichlorobenzene	1.653	1.818	-10.0	115	0.00
88 T	1,4-Dichlorobenzene	1.671	1.802	-7.8	113	0.00
89 T	n-Butylbenzene	2.893	3.116	-7.7	109	0.00
90 T	Hexachloroethane	0.631	0.667	-5.7	113	0.00
91 T	1,2-Dichlorobenzene	1.504	1.641	-9.1	114	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.154	0.179	-16.2	116	0.00
93 T	1,2,4-Trichlorobenzene	0.894	1.077	-20.5	122	0.00
94 T	Hexachlorobutadiene	0.529	0.634	-19.8	126	0.00
95 T	Naphthalene	1.948	2.532	-30.0#	125	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041923\
Data File : VY013337.D
Acq On : 19 Apr 2023 10:42
Operator : KP/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Apr 20 06:54:32 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041223S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 13 05:55:16 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.815	1.005	-23.3	126	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6