

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042622\
 Data File : VY008504.D
 Acq On : 26 Apr 2022 15:17
 Operator : KP/MD
 Sample : N2494-01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 MH-1S

Quant Time: Apr 27 01:22:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042122S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 21 14:28:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	6133	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	10317	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	9522	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	3757	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	3717	51.509	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	103.020%
35) Dibromofluoromethane	7.728	113	3369	50.598	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.200%
50) Toluene-d8	10.178	98	11890	45.034	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	90.060%
62) 4-Bromofluorobenzene	12.483	95	3685	41.100	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	82.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042622\
Data File : VY008504.D
Acq On : 26 Apr 2022 15:17
Operator : KP/MD
Sample : N2494-01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
MH-1S

Quant Time: Apr 27 01:22:52 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042122S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 21 14:28:40 2022
Response via : Initial Calibration

