

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061622\
 Data File : VY009192.D
 Acq On : 16 Jun 2022 10:47
 Operator : KP/MD
 Sample : VY0616SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0616SBL01

Quant Time: Jun 16 15:54:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061522S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 16 02:54:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	199748	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	344977	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	319998	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	140703	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	114840	55.038	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	110.080%
35) Dibromofluoromethane	7.710	113	104571	48.348	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.700%
50) Toluene-d8	10.179	98	408653	49.634	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	99.260%
62) 4-Bromofluorobenzene	12.477	95	138133	47.213	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	94.420%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061622\
Data File : VY009192.D
Acq On : 16 Jun 2022 10:47
Operator : KP/MD
Sample : VY0616SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0616SBL01

Quant Time: Jun 16 15:54:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061522S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 16 02:54:06 2022
Response via : Initial Calibration

