

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061622\
 Data File : VY009199.D
 Acq On : 16 Jun 2022 13:33
 Operator : KP/MD
 Sample : VIBLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VIBLK

Quant Time: Jun 17 02:26:13 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061522S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 16 02:54:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	257248	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	430397	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	400860	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	192343	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	143575	53.429	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	106.860%
35) Dibromofluoromethane	7.716	113	131884	48.874	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.740%
50) Toluene-d8	10.179	98	507449	49.401	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	98.800%
62) 4-Bromofluorobenzene	12.477	95	182956	50.122	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	100.240%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061622\
Data File : VY009199.D
Acq On : 16 Jun 2022 13:33
Operator : KP/MD
Sample : VIBLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Quant Time: Jun 17 02:26:13 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061522S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 16 02:54:06 2022
Response via : Initial Calibration

