

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061622\
 Data File : VY009200.D
 Acq On : 16 Jun 2022 14:17
 Operator : KP/MD
 Sample : N3347-01
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CL-02-061422

Quant Time: Jun 17 02:26:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061522S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 16 02:54:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	36992	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	62610	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	57962	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	25991	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	22005	56.946	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	113.900%
35) Dibromofluoromethane	7.716	113	20023	51.009	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.020%
50) Toluene-d8	10.173	98	72834	48.742	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	97.480%
62) 4-Bromofluorobenzene	12.483	95	25778	48.546	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	97.100%
Target Compounds						
						Qvalue
16) Acetone	3.948	43	1426	14.650	ug/l #	86
84) 1,2,4-Trimethylbenzene	13.123	105	3670	2.287	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061622\
Data File : VY009200.D
Acq On : 16 Jun 2022 14:17
Operator : KP/MD
Sample : N3347-01
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
CL-02-061422

Quant Time: Jun 17 02:26:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061522S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 16 02:54:06 2022
Response via : Initial Calibration

