

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061622\
 Data File : VY009201.D
 Acq On : 16 Jun 2022 14:41
 Operator : KP/MD
 Sample : N3348-01
 Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TR-5-061422

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 06/17/2022
 Supervised By :Mahesh Dadoda 06/17/2022

Quant Time: Jun 17 02:26:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061522S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 16 02:54:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	31032	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	52367	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	45864	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	21044	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	19387	59.807	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 119.620%			
35) Dibromofluoromethane	7.716	113	16630	50.651	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 101.300%			
50) Toluene-d8	10.179	98	60161	48.136	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery = 96.280%			
62) 4-Bromofluorobenzene	12.477	95	19919	44.850	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery = 89.700%			
Target Compounds						
16) Acetone	3.954	43	1619m	19.827	ug/l	
20) Methylene Chloride	4.710	84	3317	6.523	ug/l #	79
84) 1,2,4-Trimethylbenzene	13.117	105	1683	1.295	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061622\
Data File : VY009201.D
Acq On : 16 Jun 2022 14:41
Operator : KP/MD
Sample : N3348-01
Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TR-5-061422

Quant Time: Jun 17 02:26:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061522S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 16 02:54:06 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 06/17/2022
Supervised By :Mahesh Dadoda 06/17/2022

