

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061622\
 Data File : VY009202.D
 Acq On : 16 Jun 2022 15:04
 Operator : KP/MD
 Sample : N3353-01
 Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 PL-01-061422

Quant Time: Jun 17 02:26:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061522S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 16 02:54:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	54296	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	91315	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	84468	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	28321	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	33200	58.536	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	117.080%
35) Dibromofluoromethane	7.716	113	30038	52.467	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.940%
50) Toluene-d8	10.173	98	108294	49.691	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	99.380%
62) 4-Bromofluorobenzene	12.477	95	35787	46.210	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	92.420%
Target Compounds						
16) Acetone	3.930	43	1768	12.375	ug/l	Qvalue # 81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061622\
Data File : VY009202.D
Acq On : 16 Jun 2022 15:04
Operator : KP/MD
Sample : N3353-01
Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PL-01-061422

Quant Time: Jun 17 02:26:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061522S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 16 02:54:06 2022
Response via : Initial Calibration

