

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061622\
 Data File : VY009208.D
 Acq On : 16 Jun 2022 17:39
 Operator : KP/MD
 Sample : N3340-03
 Misc : 4.99g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CS-3

Quant Time: Jun 17 02:27:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061522S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 16 02:54:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	236467	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	411717	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	384198	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	181900	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	137605	55.707	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	111.420%	
35) Dibromofluoromethane	7.710	113	128927	49.946	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	99.900%	
50) Toluene-d8	10.173	98	485231	49.382	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	=	98.760%	
62) 4-Bromofluorobenzene	12.477	95	175092	50.144	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	=	100.280%	
Target Compounds						
11) Tert butyl alcohol	4.917	59	9156	23.364	ug/l	Qvalue # 79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061622\
Data File : VY009208.D
Acq On : 16 Jun 2022 17:39
Operator : KP/MD
Sample : N3340-03
Misc : 4.99g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
CS-3

Quant Time: Jun 17 02:27:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061522S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 16 02:54:06 2022
Response via : Initial Calibration

