

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061622\
 Data File : VY009191.D
 Acq On : 16 Jun 2022 09:53
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 16 15:40:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061522S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 16 02:54:06 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	108	0.00
2 T	Dichlorodifluoromethane	0.378	0.309	18.3	98	0.00
3 P	Chloromethane	0.469	0.419	10.7	102	0.00
4 C	Vinyl Chloride	0.492	0.465	5.5#	104	0.00
5 T	Bromomethane	0.346	0.312	9.8	103	0.00
6 T	Chloroethane	0.308	0.298	3.2	104	0.00
7 T	Trichlorofluoromethane	0.786	0.814	-3.6	118	0.00
8 T	Diethyl Ether	0.298	0.285	4.4	104	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.483	0.458	5.2	105	0.00
10 T	Methyl Iodide	0.618	0.591	4.4	99	0.00
11 T	Tert butyl alcohol	0.083	0.077	7.2	145	-0.01
12 CM	1,1-Dichloroethene	0.476	0.440	7.6#	102	0.00
13 T	Acrolein	0.017	0.012	29.4#	102	0.00
14 T	Allyl chloride	0.859	0.838	2.4	104	0.00
15 T	Acrylonitrile	0.160	0.157	1.9	105	0.00
16 T	Acetone	0.132	0.150	-13.6	135	0.00
17 T	Carbon Disulfide	1.377	1.308	5.0	102	0.00
18 T	Methyl Acetate	0.437	0.367	16.0	107	0.00
19 T	Methyl tert-butyl Ether	1.436	1.431	0.3	107	0.00
20 T	Methylene Chloride	0.670	0.551	17.8	102	0.00
21 T	trans-1,2-Dichloroethene	0.557	0.554	0.5	103	0.00
22 T	Diisopropyl ether	1.770	1.794	-1.4	107	0.00
23 T	Vinyl Acetate	1.114	1.177	-5.7	106	0.00
24 P	1,1-Dichloroethane	1.000	0.990	1.0	106	0.00
25 T	2-Butanone	0.208	0.221	-6.3	114	0.00
26 T	2,2-Dichloropropane	0.886	0.882	0.5	109	0.00
27 T	cis-1,2-Dichloroethene	0.654	0.650	0.6	104	0.00
28 T	Bromochloromethane	0.429	0.457	-6.5	111	0.00
29 T	Tetrahydrofuran	0.138	0.137	0.7	104	0.00
30 C	Chloroform	1.029	1.007	2.1#	104	0.00
31 T	Cyclohexane	0.936	0.910	2.8	105	0.00
32 T	1,1,1-Trichloroethane	0.905	0.873	3.5	103	0.00
33 S	1,2-Dichloroethane-d4	0.522	0.536	-2.7	113	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	109	0.00
35 S	Dibromofluoromethane	0.313	0.330	-5.4	112	0.00
36 T	1,1-Dichloropropene	0.482	0.472	2.1	101	0.00
37 T	Ethyl Acetate	0.287	0.288	-0.3	107	0.00
38 T	Carbon Tetrachloride	0.499	0.487	2.4	103	0.00
39 T	Methylcyclohexane	0.581	0.599	-3.1	105	0.00
40 TM	Benzene	1.408	1.412	-0.3	105	0.00
41 T	Methacrylonitrile	0.160	0.167	-4.4	98	0.00
42 TM	1,2-Dichloroethane	0.411	0.398	3.2	104	0.00
43 T	Isopropyl Acetate	0.531	0.528	0.6	106	0.00
44 TM	Trichloroethene	0.403	0.390	3.2	103	0.00
45 C	1,2-Dichloropropane	0.357	0.357	0.0	104	0.00
46 T	Dibromomethane	0.205	0.204	0.5	104	0.00
47 T	Bromodichloromethane	0.489	0.486	0.6	105	0.00
48 T	Methyl methacrylate	0.237	0.240	-1.3	104	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061622\
 Data File : VY009191.D
 Acq On : 16 Jun 2022 09:53
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 16 15:40:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061522S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 16 02:54:06 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.003	0.003	0.0	103	0.00
50 S	Toluene-d8	1.193	1.243	-4.2	111	0.00
51 T	4-Methyl-2-Pentanone	0.279	0.276	1.1	106	0.00
52 CM	Toluene	0.899	0.893	0.7#	105	0.00
53 T	t-1,3-Dichloropropene	0.513	0.519	-1.2	106	0.00
54 T	cis-1,3-Dichloropropene	0.580	0.591	-1.9	106	0.00
55 T	1,1,2-Trichloroethane	0.291	0.291	0.0	103	0.00
56 T	Ethyl methacrylate	0.394	0.399	-1.3	104	0.00
57 T	1,3-Dichloropropane	0.496	0.494	0.4	104	0.00
58 T	2-Chloroethyl Vinyl ether	0.171	0.173	-1.2	111	0.00
59 T	2-Hexanone	0.192	0.196	-2.1	110	0.00
60 T	Dibromochloromethane	0.356	0.347	2.5	102	0.00
61 T	1,2-Dibromoethane	0.283	0.282	0.4	102	0.00
62 S	4-Bromofluorobenzene	0.424	0.431	-1.7	111	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	110	0.00
64 T	Tetrachloroethene	0.463	0.453	2.2	103	0.00
65 PM	Chlorobenzene	1.083	1.055	2.6	104	0.00
66 T	1,1,1,2-Tetrachloroethane	0.397	0.392	1.3	105	0.00
67 C	Ethyl Benzene	1.890	1.837	2.8#	103	0.00
68 T	m/p-Xylenes	0.745	0.716	3.9	102	0.00
69 T	o-Xylene	0.706	0.693	1.8	104	0.00
70 T	Styrene	1.184	1.163	1.8	105	0.00
71 P	Bromoform	0.253	0.245	3.2	103	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	106	0.00
73 T	Isopropylbenzene	3.763	3.814	-1.4	102	0.00
74 T	N-amyl acetate	1.050	1.016	3.2	101	0.00
75 P	1,1,2,2-Tetrachloroethane	0.745	0.740	0.7	103	0.00
76 T	1,2,3-Trichloropropane	0.552	0.585	-6.0	113	0.00
77 T	Bromobenzene	0.925	0.932	-0.8	103	0.00
78 T	n-propylbenzene	4.553	4.577	-0.5	101	0.00
79 T	2-Chlorotoluene	2.565	2.595	-1.2	103	0.00
80 T	1,3,5-Trimethylbenzene	3.145	3.144	0.0	101	0.00
81 T	trans-1,4-Dichloro-2-butene	0.277	0.275	0.7	103	0.00
82 T	4-Chlorotoluene	2.677	2.654	0.9	102	0.00
83 T	tert-Butylbenzene	2.765	2.760	0.2	99	0.00
84 T	1,2,4-Trimethylbenzene	3.087	3.116	-0.9	101	0.00
85 T	sec-Butylbenzene	4.095	4.069	0.6	101	0.00
86 T	p-Isopropyltoluene	3.398	3.366	0.9	100	0.00
87 T	1,3-Dichlorobenzene	1.783	1.758	1.4	102	0.00
88 T	1,4-Dichlorobenzene	1.810	1.764	2.5	100	0.00
89 T	n-Butylbenzene	3.213	3.142	2.2	99	0.00
90 T	Hexachloroethane	0.641	0.633	1.2	100	0.00
91 T	1,2-Dichlorobenzene	1.613	1.602	0.7	100	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.134	0.126	6.0	94	0.00
93 T	1,2,4-Trichlorobenzene	1.014	0.990	2.4	96	0.00
94 T	Hexachlorobutadiene	0.557	0.509	8.6	92	0.00
95 T	Naphthalene	2.115	2.137	-1.0	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061622\
Data File : VY009191.D
Acq On : 16 Jun 2022 09:53
Operator : KP/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Jun 16 15:40:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061522S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 16 02:54:06 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.892	0.867	2.8	96	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 5