

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061622\
 Data File : VY009191.D
 Acq On : 16 Jun 2022 09:53
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 06/17/2022
 Supervised By :Mahesh Dadoda 06/17/2022

Quant Time: Jun 16 15:40:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061522S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 16 02:54:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	199640	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	324665	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	292618	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	136710	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	107070	51.342	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 102.680%			
35) Dibromofluoromethane	7.716	113	107042	52.587	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 105.180%			
50) Toluene-d8	10.179	98	403476	52.071	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery = 104.140%			
62) 4-Bromofluorobenzene	12.483	95	139990	50.841	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery = 101.680%			
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.900	85	61741	40.855	ug/l	99
3) Chloromethane	2.113	50	83568	44.634	ug/l	94
4) Vinyl Chloride	2.247	62	92924	47.282	ug/l	98
5) Bromomethane	2.643	94	62328	45.143	ug/l	91
6) Chloroethane	2.790	64	59494	48.377	ug/l	96
7) Trichlorofluoromethane	3.119	101	162604	51.795	ug/l	95
8) Diethyl Ether	3.527	74	56934	47.796	ug/l	87
9) 1,1,2-Trichlorotrifluo...	3.893	101	91375	47.337	ug/l	95
10) Methyl Iodide	4.082	142	118001	47.855	ug/l #	88
11) Tert butyl alcohol	4.936	59	77273	233.556	ug/l	98
12) 1,1-Dichloroethene	3.869	96	87898	46.269	ug/l	81
13) Acrolein	3.729	56	12159	207.637	ug/l	98
14) Allyl chloride	4.472	41	167371	48.822	ug/l	91
15) Acrylonitrile	5.155	53	157069	246.532	ug/l	97
16) Acetone	3.942	43	149933	285.416	ug/l	91
17) Carbon Disulfide	4.186	76	261107	47.477	ug/l	100
18) Methyl Acetate	4.472	43	73233	48.152	ug/l	94
19) Methyl tert-butyl Ether	5.210	73	285589	49.825	ug/l	97
20) Methylene Chloride	4.704	84	110012	46.314	ug/l	88
21) trans-1,2-Dichloroethene	5.210	96	110650	49.709	ug/l	88
22) Diisopropyl ether	6.112	45	358170	50.670	ug/l	94
23) Vinyl Acetate	6.051	43	1175196	264.285	ug/l #	95
24) 1,1-Dichloroethane	6.009	63	197736	49.506	ug/l	98
25) 2-Butanone	6.978	43	220768	265.583	ug/l	95
26) 2,2-Dichloropropane	6.972	77	176051	49.745	ug/l	95
27) cis-1,2-Dichloroethene	6.978	96	129822	49.728	ug/l	90
28) Bromochloromethane	7.332	49	91232	53.290	ug/l #	85
29) Tetrahydrofuran	7.344	42	136390	247.111	ug/l	90
30) Chloroform	7.502	83	201122	48.953	ug/l	99
31) Cyclohexane	7.777	56	181588	48.610	ug/l	96
32) 1,1,1-Trichloroethane	7.697	97	174346	48.262	ug/l	96
36) 1,1-Dichloropropene	7.911	75	153341	48.972	ug/l	97
37) Ethyl Acetate	7.069	43	93378	50.115	ug/l #	95
38) Carbon Tetrachloride	7.899	117	158114	48.835	ug/l	99
39) Methylcyclohexane	9.179	83	194423	51.504	ug/l	91
40) Benzene	8.155	78	458372	50.130	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061622\
 Data File : VY009191.D
 Acq On : 16 Jun 2022 09:53
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 06/17/2022
 Supervised By :Mahesh Dadoda 06/17/2022

Quant Time: Jun 16 15:40:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061522S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 16 02:54:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.307	41	54330	52.132	ug/l #	42
42) 1,2-Dichloroethane	8.234	62	129192	48.447	ug/l	91
43) Isopropyl Acetate	8.271	43	171340	49.717	ug/l #	92
44) Trichloroethene	8.935	130	126624	48.447	ug/l	97
45) 1,2-Dichloropropane	9.215	63	116021	50.058	ug/l	97
46) Dibromomethane	9.301	93	66349	49.766	ug/l	99
47) Bromodichloromethane	9.496	83	157946	49.706	ug/l	97
48) Methyl methacrylate	9.289	41	78081	50.760	ug/l #	86
49) 1,4-Dioxane	9.295	88	19086	1006.393	ug/l #	88
51) 4-Methyl-2-Pentanone	10.069	43	447844	247.041	ug/l	91
52) Toluene	10.240	92	289895	49.654	ug/l	95
53) t-1,3-Dichloropropene	10.465	75	168469	50.544	ug/l	99
54) cis-1,3-Dichloropropene	9.923	75	191907	50.979	ug/l #	87
55) 1,1,2-Trichloroethane	10.642	97	94334	50.000	ug/l	99
56) Ethyl methacrylate	10.508	69	129671	50.699	ug/l #	83
57) 1,3-Dichloropropane	10.788	76	160271	49.727	ug/l	97
58) 2-Chloroethyl Vinyl ether	9.782	63	280103	252.649	ug/l	95
59) 2-Hexanone	10.831	43	318322	255.959	ug/l	89
60) Dibromochloromethane	10.983	129	112651	48.769	ug/l	99
61) 1,2-Dibromoethane	11.087	107	91507	49.765	ug/l	100
64) Tetrachloroethene	10.715	164	132496	48.875	ug/l	99
65) Chlorobenzene	11.514	112	308768	48.711	ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.587	131	114642	49.384	ug/l	97
67) Ethyl Benzene	11.593	91	537449	48.589	ug/l	97
68) m/p-Xylenes	11.703	106	419114	96.109	ug/l	94
69) o-Xylene	12.026	106	202867	49.099	ug/l	92
70) Styrene	12.044	104	340213	49.084	ug/l	96
71) Bromoform	12.209	173	71700	48.356	ug/l #	97
73) Isopropylbenzene	12.331	105	521473	50.681	ug/l	99
74) N-amyl acetate	12.142	43	138862	48.381	ug/l #	89
75) 1,1,2,2-Tetrachloroethane	12.581	83	101129	49.659	ug/l	99
76) 1,2,3-Trichloropropane	12.629	75	80031m	53.018	ug/l	
77) Bromobenzene	12.605	156	127431	50.409	ug/l	93
78) n-propylbenzene	12.672	91	625672	50.256	ug/l	97
79) 2-Chlorotoluene	12.758	91	354793	50.580	ug/l	96
80) 1,3,5-Trimethylbenzene	12.812	105	429748	49.970	ug/l	98
81) trans-1,4-Dichloro-2-b...	12.380	75	37583	49.618	ug/l	86
82) 4-Chlorotoluene	12.855	91	362764	49.564	ug/l	97
83) tert-Butylbenzene	13.075	119	377264	49.895	ug/l	95
84) 1,2,4-Trimethylbenzene	13.117	105	425963	50.462	ug/l	98
85) sec-Butylbenzene	13.251	105	556263	49.683	ug/l	99
86) p-Isopropyltoluene	13.367	119	460201	49.531	ug/l	98
87) 1,3-Dichlorobenzene	13.367	146	240384	49.303	ug/l	99
88) 1,4-Dichlorobenzene	13.446	146	241090	48.712	ug/l	98
89) n-Butylbenzene	13.696	91	429563	48.892	ug/l	98
90) Hexachloroethane	13.959	117	86472	49.358	ug/l	89
91) 1,2-Dichlorobenzene	13.739	146	219011	49.674	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.355	75	17264	46.983	ug/l	77
93) 1,2,4-Trichlorobenzene	15.007	180	135328	48.819	ug/l	99
94) Hexachlorobutadiene	15.111	225	69592	45.676	ug/l	99
95) Naphthalene	15.239	128	292215	50.540	ug/l	98
96) 1,2,3-Trichlorobenzene	15.422	180	118594	48.609	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061622\
Data File : VY009191.D
Acq On : 16 Jun 2022 09:53
Operator : KP/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Krupa Patel 06/17/2022
Supervised By :Mahesh Dadoda 06/17/2022

Quant Time: Jun 16 15:40:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061522S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 16 02:54:06 2022
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.783 min Scan# 91
Delta R.T. -0.000 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion:168 Resp: 199640
Ion Ratio Lower Upper
168 100
99 53.6 46.9 70.3

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 06/17/2022
Supervised By :Mahesh Dadoda 06/17/2022




#2
Dichlorodifluoromethane
Concen: 40.855 ug/l
RT: 1.900 min Scan# 13
Delta R.T. -0.000 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53




Tgt Ion: 85 Resp: 61741
Ion Ratio Lower Upper
85 100
87 32.4 16.1 48.2





#3
Chloromethane
Concen: 44.634 ug/l
RT: 2.113 min Scan# 41
Delta R.T. 0.006 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 50 Resp: 8356
Ion Ratio Lower Upper
50 100
52 34.1 24.6 37.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/17/2022
Supervised By :Mahesh Dadoda 06/17/2022



#4
Vinyl Chloride
Concen: 47.282 ug/l
RT: 2.247 min Scan# 70
Delta R.T. 0.006 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53

Tgt Ion: 62 Resp: 92924
Ion Ratio Lower Upper
62 100
64 30.4 25.0 37.6









#5
Bromomethane
Concen: 45.143 ug/l
RT: 2.643 min Scan# 11
Delta R.T. 0.006 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 94 Resp: 62328
Ion Ratio Lower Upper
94 100
96 100.1 73.5 110.3




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/17/2022
Supervised By :Mahesh Dadoda 06/17/2022



#6
Chloroethane
Concen: 48.377 ug/l
RT: 2.790 min Scan# 159
Delta R.T. 0.006 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53


Tgt Ion: 64 Resp: 59494
Ion Ratio Lower Upper
64 100
66 34.0 25.4 38.0





#7
 Trichlorofluoromethane
 Concen: 51.795 ug/l
 RT: 3.119 min Scan# 21
 Delta R.T. 0.006 min
 Lab File: VY009191.D
 Acq: 16 Jun 2022 09:53




Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050




Tgt Ion:101 Resp: 162604
 Ion Ratio Lower Upper
 101 100
 103 64.1 48.0 72.0

Manual Integrations
 APPROVED



Reviewed By :Krupa Patel 06/17/2022
 Supervised By :Mahesh Dadoda 06/17/2022




#8
 Diethyl Ether
 Concen: 47.796 ug/l
 RT: 3.527 min Scan# 280
 Delta R.T. 0.006 min
 Lab File: VY009191.D
 Acq: 16 Jun 2022 09:53



Tgt Ion: 74 Resp: 56934
 Ion Ratio Lower Upper
 74 100
 45 96.5 55.1 165.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 47.337 ug/l
RT: 3.893 min Scan# 340
Delta R.T. 0.006 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion:101 Resp: 91375
Ion Ratio Lower Upper
101 100
85 45.2 40.6 60.8
151 79.0 64.9 97.3




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/17/2022
Supervised By :Mahesh Dadoda 06/17/2022



#10
Methyl Iodide
Concen: 47.855 ug/l
RT: 4.082 min Scan# 371
Delta R.T. 0.000 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53


Tgt Ion:142 Resp: 118001
Ion Ratio Lower Upper
142 100
127 41.3 41.5 62.3#
141 14.3 11.9 17.9





#11
Tert butyl alcohol
Concen: 233.556 ug/l
RT: 4.936 min Scan# 511
Delta R.T. -0.012 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 59 Resp: 7727
Ion Ratio Lower Upper
59 100
57 7.3 6.4 9.6

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 06/17/2022
Supervised By :Mahesh Dadoda 06/17/2022




#12
1,1-Dichloroethene
Concen: 46.269 ug/l
RT: 3.869 min Scan# 336
Delta R.T. 0.006 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53



Tgt Ion: 96 Resp: 87898
Ion Ratio Lower Upper
96 100
61 147.7 145.8 218.6
98 61.9 52.5 78.7





#13

Acrolein

Concen: 207.637 ug/l

RT: 3.729 min Scan# 312

Delta R.T. 0.006 min

Lab File: VY009191.D

Acq: 16 Jun 2022 09:53

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 56 Resp: 12159

Ion Ratio Lower Upper

56 100





55 70.4 57.6 86.4

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 06/17/2022

Supervised By :Mahesh Dadoda 06/17/2022



#14

Allyl chloride

Concen: 48.822 ug/l

RT: 4.472 min Scan# 435

Delta R.T. 0.006 min

Lab File: VY009191.D

Acq: 16 Jun 2022 09:53




Tgt Ion: 41 Resp: 167371


Ion Ratio Lower Upper

41 100

39 65.9 58.8 88.2


76 36.1 25.1 37.7





#15
Acrylonitrile
Concen: 246.532 ug/l
RT: 5.155 min Scan# 546
Delta R.T. 0.006 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 53 Resp: 157069
Ion Ratio Lower Upper
53 100
52 81.4 67.0 100.4
51 35.6 29.8 44.6




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/17/2022
Supervised By :Mahesh Dadoda 06/17/2022



#16
Acetone
Concen: 285.416 ug/l
RT: 3.942 min Scan# 348
Delta R.T. -0.000 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53


Tgt Ion: 43 Resp: 149933
Ion Ratio Lower Upper
43 100
58 32.3 22.0 33.0





#17
Carbon Disulfide
Concen: 47.477 ug/l
RT: 4.186 min Scan# 387
Delta R.T. 0.006 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 76 Resp: 26110
Ion Ratio Lower Upper
76 100
78 9.1 7.4 11.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/17/2022
Supervised By :Mahesh Dadoda 06/17/2022



#18
Methyl Acetate
Concen: 48.152 ug/l
RT: 4.472 min Scan# 435
Delta R.T. 0.006 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53


Tgt Ion: 43 Resp: 73233
Ion Ratio Lower Upper
43 100
74 23.5 16.6 24.8





#19
Methyl tert-butyl Ether
Concen: 49.825 ug/l
RT: 5.210 min Scan# 511
Delta R.T. -0.000 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 73 Resp: 285589
Ion Ratio Lower Upper
73 100
57 21.8 18.6 27.8

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 06/17/2022
Supervised By :Mahesh Dadoda 06/17/2022




#20
Methylene Chloride
Concen: 46.314 ug/l
RT: 4.704 min Scan# 473
Delta R.T. 0.000 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53




Tgt Ion: 84 Resp: 110012
Ion Ratio Lower Upper
84 100
49 123.7 114.4 171.6
51 38.0 35.3 52.9
86 63.0 53.7 80.5





#21
trans-1,2-Dichloroethene
Concen: 49.709 ug/l
RT: 5.210 min Scan# 511
Delta R.T. -0.000 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 96 Resp: 110650
Ion Ratio Lower Upper
96 100
61 131.8 120.5 180.7
98 65.2 49.8 74.8

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 06/17/2022
Supervised By :Mahesh Dadoda 06/17/2022




#22
Diisopropyl ether
Concen: 50.670 ug/l
RT: 6.112 min Scan# 704
Delta R.T. -0.000 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53




Tgt Ion: 45 Resp: 358170
Ion Ratio Lower Upper
45 100
43 57.0 49.3 73.9
87 28.0 19.2 28.8
59 11.2 8.4 12.6





#23
 Vinyl Acetate
 Concen: 264.285 ug/l
 RT: 6.051 min Scan# 694
 Delta R.T. -0.000 min
 Lab File: VY009191.D
 Acq: 16 Jun 2022 09:53




Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050




Tgt Ion: 43 Resp: 1175190
 Ion Ratio Lower Upper
 43 100
 86 10.8 7.1 10.7

Manual Integrations
 APPROVED



Reviewed By :Krupa Patel 06/17/2022
 Supervised By :Mahesh Dadoda 06/17/2022




#24
 1,1-Dichloroethane
 Concen: 49.506 ug/l
 RT: 6.009 min Scan# 687
 Delta R.T. -0.000 min
 Lab File: VY009191.D
 Acq: 16 Jun 2022 09:53




Tgt Ion: 63 Resp: 197736
 Ion Ratio Lower Upper
 63 100
 98 7.9 3.6 10.8
 100 4.8 2.0 5.8





#25
2-Butanone
Concen: 265.583 ug/l
RT: 6.978 min Scan# 846
Delta R.T. -0.000 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 43 Resp: 220768
Ion Ratio Lower Upper
43 100
72 24.7 17.7 26.5

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 06/17/2022
Supervised By :Mahesh Dadoda 06/17/2022




#26
2,2-Dichloropropane
Concen: 49.745 ug/l
RT: 6.972 min Scan# 845
Delta R.T. -0.000 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53




Tgt Ion: 77 Resp: 176051
Ion Ratio Lower Upper
77 100
97 23.1 10.4 31.4





#27
 cis-1,2-Dichloroethene
 Concen: 49.728 ug/l
 RT: 6.978 min Scan# 846
 Delta R.T. -0.000 min
 Lab File: VY009191.D
 Acq: 16 Jun 2022 09:53




Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050



Tgt Ion: 96 Resp: 12982
 Ion Ratio Lower Upper
 96 100
 61 141.5 0.0 319.8
 98 64.9 0.0 127.0




Manual Integrations
 APPROVED


Reviewed By :Krupa Patel 06/17/2022
 Supervised By :Mahesh Dadoda 06/17/2022



#28
 Bromochloromethane
 Concen: 53.290 ug/l
 RT: 7.332 min Scan# 904
 Delta R.T. 0.006 min
 Lab File: VY009191.D
 Acq: 16 Jun 2022 09:53

Tgt Ion: 49 Resp: 91232
 Ion Ratio Lower Upper
 49 100
 129 2.5 0.0 2.2#
 130 81.3 55.4 83.0





#29

Tetrahydrofuran

Concen: 247.111 ug/l

RT: 7.344 min Scan# 905

Delta R.T. 0.006 min

Lab File: VY009191.D


Acq: 16 Jun 2022 09:53

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050






Tgt Ion:	42	Resp:	136390
Ion Ratio	Lower	Upper	
42	100		
72	44.9	30.3	45.5
71	41.3	29.0	43.4

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 06/17/2022

Supervised By :Mahesh Dadoda 06/17/2022



#30

Chloroform

Concen: 48.953 ug/l




RT: 7.502 min Scan# 932


Delta R.T. -0.000 min

Lab File: VY009191.D

Acq: 16 Jun 2022 09:53


Tgt Ion:	83	Resp:	201122
Ion Ratio	Lower	Upper	
83	100		
85	64.1	51.8	77.6





#31
Cyclohexane
Concen: 48.610 ug/l
RT: 7.777 min Scan# 911
Delta R.T. -0.000 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 56 Resp: 181588
Ion Ratio Lower Upper
56 100
69 30.8 26.7 40.1
84 86.9 67.2 100.8




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/17/2022
Supervised By :Mahesh Dadoda 06/17/2022



#32
1,1,1-Trichloroethane
Concen: 48.262 ug/l
RT: 7.697 min Scan# 964
Delta R.T. -0.000 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53

Tgt Ion: 97 Resp: 174346
Ion Ratio Lower Upper
97 100
99 64.5 50.7 76.1
61 42.1 37.6 56.4





#33

1,2-Dichloroethane-d4

Concen: 51.342 ug/l

RT: 8.136 min Scan# 1036

Delta R.T. -0.000 min

Lab File: VY009191.D

Acq: 16 Jun 2022 09:53

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 65 Resp: 107070

Ion Ratio Lower Upper

65 100





67 56.0 0.0 99.4

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 06/17/2022

Supervised By :Mahesh Dadoda 06/17/2022



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.685 min Scan# 1126

Delta R.T. -0.000 min

Lab File: VY009191.D

Acq: 16 Jun 2022 09:53




Tgt Ion:114 Resp: 324665


Ion Ratio Lower Upper

114 100

63 19.2 0.0 43.2


88 15.2 0.0 31.8





#35
Dibromofluoromethane
Concen: 52.587 ug/l
RT: 7.716 min Scan# 911
Delta R.T. 0.006 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion:113 Resp: 107042
Ion Ratio Lower Upper
113 100
111 102.5 82.6 124.0
192 20.6 17.6 26.4




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/17/2022
Supervised By :Mahesh Dadoda 06/17/2022



#36
1,1-Dichloropropene
Concen: 48.972 ug/l
RT: 7.911 min Scan# 999
Delta R.T. -0.000 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53


Tgt Ion: 75 Resp: 153341
Ion Ratio Lower Upper
75 100
110 36.5 17.1 51.2
77 31.5 24.4 36.6





#37
Ethyl Acetate
Concen: 50.115 ug/l
RT: 7.069 min Scan# 860
Delta R.T. 0.006 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 43 Resp: 93378
Ion Ratio Lower Upper
43 100
61 14.1 9.7 14.5
70 10.9 7.1 10.7




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/17/2022
Supervised By :Mahesh Dadoda 06/17/2022



#38
Carbon Tetrachloride
Concen: 48.835 ug/l
RT: 7.899 min Scan# 997
Delta R.T. 0.006 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53


Tgt Ion:117 Resp: 158114
Ion Ratio Lower Upper
117 100
119 97.7 77.9 116.9
121 32.0 24.8 37.2





#39
Methylcyclohexane
Concen: 51.504 ug/l
RT: 9.179 min Scan# 11
Delta R.T. -0.000 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 83 Resp: 19442
Ion Ratio Lower Upper
83 100
55 73.2 66.5 99.7
98 50.3 37.9 56.9

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 06/17/2022
Supervised By :Mahesh Dadoda 06/17/2022




#40
Benzene
Concen: 50.130 ug/l
RT: 8.155 min Scan# 1039
Delta R.T. -0.000 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53




Tgt Ion: 78 Resp: 458372
Ion Ratio Lower Upper
78 100
77 24.4 20.4 30.6





#41
Methacrylonitrile
Concen: 52.132 ug/l
RT: 7.307 min Scan# 901
Delta R.T. -0.006 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 41 Resp: 54330
Ion Ratio Lower Upper
41 100
39 52.5 92.9 139.3
67 71.4 0.0 0.0
52 27.2 0.0 0.0

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 06/17/2022
Supervised By :Mahesh Dadoda 06/17/2022




#42
1,2-Dichloroethane
Concen: 48.447 ug/l
RT: 8.234 min Scan# 1052
Delta R.T. -0.000 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53




Tgt Ion: 62 Resp: 129192
Ion Ratio Lower Upper
62 100
98 10.4 0.0 14.8





#43
Isopropyl Acetate
Concen: 49.717 ug/l
RT: 8.271 min Scan# 1058
Delta R.T. 0.006 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 43 Resp: 171340
Ion Ratio Lower Upper
43 100
61 29.6 20.2 30.2
87 14.0 8.9 13.3




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/17/2022
Supervised By :Mahesh Dadoda 06/17/2022



#44
Trichloroethene
Concen: 48.447 ug/l
RT: 8.935 min Scan# 1167
Delta R.T. -0.000 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53

Tgt Ion:130 Resp: 126624
Ion Ratio Lower Upper
130 100
95 96.8 0.0 188.2





#45

1,2-Dichloropropane

Concen: 50.058 ug/l

RT: 9.215 min Scan# 11602

Delta R.T. 0.006 min

Lab File: VY009191.D

Acq: 16 Jun 2022 09:53

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 63 Resp: 11602

Ion Ratio Lower Upper

63 100





65 30.6 23.0 34.6

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 06/17/2022

Supervised By :Mahesh Dadoda 06/17/2022



#46

Dibromomethane

Concen: 49.766 ug/l

RT: 9.301 min Scan# 1227

Delta R.T. -0.000 min

Lab File: VY009191.D

Acq: 16 Jun 2022 09:53




Tgt Ion: 93 Resp: 66349


Ion Ratio Lower Upper

93 100

95 82.9 64.3 96.5


174 104.6 83.4 125.0





#47
Bromodichloromethane
Concen: 49.706 ug/l
RT: 9.496 min Scan# 1258
Delta R.T. 0.006 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 83 Resp: 157940
Ion Ratio Lower Upper
83 100
85 62.7 51.9 77.9
127 8.6 6.6 10.0

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 06/17/2022
Supervised By :Mahesh Dadoda 06/17/2022




#48
Methyl methacrylate
Concen: 50.760 ug/l
RT: 9.289 min Scan# 1225
Delta R.T. -0.000 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53



Tgt Ion: 41 Resp: 78081
Ion Ratio Lower Upper
41 100
69 88.2 58.2 87.2#
39 52.9 47.4 71.0





#49

1,4-Dioxane

Concen: 1006.393 ug/l

RT: 9.295 min Scan# 1226

Delta R.T. -0.000 min

Lab File: VY009191.D

Acq: 16 Jun 2022 09:53

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 88 Resp: 19080





Ion	Ratio	Lower	Upper
88	100		
43	28.0	32.8	49.2
58	68.7	59.8	89.6

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 06/17/2022

Supervised By :Mahesh Dadoda 06/17/2022



#50

Toluene-d8

Concen: 52.071 ug/l

RT: 10.179 min Scan# 1371




Delta R.T. 0.006 min


Lab File: VY009191.D

Acq: 16 Jun 2022 09:53

Tgt Ion: 98 Resp: 403476


Ion	Ratio	Lower	Upper
98	100		
100	65.3	50.2	75.2





#51
4-Methyl-2-Pentanone
Concen: 247.041 ug/l
RT: 10.069 min Scan# 11
Delta R.T. 0.006 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 43 Resp: 447844
Ion Ratio Lower Upper
43 100
58 40.2 28.1 42.1




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/17/2022
Supervised By :Mahesh Dadoda 06/17/2022



#52
Toluene
Concen: 49.654 ug/l
RT: 10.240 min Scan# 1381
Delta R.T. -0.000 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53


Tgt Ion: 92 Resp: 289895
Ion Ratio Lower Upper
92 100
91 171.5 142.6 213.8





#53
t-1,3-Dichloropropene
Concen: 50.544 ug/l
RT: 10.465 min Scan# 1418
Delta R.T. 0.006 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 75 Resp: 168469
Ion Ratio Lower Upper
75 100
77 31.5 25.4 38.2

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 06/17/2022
Supervised By :Mahesh Dadoda 06/17/2022




#54
cis-1,3-Dichloropropene
Concen: 50.979 ug/l
RT: 9.923 min Scan# 1329
Delta R.T. -0.000 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53




Tgt Ion: 75 Resp: 191907
Ion Ratio Lower Upper
75 100
77 30.8 25.6 38.4
39 44.9 47.6 71.4#





#55
1,1,2-Trichloroethane
Concen: 50.000 ug/l
RT: 10.642 min Scan# 1447
Delta R.T. -0.000 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 97 Resp: 94334
Ion Ratio Lower Upper
97 100
83 87.1 69.8 104.8
85 54.2 44.3 66.5
99 61.6 49.8 74.6

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 06/17/2022
Supervised By :Mahesh Dadoda 06/17/2022




#56
Ethyl methacrylate
Concen: 50.699 ug/l
RT: 10.508 min Scan# 1425
Delta R.T. 0.006 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53



Tgt Ion: 69 Resp: 129671
Ion Ratio Lower Upper
69 100
41 63.4 63.4 95.0
39 34.5 35.8 53.8#





#57

1,3-Dichloropropane

Concen: 49.727 ug/l

RT: 10.788 min Scan# 1471

Delta R.T. -0.000 min

Lab File: VY009191.D

Acq: 16 Jun 2022 09:53

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 76 Resp: 16027

Ion Ratio Lower Upper

76 100





78 33.0 25.0 37.6

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 06/17/2022

Supervised By :Mahesh Dadoda 06/17/2022



#58

2-Chloroethyl Vinyl ether

Concen: 252.649 ug/l

RT: 9.782 min Scan# 1306

Delta R.T. 0.006 min

Lab File: VY009191.D




Acq: 16 Jun 2022 09:53


Tgt Ion: 63 Resp: 280103

Ion Ratio Lower Upper

63 100


106 29.5 21.5 32.3





#59
2-Hexanone
Concen: 255.959 ug/l
RT: 10.831 min Scan# 1477
Delta R.T. 0.006 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 43 Resp: 31832
Ion Ratio Lower Upper
43 100
58 56.2 24.4 73.4




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/17/2022
Supervised By :Mahesh Dadoda 06/17/2022



#60
Dibromochloromethane
Concen: 48.769 ug/l
RT: 10.983 min Scan# 1503
Delta R.T. 0.006 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53


Tgt Ion:129 Resp: 112651
Ion Ratio Lower Upper
129 100
127 78.1 38.7 116.1





#61
1,2-Dibromoethane
Concen: 49.765 ug/l
RT: 11.087 min Scan# 1107
Delta R.T. 0.006 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion:107 Resp: 91507
Ion Ratio Lower Upper
107 100
109 94.6 75.4 113.2

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 06/17/2022
Supervised By :Mahesh Dadoda 06/17/2022




#62
4-Bromofluorobenzene
Concen: 50.841 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. 0.006 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53




Tgt Ion: 95 Resp: 139990
Ion Ratio Lower Upper
95 100
174 85.6 0.0 177.0
176 80.7 0.0 175.8





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.489 min Scan# 1586
Delta R.T. -0.000 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion:117 Resp: 292618
Ion Ratio Lower Upper
117 100
82 52.8 43.0 64.4
119 31.7 25.4 38.0

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 06/17/2022
Supervised By :Mahesh Dadoda 06/17/2022




#64
Tetrachloroethene
Concen: 48.875 ug/l
RT: 10.715 min Scan# 1459
Delta R.T. -0.000 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53



Tgt Ion:164 Resp: 132496
Ion Ratio Lower Upper
164 100
166 123.3 100.4 150.6
129 90.0 71.4 107.0
131 88.9 70.8 106.2









#65
Chlorobenzene
Concen: 48.711 ug/l
RT: 11.514 min Scan# 11590
Delta R.T. -0.000 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion:112 Resp: 308768
Ion Ratio Lower Upper
112 100
114 32.3 25.8 38.6




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/17/2022
Supervised By :Mahesh Dadoda 06/17/2022



#66
1,1,1,2-Tetrachloroethane
Concen: 49.384 ug/l
RT: 11.587 min Scan# 1602
Delta R.T. -0.000 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53


Tgt Ion:131 Resp: 114642
Ion Ratio Lower Upper
131 100
133 93.8 47.9 143.6
119 64.1 30.8 92.3





#67
Ethyl Benzene
Concen: 48.589 ug/l
RT: 11.593 min Scan# 1603
Delta R.T. 0.006 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 91 Resp: 537449
Ion Ratio Lower Upper
91 100
106 32.5 24.5 36.7




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/17/2022
Supervised By :Mahesh Dadoda 06/17/2022



#68
m/p-Xylenes
Concen: 96.109 ug/l
RT: 11.703 min Scan# 1621
Delta R.T. 0.006 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53


Tgt Ion:106 Resp: 419114
Ion Ratio Lower Upper
106 100
91 198.4 166.6 249.8





#69
o-Xylene
Concen: 49.099 ug/l
RT: 12.026 min Scan# 1674
Delta R.T. -0.000 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion:106 Resp: 20286
Ion Ratio Lower Upper
106 100
91 208.0 110.3 331.0

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 06/17/2022
Supervised By :Mahesh Dadoda 06/17/2022




#70
Styrene
Concen: 49.084 ug/l
RT: 12.044 min Scan# 1677
Delta R.T. 0.006 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53




Tgt Ion:104 Resp: 340213
Ion Ratio Lower Upper
104 100
78 48.2 42.4 63.6
103 54.0 44.4 66.6





#71
Bromoform
Concen: 48.356 ug/l
RT: 12.209 min Scan# 1703
Delta R.T. 0.006 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion:173 Resp: 71700
Ion Ratio Lower Upper
173 100
175 47.5 24.6 74.0
254 0.0 0.0 0.0

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 06/17/2022
Supervised By :Mahesh Dadoda 06/17/2022




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.428 min Scan# 1904
Delta R.T. 0.006 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53




Tgt Ion:152 Resp: 136710
Ion Ratio Lower Upper
152 100
115 57.2 28.2 84.7
150 176.1 0.0 347.6





#73
Isopropylbenzene
Concen: 50.681 ug/l
RT: 12.331 min Scan# 1724
Delta R.T. 0.006 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion:105 Resp: 52147
Ion Ratio Lower Upper
105 100
120 27.4 13.4 40.2

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 06/17/2022
Supervised By :Mahesh Dadoda 06/17/2022




#74
N-amy1 acetate
Concen: 48.381 ug/l
RT: 12.142 min Scan# 1693
Delta R.T. -0.000 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53




Tgt Ion: 43 Resp: 138862
Ion Ratio Lower Upper
43 100
70 44.9 28.7 43.1#
55 24.8 17.3 25.9
61 25.6 17.0 25.6#





#75
1,1,2,2-Tetrachloroethane
Concen: 49.659 ug/l
RT: 12.581 min Scan# 1765
Delta R.T. -0.000 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 83 Resp: 101129
Ion Ratio Lower Upper
83 100
131 11.5 5.6 16.8
85 65.1 32.3 96.8

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 06/17/2022
Supervised By :Mahesh Dadoda 06/17/2022




#76
1,2,3-Trichloropropane
Concen: 53.018 ug/l m
RT: 12.629 min Scan# 1773
Delta R.T. -0.000 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53




Tgt Ion: 75 Resp: 80031
Ion Ratio Lower Upper
75 100
77 0.0 0.0 0.0





#77
Bromobenzene
Concen: 50.409 ug/l
RT: 12.605 min Scan# 1769
Delta R.T. -0.000 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion:156 Resp: 12743
Ion Ratio Lower Upper
156 100
77 176.7 95.5 286.4
158 95.5 48.0 144.0

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 06/17/2022
Supervised By :Mahesh Dadoda 06/17/2022




#78
n-propylbenzene
Concen: 50.256 ug/l
RT: 12.672 min Scan# 1780
Delta R.T. 0.006 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53



Tgt Ion: 91 Resp: 625672
Ion Ratio Lower Upper
91 100
120 24.0 11.3 33.9





#79

2-Chlorotoluene

Concen: 50.580 ug/l

RT: 12.758 min Scan# 1793

Delta R.T. 0.006 min

Lab File: VY009191.D

Acq: 16 Jun 2022 09:53

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 91 Resp: 354793

Ion Ratio Lower Upper

91 100





126 35.9 16.9 50.6

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 06/17/2022

Supervised By :Mahesh Dadoda 06/17/2022



#80

1,3,5-Trimethylbenzene

Concen: 49.970 ug/l

RT: 12.812 min Scan# 1803

Delta R.T. 0.006 min

Lab File: VY009191.D




Acq: 16 Jun 2022 09:53


Tgt Ion:105 Resp: 429748

Ion Ratio Lower Upper

105 100


120 50.6 24.4 73.4





#81
trans-1,4-Dichloro-2-butene
Concen: 49.618 ug/l
RT: 12.380 min Scan# 1731
Delta R.T. 0.006 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 75 Resp: 3758
Ion Ratio Lower Upper
75 100
53 85.6 85.3 127.9
89 46.6 37.0 55.4




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/17/2022
Supervised By :Mahesh Dadoda 06/17/2022



#82
4-Chlorotoluene
Concen: 49.564 ug/l
RT: 12.855 min Scan# 1810
Delta R.T. 0.006 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53


Tgt Ion: 91 Resp: 362764
Ion Ratio Lower Upper
91 100
126 35.9 17.0 51.0





#83
tert-Butylbenzene
Concen: 49.895 ug/l
RT: 13.075 min Scan# 1845
Delta R.T. 0.006 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion:119 Resp: 377264
Ion Ratio Lower Upper
119 100
91 64.2 34.8 104.4
134 27.0 13.3 39.9

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 06/17/2022
Supervised By :Mahesh Dadoda 06/17/2022




#84
1,2,4-Trimethylbenzene
Concen: 50.462 ug/l
RT: 13.117 min Scan# 1853
Delta R.T. -0.000 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53



Tgt Ion:105 Resp: 425963
Ion Ratio Lower Upper
105 100
120 46.1 22.3 66.9





#85

sec-Butylbenzene

Concen: 49.683 ug/l

RT: 13.251 min Scan# 1875

Delta R.T. -0.000 min

Lab File: VY009191.D

Acq: 16 Jun 2022 09:53

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion:105 Resp: 55626

Ion Ratio Lower Upper

105 100





134 20.8 10.2 30.4

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 06/17/2022

Supervised By :Mahesh Dadoda 06/17/2022



#86

p-Isopropyltoluene

Concen: 49.531 ug/l

RT: 13.367 min Scan# 1894

Delta R.T. -0.000 min

Lab File: VY009191.D

Acq: 16 Jun 2022 09:53




Tgt Ion:119 Resp: 460201


Ion Ratio Lower Upper

119 100

134 27.1 13.2 39.5


91 23.3 12.4 37.2





#87
1,3-Dichlorobenzene
Concen: 49.303 ug/l
RT: 13.367 min Scan# 1893
Delta R.T. 0.006 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion:146 Resp: 240384
Ion Ratio Lower Upper
146 100
111 39.3 19.5 58.5
148 63.8 31.6 95.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/17/2022
Supervised By :Mahesh Dadoda 06/17/2022



#88
1,4-Dichlorobenzene
Concen: 48.712 ug/l
RT: 13.446 min Scan# 1907
Delta R.T. 0.006 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53

Tgt Ion:146 Resp: 241090
Ion Ratio Lower Upper
146 100
111 37.8 19.3 57.8
148 63.2 32.6 97.8









#89
n-Butylbenzene
Concen: 48.892 ug/l
RT: 13.696 min Scan# 1947
Delta R.T. 0.006 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 91 Resp: 42956
Ion Ratio Lower Upper
91 100
92 53.6 26.3 78.9
134 27.1 13.0 38.9




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/17/2022
Supervised By :Mahesh Dadoda 06/17/2022



#90
Hexachloroethane
Concen: 49.358 ug/l
RT: 13.959 min Scan# 1991
Delta R.T. -0.000 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53

Tgt Ion:117 Resp: 86472
Ion Ratio Lower Upper
117 100
201 76.0 43.0 128.8









#91
1,2-Dichlorobenzene
Concen: 49.674 ug/l
RT: 13.739 min Scan# 1954
Delta R.T. 0.006 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion:146 Resp: 21901
Ion Ratio Lower Upper
146 100
111 40.4 20.3 60.9
148 64.0 32.3 96.9




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/17/2022
Supervised By :Mahesh Dadoda 06/17/2022



#92
1,2-Dibromo-3-Chloropropane
Concen: 46.983 ug/l
RT: 14.355 min Scan# 2056
Delta R.T. -0.000 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53

Tgt Ion: 75 Resp: 17264
Ion Ratio Lower Upper
75 100
155 91.8 38.0 113.9
157 121.5 47.8 143.3





#93

1,2,4-Trichlorobenzene

Concen: 48.819 ug/l

RT: 15.007 min Scan# 2163

Delta R.T. -0.000 min

Lab File: VY009191.D

Acq: 16 Jun 2022 09:53

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion:180 Resp: 135328

Ion Ratio Lower Upper

180 100

182 94.6 47.3 141.9





145 30.4 14.6 43.8

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 06/17/2022

Supervised By :Mahesh Dadoda 06/17/2022



#94

Hexachlorobutadiene

Concen: 45.676 ug/l

RT: 15.111 min Scan# 2180

Delta R.T. -0.000 min

Lab File: VY009191.D

Acq: 16 Jun 2022 09:53




Tgt Ion:225 Resp: 69592


Ion Ratio Lower Upper

225 100

223 62.7 30.8 92.3


227 62.2 31.6 94.7





#95
Naphthalene
Concen: 50.540 ug/l
RT: 15.239 min Scan# 2201
Delta R.T. 0.006 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion:128 Resp: 292215
Ion Ratio Lower Upper
128 100
127 12.8 11.0 16.4
129 10.9 8.4 12.6

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 06/17/2022
Supervised By :Mahesh Dadoda 06/17/2022



#96
1,2,3-Trichlorobenzene
Concen: 48.609 ug/l
RT: 15.422 min Scan# 2231
Delta R.T. -0.000 min
Lab File: VY009191.D
Acq: 16 Jun 2022 09:53



Tgt Ion:180 Resp: 118594
Ion Ratio Lower Upper
180 100
182 94.4 47.3 141.9
145 32.2 15.3 45.9

