

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062022\
 Data File : VY009241.D
 Acq On : 20 Jun 2022 13:31
 Operator : KP/MD
 Sample : N3371-01
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VNJ-211

Quant Time: Jun 21 08:12:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061522S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 16 02:54:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	218986	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	379297	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	347466	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	152176	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	125724	54.961	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 109.920%	
35) Dibromofluoromethane	7.716	113	115503	48.570	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 97.140%	
50) Toluene-d8	10.179	98	446974	49.376	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 98.760%	
62) 4-Bromofluorobenzene	12.483	95	151968	47.242	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 94.480%	
Target Compounds						Qvalue
65) Chlorobenzene	11.514	112	9166	1.218	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062022\
Data File : VY009241.D
Acq On : 20 Jun 2022 13:31
Operator : KP/MD
Sample : N3371-01
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VNJ-211

Quant Time: Jun 21 08:12:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061522S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 16 02:54:06 2022
Response via : Initial Calibration

