

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062022\
 Data File : VY009243.D
 Acq On : 20 Jun 2022 14:41
 Operator : KP/MD
 Sample : N3397-01
 Misc : 5.38g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OR-2-061722

Quant Time: Jun 21 08:13:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061522S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 16 02:54:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.782	168	204690	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	359376	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	343841	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	159413	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	123354	57.691	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.380%
35) Dibromofluoromethane	7.715	113	111962	49.691	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.380%
50) Toluene-d8	10.178	98	438845	51.165	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	102.340%
62) 4-Bromofluorobenzene	12.477	95	153288	50.293	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	100.580%
Target Compounds						
					Qvalue	
65) Chlorobenzene	11.514	112	9710	1.304	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062022\
Data File : VY009243.D
Acq On : 20 Jun 2022 14:41
Operator : KP/MD
Sample : N3397-01
Misc : 5.38g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OR-2-061722

Quant Time: Jun 21 08:13:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061522S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 16 02:54:06 2022
Response via : Initial Calibration

