

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062222\
 Data File : VY009271.D
 Acq On : 22 Jun 2022 14:01
 Operator : KP/MD
 Sample : N3412-01
 Misc : 4.99g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 NB-7-062022

Quant Time: Jun 23 01:54:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062122S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 21 14:08:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	184004	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	322743	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	299541	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	137106	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	107641	53.265	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	106.540%
35) Dibromofluoromethane	7.716	113	99198	47.460	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	94.920%
50) Toluene-d8	10.179	98	390221	48.600	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	97.200%
62) 4-Bromofluorobenzene	12.477	95	128203	45.439	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	90.880%
Target Compounds						
65) Chlorobenzene	11.514	112	6840	1.014	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062222\
Data File : VY009271.D
Acq On : 22 Jun 2022 14:01
Operator : KP/MD
Sample : N3412-01
Misc : 4.99g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
NB-7-062022

Quant Time: Jun 23 01:54:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062122S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 21 14:08:27 2022
Response via : Initial Calibration

