



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062222\
 Data File : VY009263.D
 Acq On : 22 Jun 2022 08:41
 Operator : KP/MD
 Sample : BFB
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062122S.M
 Title : SW846 8260
 Last Update : Tue Jun 21 14:08:27 2022



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1739

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.3	13996	PASS
75	95	30	60	52.4	38099	PASS
95	95	100	100	100.0	72664	PASS
96	95	5	9	6.7	4890	PASS
173	174	0.00	2	1.5	918	PASS
174	95	50	100	86.1	62539	PASS
175	174	5	9	7.6	4732	PASS
176	174	95	101	98.2	61440	PASS
177	176	5	9	7.0	4308	PASS