

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071124\
 Data File : VY018774.D
 Acq On : 11 Jul 2024 12:55
 Operator : SY/MD
 Sample : VY0711SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0711SBL01

Quant Time: Jul 12 01:36:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y070924S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 09 14:28:51 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	71843	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	126430	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	114680	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	45731	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	38294	54.168	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	108.340%
35) Dibromofluoromethane	7.710	113	40189	50.753	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.500%
50) Toluene-d8	10.173	98	147012	49.004	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	98.000%
62) 4-Bromofluorobenzene	12.477	95	44913	43.838	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	87.680%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071124\
Data File : VY018774.D
Acq On : 11 Jul 2024 12:55
Operator : SY/MD
Sample : VY0711SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0711SBL01

Quant Time: Jul 12 01:36:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y070924S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 09 14:28:51 2024
Response via : Initial Calibration

