

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080522\
 Data File : VY009858.D
 Acq On : 05 Aug 2022 18:10
 Operator : KP/MD
 Sample : N4023-01
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP1-1-6

Quant Time: Aug 06 01:20:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080322S.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 04 02:27:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	2640	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	4063	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	3690	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	1210	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	1196	55.814	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	111.620%
35) Dibromofluoromethane	7.703	113	422	17.035	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	34.060%#
50) Toluene-d8	10.179	98	3843	53.297	ug/l	0.00
Spiked Amount	50.000	Range	78 - 125	Recovery	=	106.600%
62) 4-Bromofluorobenzene	12.477	95	836	24.922	ug/l	0.00
Spiked Amount	50.000	Range	50 - 146	Recovery	=	49.840%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080522\
Data File : VY009858.D
Acq On : 05 Aug 2022 18:10
Operator : KP/MD
Sample : N4023-01
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP1-1-6

Quant Time: Aug 06 01:20:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080322S.M
Quant Title : SW846 8260
QLast Update : Thu Aug 04 02:27:45 2022
Response via : Initial Calibration

