

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080522\
 Data File : VY009860.D
 Acq On : 05 Aug 2022 18:58
 Operator : KP/MD
 Sample : N4023-07
 Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP3-1-6

Quant Time: Aug 06 01:21:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080322S.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 04 02:27:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	2296	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	3655	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	3885	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	1537	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	1353	74.681	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	149.360%
35) Dibromofluoromethane	7.710	113	445	19.969	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	39.940%#
50) Toluene-d8	10.179	98	3204	48.808	ug/l	0.00
Spiked Amount	50.000	Range	78 - 125	Recovery	=	97.620%
62) 4-Bromofluorobenzene	12.471	95	1020	33.801	ug/l	0.00
Spiked Amount	50.000	Range	50 - 146	Recovery	=	67.600%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080522\
Data File : VY009860.D
Acq On : 05 Aug 2022 18:58
Operator : KP/MD
Sample : N4023-07
Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP3-1-6

Quant Time: Aug 06 01:21:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080322S.M
Quant Title : SW846 8260
QLast Update : Thu Aug 04 02:27:45 2022
Response via : Initial Calibration

