

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080522\
 Data File : VY009866.D
 Acq On : 05 Aug 2022 21:20
 Operator : KP/MD
 Sample : N4064-05
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-13

Quant Time: Aug 06 01:22:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080322S.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 04 02:27:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.777	168	17039	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	32143	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	33619	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	14273	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	8979	66.052	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	132.100%
35) Dibromofluoromethane	7.710	113	8426	42.994	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	85.980%
50) Toluene-d8	10.179	98	29077	50.623	ug/l	0.00
Spiked Amount	50.000	Range	78 - 125	Recovery	=	101.240%
62) 4-Bromofluorobenzene	12.477	95	7752	29.211	ug/l	0.00
Spiked Amount	50.000	Range	50 - 146	Recovery	=	58.420%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080522\
Data File : VY009866.D
Acq On : 05 Aug 2022 21:20
Operator : KP/MD
Sample : N4064-05
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-13

Quant Time: Aug 06 01:22:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080322S.M
Quant Title : SW846 8260
QLast Update : Thu Aug 04 02:27:45 2022
Response via : Initial Calibration

