

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080522\
 Data File : VY009867.D
 Acq On : 05 Aug 2022 21:43
 Operator : KP/MD
 Sample : N4064-06
 Misc : 5.07g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-14

Quant Time: Aug 06 01:23:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080322S.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 04 02:27:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	2506	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.685	114	3952	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	4203	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	1501	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	1528	77.512	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	155.020%
35) Dibromofluoromethane	7.704	113	522	21.664	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	43.320%#
50) Toluene-d8	10.179	98	3756	53.592	ug/l	0.00
Spiked Amount	50.000	Range	78 - 125	Recovery	=	107.180%
62) 4-Bromofluorobenzene	12.477	95	914	28.012	ug/l	0.00
Spiked Amount	50.000	Range	50 - 146	Recovery	=	56.020%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080522\
Data File : VY009867.D
Acq On : 05 Aug 2022 21:43
Operator : KP/MD
Sample : N4064-06
Misc : 5.07g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-14

Quant Time: Aug 06 01:23:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080322S.M
Quant Title : SW846 8260
QLast Update : Thu Aug 04 02:27:45 2022
Response via : Initial Calibration

