

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080822\
 Data File : VY009901.D
 Acq On : 09 Aug 2022 03:32
 Operator : KP/MD
 Sample : N4044-01
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VNJ-230

Quant Time: Aug 09 07:00:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080822S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 08 17:55:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	48767	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	98367	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	102243	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	46347	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	23839	64.059	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	128.120%
35) Dibromofluoromethane	7.710	113	22296	41.754	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	83.500%
50) Toluene-d8	10.173	98	79418	48.959	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	97.920%
62) 4-Bromofluorobenzene	12.477	95	24938	33.447	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	66.900%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080822\
Data File : VY009901.D
Acq On : 09 Aug 2022 03:32
Operator : KP/MD
Sample : N4044-01
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VNJ-230

Quant Time: Aug 09 07:00:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080822S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 08 17:55:50 2022
Response via : Initial Calibration

