

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080822\
 Data File : VY009904.D
 Acq On : 09 Aug 2022 04:43
 Operator : KP/MD
 Sample : N3962-06
 Misc : 3.86g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RVE-192

Quant Time: Aug 09 07:01:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080822S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 08 17:55:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	52542	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	106322	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	108300	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	48790	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	24663	61.253	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	122.500%
35) Dibromofluoromethane	7.710	113	23423	40.583	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	81.160%
50) Toluene-d8	10.179	98	85561	48.774	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	97.540%
62) 4-Bromofluorobenzene	12.477	95	26425	32.790	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	65.580%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080822\
Data File : VY009904.D
Acq On : 09 Aug 2022 04:43
Operator : KP/MD
Sample : N3962-06
Misc : 3.86g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RVE-192

Quant Time: Aug 09 07:01:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080822S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 08 17:55:50 2022
Response via : Initial Calibration

