

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080822\
 Data File : VY080822.D
 Acq On : 08 Aug 2022 18:55
 Operator : KP/MD
 Sample : VY0808SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0808SBL01

Quant Time: Aug 08 19:09:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080822S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 08 17:55:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.777	168	62130	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	120032	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	118890	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	54951	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	29726	62.560	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	125.120%
35) Dibromofluoromethane	7.710	113	27399	42.050	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	84.100%
50) Toluene-d8	10.173	98	101931	51.898	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.800%
62) 4-Bromofluorobenzene	12.477	95	33127	36.411	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	72.820%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080822\
Data File : VY080822.D
Acq On : 08 Aug 2022 18:55
Operator : KP/MD
Sample : VY0808SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0808SBL01

Quant Time: Aug 08 19:09:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080822S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 08 17:55:50 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.777 min Scan# 91
Delta R.T. -0.005 min
Lab File: VY009883.D
Acq: 08 Aug 2022 18:55




Instrument :
MSVOA_Y
ClientSampleId :
VY0808SBL01


Tgt Ion:168 Resp: 62130
Ion Ratio Lower Upper
168 100
99 55.4 42.1 63.1



#33
1,2-Dichloroethane-d4
Concen: 62.560 ug/l
RT: 8.130 min Scan# 1035
Delta R.T. 0.000 min
Lab File: VY009883.D
Acq: 08 Aug 2022 18:55

Tgt Ion: 65 Resp: 29726
Ion Ratio Lower Upper
65 100
67 54.1 0.0 108.4





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.685 min Scan# 1126

Delta R.T. 0.000 min

Lab File: VY009883.D

Acq: 08 Aug 2022 18:55

Instrument :

MSVOA_Y

ClientSampleId :

VY0808SBL01





Tgt Ion:114 Resp: 120032

Ion Ratio Lower Upper

114 100

63 19.0 0.0 38.0

88 13.8 0.0 30.0



#35

Dibromofluoromethane

Concen: 42.050 ug/l

RT: 7.710 min Scan# 966

Delta R.T. 0.001 min

Lab File: VY009883.D

Acq: 08 Aug 2022 18:55




Tgt Ion:113 Resp: 27399


Ion Ratio Lower Upper

113 100

111 102.9 80.6 120.8

192 19.1 16.2 24.4









#50
Toluene-d8
Concen: 51.898 ug/l
RT: 10.173 min Scan# 11
Delta R.T. 0.001 min
Lab File: VY009883.D
Acq: 08 Aug 2022 18:55




Instrument :
MSVOA_Y
ClientSampleId :
VY0808SBL01


Tgt Ion: 98 Resp: 101931
Ion Ratio Lower Upper
98 100
100 65.6 52.9 79.3



#62
4-Bromofluorobenzene
Concen: 36.411 ug/l
RT: 12.477 min Scan# 1748
Delta R.T. 0.000 min
Lab File: VY009883.D
Acq: 08 Aug 2022 18:55

Tgt Ion: 95 Resp: 33127
Ion Ratio Lower Upper
95 100
174 85.1 0.0 173.8
176 82.9 0.0 167.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.489 min Scan# 11
Delta R.T. 0.000 min
Lab File: VY009883.D
Acq: 08 Aug 2022 18:55

Instrument :
MSVOA_Y
ClientSampleId :
VY0808SBL01

Tgt Ion:117 Resp: 118890
Ion Ratio Lower Upper
117 100
82 51.5 41.8 62.6
119 31.6 25.6 38.4



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.422 min Scan# 1903
Delta R.T. 0.000 min
Lab File: VY009883.D
Acq: 08 Aug 2022 18:55

Tgt Ion:152 Resp: 54951
Ion Ratio Lower Upper
152 100
115 58.7 28.6 85.9
150 159.1 0.0 349.8

