



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080822\
 Data File : VY009873.D
 Acq On : 08 Aug 2022 09:28
 Operator : KP/MD
 Sample : BFB
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080822S.M
 Title : SW846 8260
 Last Update : Mon Aug 08 17:55:50 2022



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.5	2494	PASS
75	95	30	60	51.8	6304	PASS
95	95	100	100	100.0	12164	PASS
96	95	5	9	6.8	828	PASS
173	174	0.00	2	1.0	106	PASS
174	95	50	100	88.7	10786	PASS
175	174	5	9	6.8	735	PASS
176	174	95	101	95.9	10347	PASS
177	176	5	9	6.6	681	PASS