



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080822\
 Data File : VY009888.D
 Acq On : 08 Aug 2022 21:14
 Operator : KP/MD
 Sample : BFB
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080822S.M
 Title : SW846 8260
 Last Update : Mon Aug 08 17:55:50 2022



AutoFind: Scans 1747, 1748, 1749; Background Corrected with Scan 1740

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.0	2860	PASS
75	95	30	60	51.9	7437	PASS
95	95	100	100	100.0	14316	PASS
96	95	5	9	6.9	982	PASS
173	174	0.00	2	1.7	194	PASS
174	95	50	100	78.8	11288	PASS
175	174	5	9	7.1	801	PASS
176	174	95	101	97.6	11016	PASS
177	176	5	9	6.9	762	PASS