

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080922\
 Data File : VY009931.D
 Acq On : 09 Aug 2022 19:47
 Operator : KP/MD
 Sample : N4068-01
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BU-02-080522

Quant Time: Aug 09 23:10:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080822S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 08 17:55:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	64057	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	132597	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	134355	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	62354	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	25972	52.024	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	104.040%
35) Dibromofluoromethane	7.704	113	24926	34.629	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	69.260%
50) Toluene-d8	10.173	98	92720	41.363	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	82.720%
62) 4-Bromofluorobenzene	12.477	95	30161	30.010	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	60.020%
Target Compounds						
20) Methylene Chloride	4.692	84	13974	9.089	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080922\
Data File : VY009931.D
Acq On : 09 Aug 2022 19:47
Operator : KP/MD
Sample : N4068-01
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BU-02-080522

Quant Time: Aug 09 23:10:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080822S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 08 17:55:50 2022
Response via : Initial Calibration

