

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080922\
 Data File : VY009943.D
 Acq On : 10 Aug 2022 02:06
 Operator : KP/MD
 Sample : N4017-06
 Misc : 5.08g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB07

Quant Time: Aug 10 08:13:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080822S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 08 17:55:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	20117	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	40328	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.484	117	48102	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	18256	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	10013	65.344	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	130.680%
35) Dibromofluoromethane	7.710	113	8857	40.458	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	80.920%
50) Toluene-d8	10.179	98	29892	44.312	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	88.620%
62) 4-Bromofluorobenzene	12.477	95	9313	30.467	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	60.940%
Target Compounds						
20) Methylene Chloride	4.704	84	3909	6.939	ug/l	Qvalue 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080922\
Data File : VY009943.D
Acq On : 10 Aug 2022 02:06
Operator : KP/MD
Sample : N4017-06
Misc : 5.08g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB07

Quant Time: Aug 10 08:13:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080822S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 08 17:55:50 2022
Response via : Initial Calibration

