

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081222\
 Data File : VY009996.D
 Acq On : 12 Aug 2022 19:12
 Operator : KP/MD
 Sample : N4143-05
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CHRT-26414

Quant Time: Aug 13 02:57:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081222S.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 13 02:54:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	114858	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	231172	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	235176	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	113987	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	47011	40.348	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	80.700%
35) Dibromofluoromethane	7.710	113	47299	33.430	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	66.860%
50) Toluene-d8	10.173	98	173582	32.019	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	64.040%
62) 4-Bromofluorobenzene	12.477	95	69445	34.962	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	69.920%
Target Compounds						
16) Acetone	3.936	43	4206	15.031	ug/l	Qvalue # 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081222\
Data File : VY009996.D
Acq On : 12 Aug 2022 19:12
Operator : KP/MD
Sample : N4143-05
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
CHRT-26414

Quant Time: Aug 13 02:57:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081222S.M
Quant Title : SW846 8260
QLast Update : Sat Aug 13 02:54:57 2022
Response via : Initial Calibration

