

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081222\
 Data File : VY009990.D
 Acq On : 12 Aug 2022 16:40
 Operator : KP/MD
 Sample : VSTDICV050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ICVVY081222

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 08/15/2022
 Supervised By :Mahesh Dadoda 08/16/2022

Quant Time: Aug 13 02:56:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081222S.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 13 02:54:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.777	168	111007	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.679	114	180874	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	160968	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	77560	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	54026	47.978	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	95.960%		
35) Dibromofluoromethane	7.704	113	56872	51.374	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	102.740%		
50) Toluene-d8	10.173	98	218244	51.452	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery =	102.900%		
62) 4-Bromofluorobenzene	12.477	95	76000	48.903	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery =	97.800%		
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.900	85	49985	50.158	ug/l	99
3) Chloromethane	2.107	50	35586	48.615	ug/l	98
4) Vinyl Chloride	2.241	62	38885	50.417	ug/l	98
5) Bromomethane	2.637	94	29107	54.142	ug/l	95
6) Chloroethane	2.778	64	24033	47.132	ug/l	95
7) Trichlorofluoromethane	3.113	101	94346	49.379	ug/l	97
8) Diethyl Ether	3.515	74	34584	48.086	ug/l	89
9) 1,1,2-Trichlorotrifluo...	3.881	101	62013	50.509	ug/l	95
10) Methyl Iodide	4.076	142	73255	49.877	ug/l	97
11) Tert butyl alcohol	4.948	59	27641	251.033	ug/l	96
12) 1,1-Dichloroethene	3.857	96	59867	50.222	ug/l	92
13) Acrolein	3.716	56	32703	225.708	ug/l	98
14) Allyl chloride	4.460	41	82556	49.675	ug/l	93
15) Acrylonitrile	5.149	53	82944	240.946	ug/l	99
16) Acetone	3.936	43	73072	270.192	ug/l #	87
17) Carbon Disulfide	4.180	76	194335	50.093	ug/l	98
18) Methyl Acetate	4.460	43	37113	49.876	ug/l	91
19) Methyl tert-butyl Ether	5.210	73	147452	50.621	ug/l	97
20) Methylene Chloride	4.704	84	72345	51.394	ug/l #	85
21) trans-1,2-Dichloroethene	5.198	96	67388	50.869	ug/l	92
22) Diisopropyl ether	6.106	45	163553	48.677	ug/l	96
23) Vinyl Acetate	6.045	43	527607	254.136	ug/l	95
24) 1,1-Dichloroethane	6.003	63	105513	49.212	ug/l	99
25) 2-Butanone	6.972	43	100287	238.976	ug/l	94
26) 2,2-Dichloropropane	6.966	77	94590	49.775	ug/l	98
27) cis-1,2-Dichloroethene	6.972	96	71165	50.008	ug/l	94
28) Bromochloromethane	7.319	49	40833	47.682	ug/l	84
29) Tetrahydrofuran	7.338	42	63122	238.001	ug/l #	87
30) Chloroform	7.496	83	109677	49.470	ug/l	98
31) Cyclohexane	7.771	56	99968	49.598	ug/l	92
32) 1,1,1-Trichloroethane	7.691	97	97180	51.585	ug/l	99
36) 1,1-Dichloropropene	7.905	75	85831	52.067	ug/l	99
37) Ethyl Acetate	7.063	43	44395	49.900	ug/l #	95
38) Carbon Tetrachloride	7.886	117	89755	54.560	ug/l	100
39) Methylcyclohexane	9.173	83	110289	55.564	ug/l	98
40) Benzene	8.149	78	253677	50.852	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081222\
 Data File : VY009990.D
 Acq On : 12 Aug 2022 16:40
 Operator : KP/MD
 Sample : VSTDICV050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ICVVY081222

Quant Time: Aug 13 02:56:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081222S.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 13 02:54:57 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Krupa Patel 08/15/2022
 Supervised By :Mahesh Dadoda 08/16/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.307	41	25089m	53.091	ug/l	
42) 1,2-Dichloroethane	8.228	62	65865	50.969	ug/l	98
43) Isopropyl Acetate	8.264	43	77082	49.914	ug/l	93
44) Trichloroethene	8.935	130	70500	52.186	ug/l	96
45) 1,2-Dichloropropane	9.209	63	60147	50.250	ug/l	95
46) Dibromomethane	9.301	93	35942	49.876	ug/l	99
47) Bromodichloromethane	9.490	83	82305	51.374	ug/l	100
48) Methyl methacrylate	9.289	41	36005	53.136	ug/l	90
49) 1,4-Dioxane	9.289	88	8624	957.664	ug/l	97
51) 4-Methyl-2-Pentanone	10.063	43	210424	251.323	ug/l	92
52) Toluene	10.240	92	158693	52.824	ug/l	99
53) t-1,3-Dichloropropene	10.459	75	85202	51.748	ug/l	99
54) cis-1,3-Dichloropropene	9.923	75	99664	52.352	ug/l	95
55) 1,1,2-Trichloroethane	10.642	97	51595	50.061	ug/l	99
56) Ethyl methacrylate	10.508	69	68743	53.946	ug/l	91
57) 1,3-Dichloropropane	10.788	76	84960	50.820	ug/l	99
58) 2-Chloroethyl Vinyl ether	9.776	63	122737	220.051	ug/l	97
59) 2-Hexanone	10.825	43	151659	262.501	ug/l	92
60) Dibromochloromethane	10.983	129	61769	52.477	ug/l	99
61) 1,2-Dibromoethane	11.087	107	50207	50.863	ug/l	98
64) Tetrachloroethene	10.715	164	63772	54.230	ug/l	96
65) Chlorobenzene	11.514	112	168294	51.508	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.587	131	61005	52.413	ug/l	98
67) Ethyl Benzene	11.587	91	297795	54.208	ug/l	98
68) m/p-Xylenes	11.697	106	230552	108.418	ug/l	95
69) o-Xylene	12.026	106	108105	53.605	ug/l	97
70) Styrene	12.044	104	181940	52.962	ug/l	98
71) Bromoform	12.203	173	38187	51.936	ug/l #	99
73) Isopropylbenzene	12.325	105	287422	55.215	ug/l	100
74) N-amyl acetate	12.142	43	70158	52.686	ug/l	94
75) 1,1,2,2-Tetrachloroethane	12.581	83	60021	49.956	ug/l	97
76) 1,2,3-Trichloropropane	12.629	75	42907m	47.873	ug/l	
77) Bromobenzene	12.605	156	66991	53.197	ug/l	92
78) n-propylbenzene	12.666	91	362094	55.424	ug/l	98
79) 2-Chlorotoluene	12.751	91	203976	54.965	ug/l	98
80) 1,3,5-Trimethylbenzene	12.812	105	242305	56.971	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.373	75	21686	52.597	ug/l	93
82) 4-Chlorotoluene	12.855	91	211628	54.744	ug/l	96
83) tert-Butylbenzene	13.075	119	209957	56.161	ug/l	98
84) 1,2,4-Trimethylbenzene	13.117	105	239114	55.987	ug/l	98
85) sec-Butylbenzene	13.251	105	317424	55.421	ug/l	98
86) p-Isopropyltoluene	13.367	119	258130	57.094	ug/l	99
87) 1,3-Dichlorobenzene	13.367	146	130456	52.556	ug/l	98
88) 1,4-Dichlorobenzene	13.446	146	131294	50.880	ug/l	98
89) n-Butylbenzene	13.696	91	249410	57.136	ug/l	99
90) Hexachloroethane	13.959	117	51273	52.884	ug/l	92
91) 1,2-Dichlorobenzene	13.739	146	119467	53.506	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	14.355	75	9924	51.429	ug/l	97
93) 1,2,4-Trichlorobenzene	15.007	180	68614	54.010	ug/l	98
94) Hexachlorobutadiene	15.111	225	38908	54.234	ug/l	99
95) Naphthalene	15.239	128	151566	49.507	ug/l	100
96) 1,2,3-Trichlorobenzene	15.422	180	60352	53.118	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081222\
Data File : VY009990.D
Acq On : 12 Aug 2022 16:40
Operator : KP/MD
Sample : VSTDICV050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
ICVVY081222

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 08/15/2022
Supervised By :Mahesh Dadoda 08/16/2022

Quant Time: Aug 13 02:56:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081222S.M
Quant Title : SW846 8260
QLast Update : Sat Aug 13 02:54:57 2022
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081222\
 Data File : VY009990.D
 Acq On : 12 Aug 2022 16:40
 Operator : KP/MD
 Sample : VSTDICV050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1


Instrument :
 MSVOA_Y
 ClientSampleId :
 ICVVY081222

Quant Time: Aug 13 02:56:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081222S.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 13 02:54:57 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Krupa Patel 08/15/2022
 Supervised By :Mahesh Dadoda 08/16/2022









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.777 min Scan# 91
Delta R.T. -0.006 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40

Instrument :
MSVOA_Y
ClientSampleId :
ICVVY081222

Tgt Ion:168 Resp: 11100
Ion Ratio Lower Upper
168 100
99 57.3 42.1 63.1




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 08/15/2022
Supervised By :Mahesh Dadoda 08/16/2022



#2
Dichlorodifluoromethane
Concen: 50.158 ug/l
RT: 1.900 min Scan# 13
Delta R.T. -0.000 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40


Tgt Ion: 85 Resp: 49985
Ion Ratio Lower Upper
85 100
87 33.1 16.2 48.5





#3
Chloromethane
Concen: 48.615 ug/l
RT: 2.107 min Scan# 41
Delta R.T. 0.000 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40




Instrument :
MSVOA_Y
ClientSampleId :
ICVVY081222



Tgt Ion: 50 Resp: 3558
Ion Ratio Lower Upper
50 100
52 33.7 27.9 41.9




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 08/15/2022
Supervised By :Mahesh Dadoda 08/16/2022



#4
Vinyl Chloride
Concen: 50.417 ug/l
RT: 2.241 min Scan# 69
Delta R.T. 0.000 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40


Tgt Ion: 62 Resp: 38885
Ion Ratio Lower Upper
62 100
64 32.5 25.0 37.4





#5
Bromomethane
Concen: 54.142 ug/l
RT: 2.637 min Scan# 11
Delta R.T. 0.000 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40




Instrument :
MSVOA_Y
ClientSampleId :
ICVVY081222



Tgt Ion: 94 Resp: 2910
Ion Ratio Lower Upper
94 100
96 97.8 74.6 111.8




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 08/15/2022
Supervised By :Mahesh Dadoda 08/16/2022



#6
Chloroethane
Concen: 47.132 ug/l
RT: 2.778 min Scan# 157
Delta R.T. 0.000 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40


Tgt Ion: 64 Resp: 24033
Ion Ratio Lower Upper
64 100
66 29.6 25.8 38.8





#7
 Trichlorofluoromethane
 Concen: 49.379 ug/l
 RT: 3.113 min Scan# 212
 Delta R.T. 0.000 min
 Lab File: VY009990.D
 Acq: 12 Aug 2022 16:40




Instrument : MSVOA_Y
 ClientSampleId : ICVVY081222




Tgt Ion: 101 Resp: 94340
 Ion Ratio Lower Upper
 101 100
 103 62.2 51.4 77.2

Manual Integrations
 APPROVED



Reviewed By :Krupa Patel 08/15/2022
 Supervised By :Mahesh Dadoda 08/16/2022




#8
 Diethyl Ether
 Concen: 48.086 ug/l
 RT: 3.515 min Scan# 278
 Delta R.T. 0.000 min
 Lab File: VY009990.D
 Acq: 12 Aug 2022 16:40



Tgt Ion: 74 Resp: 34584
 Ion Ratio Lower Upper
 74 100
 45 87.3 49.1 147.4









#9
1,1,2-Trichlorotrifluoroethane
Concen: 50.509 ug/l
RT: 3.881 min Scan# 31
Delta R.T. 0.000 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40

Instrument :
MSVOA_Y
ClientSampleId :
ICVY081222

Tgt Ion:101 Resp: 6201
Ion Ratio Lower Upper
101 100
85 41.5 36.2 54.2
151 75.6 63.6 95.4




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 08/15/2022
Supervised By :Mahesh Dadoda 08/16/2022



#10
Methyl Iodide
Concen: 49.877 ug/l
RT: 4.076 min Scan# 370
Delta R.T. 0.000 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40

Tgt Ion:142 Resp: 73255
Ion Ratio Lower Upper
142 100
127 41.9 35.2 52.8
141 14.5 11.4 17.2









#11
Tert butyl alcohol
Concen: 251.033 ug/l
RT: 4.948 min Scan# 513
Delta R.T. -0.000 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40

Instrument :
MSVOA_Y
ClientSampleId :
ICVY081222

Tgt Ion: 59 Resp: 2764
Ion Ratio Lower Upper
59 100
57 9.4 6.3 9.5




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 08/15/2022
Supervised By :Mahesh Dadoda 08/16/2022



#12
1,1-Dichloroethene
Concen: 50.222 ug/l
RT: 3.857 min Scan# 334
Delta R.T. 0.000 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40


Tgt Ion: 96 Resp: 59867
Ion Ratio Lower Upper
96 100
61 138.6 121.9 182.9
98 61.0 48.2 72.4





#13
Acrolein
Concen: 225.708 ug/l
RT: 3.716 min Scan# 311
Delta R.T. 0.000 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40




Instrument :
MSVOA_Y
ClientSampleId :
ICVVY081222



Tgt Ion: 56 Resp: 3270
Ion Ratio Lower Upper
56 100
55 70.2 54.8 82.2




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 08/15/2022
Supervised By :Mahesh Dadoda 08/16/2022



#14
Allyl chloride
Concen: 49.675 ug/l
RT: 4.460 min Scan# 433
Delta R.T. 0.000 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40


Tgt Ion: 41 Resp: 82556
Ion Ratio Lower Upper
41 100
39 63.8 53.5 80.3
76 43.8 29.4 44.0





#15
Acrylonitrile
Concen: 240.946 ug/l
RT: 5.149 min Scan# 547
Delta R.T. -0.006 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40




Instrument :
MSVOA_Y
ClientSampleId :
ICVY081222



Tgt Ion: 53 Resp: 82944
Ion Ratio Lower Upper
53 100
52 80.7 65.4 98.2
51 34.5 28.4 42.6




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 08/15/2022
Supervised By :Mahesh Dadoda 08/16/2022



#16
Acetone
Concen: 270.192 ug/l
RT: 3.936 min Scan# 347
Delta R.T. 0.000 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40


Tgt Ion: 43 Resp: 73072
Ion Ratio Lower Upper
43 100
58 37.3 24.0 36.0#





#17
Carbon Disulfide
Concen: 50.093 ug/l
RT: 4.180 min Scan# 386
Delta R.T. 0.006 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40




Instrument :
MSVOA_Y
ClientSampleId :
ICVY081222



Tgt Ion: 76 Resp: 19433
Ion Ratio Lower Upper
76 100
78 9.4 6.8 10.2




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 08/15/2022
Supervised By :Mahesh Dadoda 08/16/2022



#18
Methyl Acetate
Concen: 49.876 ug/l
RT: 4.460 min Scan# 433
Delta R.T. -0.006 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40


Tgt Ion: 43 Resp: 37113
Ion Ratio Lower Upper
43 100
74 27.9 18.8 28.2





#19
Methyl tert-butyl Ether
Concen: 50.621 ug/l
RT: 5.210 min Scan# 51
Delta R.T. 0.000 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40




Instrument :
MSVOA_Y
ClientSampleId :
ICVVY081222



Tgt Ion: 73 Resp: 147452
Ion Ratio Lower Upper
73 100
57 21.9 18.5 27.7




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 08/15/2022
Supervised By :Mahesh Dadoda 08/16/2022



#20
Methylene Chloride
Concen: 51.394 ug/l
RT: 4.704 min Scan# 473
Delta R.T. 0.006 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40


Tgt Ion: 84 Resp: 72345
Ion Ratio Lower Upper
84 100
49 101.1 98.2 147.4
51 32.1 33.4 50.0#
86 61.1 52.3 78.5





#21
trans-1,2-Dichloroethene
Concen: 50.869 ug/l
RT: 5.198 min Scan# 511
Delta R.T. -0.006 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40




Instrument :
MSVOA_Y
ClientSampleId :
ICVVY081222



Tgt Ion: 96 Resp: 67388
Ion Ratio Lower Upper
96 100
61 119.1 105.8 158.6
98 60.9 49.7 74.5




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 08/15/2022
Supervised By :Mahesh Dadoda 08/16/2022



#22
Diisopropyl ether
Concen: 48.677 ug/l
RT: 6.106 min Scan# 703
Delta R.T. 0.000 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40

Tgt Ion: 45 Resp: 163553
Ion Ratio Lower Upper
45 100
43 56.7 43.7 65.5
87 31.4 22.2 33.4
59 12.6 9.5 14.3





#23

Vinyl Acetate

Concen: 254.136 ug/l

RT: 6.045 min Scan# 693

Delta R.T. 0.000 min

Lab File: VY009990.D

Acq: 12 Aug 2022 16:40

Instrument :

MSVOA_Y

ClientSampleId :

ICVY081222

Tgt Ion: 43 Resp: 52760

Ion Ratio Lower Upper

43 100





86 13.3 9.0 13.6

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 08/15/2022

Supervised By :Mahesh Dadoda 08/16/2022



#24

1,1-Dichloroethane

Concen: 49.212 ug/l

RT: 6.003 min Scan# 686

Delta R.T. 0.000 min

Lab File: VY009990.D

Acq: 12 Aug 2022 16:40




Tgt Ion: 63 Resp: 105513


Ion Ratio Lower Upper

63 100

98 7.9 3.9 11.7


100 5.8 2.8 8.3





#25
2-Butanone
Concen: 238.976 ug/l
RT: 6.972 min Scan# 845
Delta R.T. -0.006 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40




Instrument :
MSVOA_Y
ClientSampleId :
ICVY081222



Tgt Ion: 43 Resp: 10028
Ion Ratio Lower Upper
43 100
72 29.9 21.5 32.3




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 08/15/2022
Supervised By :Mahesh Dadoda 08/16/2022



#26
2,2-Dichloropropane
Concen: 49.775 ug/l
RT: 6.966 min Scan# 844
Delta R.T. -0.006 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40


Tgt Ion: 77 Resp: 94590
Ion Ratio Lower Upper
77 100
97 24.4 11.7 35.0





#27
 cis-1,2-Dichloroethene
 Concen: 50.008 ug/l
 RT: 6.972 min Scan# 845
 Delta R.T. 0.000 min
 Lab File: VY009990.D
 Acq: 12 Aug 2022 16:40




Instrument :
 MSVOA_Y
 ClientSampleId :
 ICVVY081222



Tgt Ion: 96 Resp: 7116
 Ion Ratio Lower Upper
 96 100
 61 132.2 0.0 285.6
 98 64.5 0.0 131.0




Manual Integrations
 APPROVED


Reviewed By :Krupa Patel 08/15/2022
 Supervised By :Mahesh Dadoda 08/16/2022



#28
 Bromochloromethane
 Concen: 47.682 ug/l
 RT: 7.319 min Scan# 902
 Delta R.T. -0.006 min
 Lab File: VY009990.D
 Acq: 12 Aug 2022 16:40

Tgt Ion: 49 Resp: 40833
 Ion Ratio Lower Upper
 49 100
 129 2.6 0.0 4.6
 130 94.8 64.1 96.1





#29

Tetrahydrofuran

Concen: 238.001 ug/l

RT: 7.338 min Scan# 905

Delta R.T. 0.000 min

Lab File: VY009990.D

Acq: 12 Aug 2022 16:40

Instrument :

MSVOA_Y

ClientSampleId :

ICVY081222





Tgt Ion:	42	Resp:	6312
Ion Ratio	Lower	Upper	
42	100		
72	52.3	34.6	52.0
71	48.4	32.7	49.1

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 08/15/2022

Supervised By :Mahesh Dadoda 08/16/2022



#30

Chloroform

Concen: 49.470 ug/l




RT: 7.496 min Scan# 931


Delta R.T. 0.000 min

Lab File: VY009990.D

Acq: 12 Aug 2022 16:40


Tgt Ion:	83	Resp:	109677
Ion Ratio	Lower	Upper	
83	100		
85	64.5	50.5	75.7





#31
Cyclohexane
Concen: 49.598 ug/l
RT: 7.771 min Scan# 91
Delta R.T. -0.006 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40




Instrument :
MSVOA_Y
ClientSampleId :
ICVY081222



Tgt Ion: 56 Resp: 9996
Ion Ratio Lower Upper
56 100
69 36.1 26.0 39.0
84 94.2 69.5 104.3




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 08/15/2022
Supervised By :Mahesh Dadoda 08/16/2022



#32
1,1,1-Trichloroethane
Concen: 51.585 ug/l
RT: 7.691 min Scan# 963
Delta R.T. 0.000 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40

Tgt Ion: 97 Resp: 97180
Ion Ratio Lower Upper
97 100
99 64.5 51.8 77.8
61 40.2 33.7 50.5





#33

1,2-Dichloroethane-d4

Concen: 47.978 ug/l

RT: 8.130 min Scan# 1035

Delta R.T. 0.000 min

Lab File: VY009990.D

Acq: 12 Aug 2022 16:40

Instrument :

MSVOA_Y

ClientSampleId :

ICVY081222

Tgt Ion: 65 Resp: 54020

Ion Ratio Lower Upper

65 100





67 55.8 0.0 108.4

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 08/15/2022

Supervised By :Mahesh Dadoda 08/16/2022



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.679 min Scan# 1125

Delta R.T. -0.006 min

Lab File: VY009990.D

Acq: 12 Aug 2022 16:40




Tgt Ion:114 Resp: 180874


Ion Ratio Lower Upper

114 100

63 19.2 0.0 38.0


88 15.4 0.0 30.0





#35
Dibromofluoromethane
Concen: 51.374 ug/l
RT: 7.704 min Scan# 911
Delta R.T. -0.006 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40




Instrument :
MSVOA_Y
ClientSampleId :
ICVY081222




Tgt Ion: 113 Resp: 56872
Ion Ratio Lower Upper
113 100
111 103.0 80.6 120.8
192 18.9 16.2 24.4

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 08/15/2022
Supervised By :Mahesh Dadoda 08/16/2022




#36
1,1-Dichloropropene
Concen: 52.067 ug/l
RT: 7.905 min Scan# 998
Delta R.T. -0.006 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40



Tgt Ion: 75 Resp: 85831
Ion Ratio Lower Upper
75 100
110 37.6 18.6 56.0
77 31.9 25.0 37.4





#37

Ethyl Acetate

Concen: 49.900 ug/l

RT: 7.063 min Scan# 80

Delta R.T. 0.000 min

Lab File: VY009990.D

Acq: 12 Aug 2022 16:40

Instrument :

MSVOA_Y

ClientSampleId :

ICVVY081222

Tgt Ion: 43 Resp: 4439





Ion	Ratio	Lower	Upper
43	100		
61	15.9	11.2	16.8
70	12.6	8.4	12.6

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 08/15/2022

Supervised By :Mahesh Dadoda 08/16/2022



#38

Carbon Tetrachloride

Concen: 54.560 ug/l

RT: 7.886 min Scan# 995




Delta R.T. -0.006 min


Lab File: VY009990.D

Acq: 12 Aug 2022 16:40

Tgt Ion:117 Resp: 89755

Ion	Ratio	Lower	Upper
117	100		
119	95.0	75.9	113.9
121	30.6	24.9	37.3





#39

Methylcyclohexane

Concen: 55.564 ug/l

RT: 9.173 min Scan# 110289

Delta R.T. -0.006 min

Lab File: VY009990.D

Acq: 12 Aug 2022 16:40

Instrument :

MSVOA_Y

ClientSampleId :

ICVY081222

Tgt Ion: 83 Resp: 110289





Ion	Ratio	Lower	Upper
83	100		
55	72.9	56.7	85.1
98	50.3	39.9	59.9

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 08/15/2022

Supervised By :Mahesh Dadoda 08/16/2022



#40

Benzene

Concen: 50.852 ug/l

RT: 8.149 min Scan# 1038




Delta R.T. 0.000 min


Lab File: VY009990.D

Acq: 12 Aug 2022 16:40

Tgt Ion: 78 Resp: 253677


Ion	Ratio	Lower	Upper
78	100		
77	23.3	19.4	29.0





#41
Methacrylonitrile
Concen: 53.091 ug/l m
RT: 7.307 min Scan# 901
Delta R.T. -0.006 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40




Instrument :
MSVOA_Y
ClientSampleId :
ICVY081222



Tgt Ion: 41 Resp: 25089
Ion Ratio Lower Upper
41 100
39 0.0 0.0 0.0
67 0.0 0.0 0.0
52 0.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 08/15/2022
Supervised By :Mahesh Dadoda 08/16/2022



#42
1,2-Dichloroethane
Concen: 50.969 ug/l
RT: 8.228 min Scan# 1051
Delta R.T. 0.000 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40


Tgt Ion: 62 Resp: 65865
Ion Ratio Lower Upper
62 100
98 11.1 0.0 20.6





#43
Isopropyl Acetate
Concen: 49.914 ug/l
RT: 8.264 min Scan# 1057
Delta R.T. 0.000 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40




Instrument :
MSVOA_Y
ClientSampleId :
ICVVY081222




Tgt Ion: 43 Resp: 77082
Ion Ratio Lower Upper
43 100
61 33.6 23.6 35.4
87 16.6 11.3 16.9

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 08/15/2022
Supervised By :Mahesh Dadoda 08/16/2022




#44
Trichloroethene
Concen: 52.186 ug/l
RT: 8.935 min Scan# 1167
Delta R.T. 0.006 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40




Tgt Ion:130 Resp: 70500
Ion Ratio Lower Upper
130 100
95 93.8 0.0 194.6





#45
1,2-Dichloropropane
Concen: 50.250 ug/l
RT: 9.209 min Scan# 1212
Delta R.T. 0.000 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40




Instrument :
MSVOA_Y
ClientSampleId :
ICVVY081222



Tgt Ion: 63 Resp: 6014
Ion Ratio Lower Upper
63 100
65 32.3 23.7 35.5




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 08/15/2022
Supervised By :Mahesh Dadoda 08/16/2022



#46
Dibromomethane
Concen: 49.876 ug/l
RT: 9.301 min Scan# 1227
Delta R.T. 0.000 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40

Tgt Ion: 93 Resp: 35942
Ion Ratio Lower Upper
93 100
95 83.9 66.3 99.5
174 100.4 80.6 120.8









#47
Bromodichloromethane
Concen: 51.374 ug/l
RT: 9.490 min Scan# 1258
Delta R.T. 0.000 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40

Instrument :
MSVOA_Y
ClientSampleId :
ICVVY081222

Tgt Ion: 83 Resp: 82305
Ion Ratio Lower Upper
83 100
85 63.7 51.2 76.8
127 8.4 7.1 10.7




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 08/15/2022
Supervised By :Mahesh Dadoda 08/16/2022



#48
Methyl methacrylate
Concen: 53.136 ug/l
RT: 9.289 min Scan# 1225
Delta R.T. 0.000 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40

Tgt Ion: 41 Resp: 36005
Ion Ratio Lower Upper
41 100
69 101.6 73.2 109.8
39 53.4 47.5 71.3





#49

1,4-Dioxane

Concen: 957.664 ug/l

RT: 9.289 min Scan# 1225

Delta R.T. -0.000 min

Lab File: VY009990.D

Acq: 12 Aug 2022 16:40

Instrument :

MSVOA_Y

ClientSampleId :

ICVY081222

Tgt Ion: 88 Resp: 8624

Ion Ratio Lower Upper

88 100

43 29.0 26.3 39.5





58 66.1 52.6 79.0

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 08/15/2022

Supervised By :Mahesh Dadoda 08/16/2022



#50

Toluene-d8

Concen: 51.452 ug/l

RT: 10.173 min Scan# 1370

Delta R.T. 0.000 min

Lab File: VY009990.D




Acq: 12 Aug 2022 16:40


Tgt Ion: 98 Resp: 218244

Ion Ratio Lower Upper

98 100


100 65.7 52.9 79.3





#51
4-Methyl-2-Pentanone
Concen: 251.323 ug/l
RT: 10.063 min Scan# 1352
Delta R.T. 0.000 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40




Instrument :
MSVOA_Y
ClientSampleId :
ICVVY081222



Tgt Ion: 43 Resp: 210424
Ion Ratio Lower Upper
43 100
58 44.1 31.5 47.3




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 08/15/2022
Supervised By :Mahesh Dadoda 08/16/2022



#52
Toluene
Concen: 52.824 ug/l
RT: 10.240 min Scan# 1381
Delta R.T. 0.000 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40


Tgt Ion: 92 Resp: 158693
Ion Ratio Lower Upper
92 100
91 171.9 138.7 208.1





#53
t-1,3-Dichloropropene
Concen: 51.748 ug/l
RT: 10.459 min Scan# 1417
Delta R.T. 0.000 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40




Instrument : MSVOA_Y
ClientSampleId : ICVVY081222



Tgt Ion: 75 Resp: 8520
Ion Ratio Lower Upper
75 100
77 31.2 25.4 38.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 08/15/2022
Supervised By :Mahesh Dadoda 08/16/2022



#54
cis-1,3-Dichloropropene
Concen: 52.352 ug/l
RT: 9.923 min Scan# 1329
Delta R.T. 0.000 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40


Tgt Ion: 75 Resp: 99664
Ion Ratio Lower Upper
75 100
77 31.7 25.2 37.8
39 39.2 35.8 53.8





#55
1,1,2-Trichloroethane
Concen: 50.061 ug/l
RT: 10.642 min Scan# 1447
Delta R.T. 0.000 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40




Instrument :
MSVOA_Y
ClientSampleId :
ICVVY081222




Tgt Ion:	97	Resp:	5159
Ion Ratio	100	Lower	Upper
97	100		
83	83.5	66.9	100.3
85	55.2	44.2	66.4
99	61.4	50.0	75.0

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 08/15/2022
Supervised By :Mahesh Dadoda 08/16/2022




#56
Ethyl methacrylate
Concen: 53.946 ug/l
RT: 10.508 min Scan# 1425
Delta R.T. 0.000 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40




Tgt Ion:	69	Resp:	68743
Ion Ratio	100	Lower	Upper
69	100		
41	55.0	50.4	75.6
39	30.5	27.6	41.4





#57
1,3-Dichloropropane
Concen: 50.820 ug/l
RT: 10.788 min Scan# 1471
Delta R.T. 0.006 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40




Instrument :
MSVOA_Y
ClientSampleId :
ICVY081222



Tgt Ion: 76 Resp: 84960
Ion Ratio Lower Upper
76 100
78 32.1 26.1 39.1




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 08/15/2022
Supervised By :Mahesh Dadoda 08/16/2022



#58
2-Chloroethyl Vinyl ether
Concen: 220.051 ug/l
RT: 9.776 min Scan# 1305
Delta R.T. 0.000 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40


Tgt Ion: 63 Resp: 122737
Ion Ratio Lower Upper
63 100
106 30.6 23.3 34.9





#59
2-Hexanone
Concen: 262.501 ug/l
RT: 10.825 min Scan# 1477
Delta R.T. 0.000 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40




Instrument :
MSVOA_Y
ClientSampleId :
ICVVY081222



Tgt Ion: 43 Resp: 151659
Ion Ratio Lower Upper
43 100
58 61.3 27.6 82.8




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 08/15/2022
Supervised By :Mahesh Dadoda 08/16/2022



#60
Dibromochloromethane
Concen: 52.477 ug/l
RT: 10.983 min Scan# 1503
Delta R.T. 0.006 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40


Tgt Ion:129 Resp: 61769
Ion Ratio Lower Upper
129 100
127 76.6 38.8 116.4





#61
1,2-Dibromoethane
Concen: 50.863 ug/l
RT: 11.087 min Scan# 1520
Delta R.T. 0.000 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40




Instrument :
MSVOA_Y
ClientSampleId :
ICVY081222




Tgt Ion:107 Resp: 5020
Ion Ratio Lower Upper
107 100
109 94.4 73.9 110.9

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 08/15/2022
Supervised By :Mahesh Dadoda 08/16/2022




#62
4-Bromofluorobenzene
Concen: 48.903 ug/l
RT: 12.477 min Scan# 1748
Delta R.T. 0.000 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40




Tgt Ion: 95 Resp: 76000
Ion Ratio Lower Upper
95 100
174 76.5 0.0 173.8
176 74.0 0.0 167.0





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.489 min Scan# 1586
Delta R.T. 0.000 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40




Instrument :
MSVOA_Y
ClientSampleId :
ICVVY081222



Tgt Ion:117 Resp: 160968
Ion Ratio Lower Upper
117 100
82 52.0 41.8 62.6
119 31.3 25.6 38.4




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 08/15/2022
Supervised By :Mahesh Dadoda 08/16/2022



#64
Tetrachloroethene
Concen: 54.230 ug/l
RT: 10.715 min Scan# 1459
Delta R.T. 0.000 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40


Tgt Ion:164 Resp: 63772
Ion Ratio Lower Upper
164 100
166 124.2 104.9 157.3
129 94.5 76.5 114.7
131 94.6 72.3 108.5





#65
Chlorobenzene
Concen: 51.508 ug/l
RT: 11.514 min Scan# 1590
Delta R.T. 0.000 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40




Instrument :
MSVOA_Y
ClientSampleId :
ICVVY081222




Tgt Ion:112 Resp: 168294
Ion Ratio Lower Upper
112 100
114 32.7 25.3 37.9

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 08/15/2022
Supervised By :Mahesh Dadoda 08/16/2022




#66
1,1,1,2-Tetrachloroethane
Concen: 52.413 ug/l
RT: 11.587 min Scan# 1602
Delta R.T. 0.000 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40



Tgt Ion:131 Resp: 61005
Ion Ratio Lower Upper
131 100
133 95.0 47.3 142.0
119 68.0 32.1 96.3





#67

Ethyl Benzene

Concen: 54.208 ug/l

RT: 11.587 min Scan# 1602

Delta R.T. 0.000 min

Lab File: VY009990.D

Acq: 12 Aug 2022 16:40

Instrument :

MSVOA_Y

ClientSampleId :

ICVY081222

Tgt Ion: 91 Resp: 29779

Ion Ratio Lower Upper

91 100





106 31.5 26.2 39.4

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 08/15/2022

Supervised By :Mahesh Dadoda 08/16/2022



#68

m/p-Xylenes

Concen: 108.418 ug/l

RT: 11.697 min Scan# 1620

Delta R.T. 0.000 min

Lab File: VY009990.D




Acq: 12 Aug 2022 16:40


Tgt Ion:106 Resp: 230552

Ion Ratio Lower Upper

106 100


91 203.3 156.2 234.2





#69
o-Xylene
Concen: 53.605 ug/l
RT: 12.026 min Scan# 1674
Delta R.T. 0.000 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40




Instrument :
MSVOA_Y
ClientSampleId :
ICVVY081222




Tgt Ion:106 Resp: 10810
Ion Ratio Lower Upper
106 100
91 211.0 103.2 309.4

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 08/15/2022
Supervised By :Mahesh Dadoda 08/16/2022




#70
Styrene
Concen: 52.962 ug/l
RT: 12.044 min Scan# 1677
Delta R.T. 0.006 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40



Tgt Ion:104 Resp: 181940
Ion Ratio Lower Upper
104 100
78 49.7 38.6 57.8
103 55.2 43.3 64.9





#71

Bromoform

Concen: 51.936 ug/l

RT: 12.203 min Scan# 1703

Delta R.T. 0.000 min

Lab File: VY009990.D

Acq: 12 Aug 2022 16:40

Instrument :

MSVOA_Y

ClientSampleId :

ICVY081222





Tgt Ion:	173	Resp:	3818
Ion Ratio	Lower	Upper	
173	100		
175	48.0	24.4	73.4
254	0.0	0.1	0.1

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 08/15/2022

Supervised By :Mahesh Dadoda 08/16/2022



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l




RT: 13.422 min Scan# 1903


Delta R.T. 0.000 min

Lab File: VY009990.D

Acq: 12 Aug 2022 16:40


Tgt Ion:	152	Resp:	77560
Ion Ratio	Lower	Upper	
152	100		
115	59.4	28.6	85.9
150	172.7	0.0	349.8





#73
Isopropylbenzene
Concen: 55.215 ug/l
RT: 12.325 min Scan# 1723
Delta R.T. 0.000 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40




Instrument :
MSVOA_Y
ClientSampleId :
ICVVY081222



Tgt Ion:105 Resp: 28742
Ion Ratio Lower Upper
105 100
120 27.3 13.7 41.1




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 08/15/2022
Supervised By :Mahesh Dadoda 08/16/2022



#74
N-ethyl acetate
Concen: 52.686 ug/l
RT: 12.142 min Scan# 1693
Delta R.T. 0.000 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40


Tgt Ion: 43 Resp: 70158
Ion Ratio Lower Upper
43 100
70 49.5 35.6 53.4
55 27.0 19.5 29.3
61 26.8 19.7 29.5





#75
1,1,2,2-Tetrachloroethane
Concen: 49.956 ug/l
RT: 12.581 min Scan# 1765
Delta R.T. 0.000 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40




Instrument :
MSVOA_Y
ClientSampleId :
ICVVY081222




Tgt Ion: 83 Resp: 6002
Ion Ratio Lower Upper
83 100
131 9.9 5.4 16.2
85 63.5 32.8 98.4

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 08/15/2022
Supervised By :Mahesh Dadoda 08/16/2022




#76
1,2,3-Trichloropropane
Concen: 47.873 ug/l m
RT: 12.629 min Scan# 1773
Delta R.T. 0.000 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40



Tgt Ion: 75 Resp: 42907
Ion Ratio Lower Upper
75 100
77 0.0 0.0 0.0





#77

Bromobenzene

Concen: 53.197 ug/l

RT: 12.605 min Scan# 1769

Delta R.T. 0.000 min

Lab File: VY009990.D

Acq: 12 Aug 2022 16:40

Instrument :

MSVOA_Y

ClientSampleId :

ICVVY081222





Tgt Ion:	156	Resp:	6699
Ion Ratio	Lower	Upper	
156	100		
77	193.6	88.1	264.3
158	95.9	48.6	145.8

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 08/15/2022

Supervised By :Mahesh Dadoda 08/16/2022



#78

n-propylbenzene

Concen: 55.424 ug/l




RT: 12.666 min Scan# 1779


Delta R.T. 0.000 min

Lab File: VY009990.D

Acq: 12 Aug 2022 16:40


Tgt Ion:	91	Resp:	362094
Ion Ratio	Lower	Upper	
91	100		
120	22.5	11.9	35.5





#79
2-Chlorotoluene
Concen: 54.965 ug/l
RT: 12.751 min Scan# 1793
Delta R.T. 0.000 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40




Instrument :
MSVOA_Y
ClientSampleId :
ICVVY081222



Tgt Ion: 91 Resp: 203970
Ion Ratio Lower Upper
91 100
126 34.4 17.8 53.4




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 08/15/2022
Supervised By :Mahesh Dadoda 08/16/2022



#80
1,3,5-Trimethylbenzene
Concen: 56.971 ug/l
RT: 12.812 min Scan# 1803
Delta R.T. 0.006 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40

Tgt Ion:105 Resp: 242305
Ion Ratio Lower Upper
105 100
120 49.7 25.3 75.8









#81
trans-1,4-Dichloro-2-butene
Concen: 52.597 ug/l
RT: 12.373 min Scan# 1731
Delta R.T. 0.000 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40

Instrument :
MSVOA_Y
ClientSampleId :
ICVVY081222

Tgt Ion: 75 Resp: 21680
Ion Ratio Lower Upper
75 100
53 81.3 71.4 107.2
89 45.3 38.1 57.1




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 08/15/2022
Supervised By :Mahesh Dadoda 08/16/2022



#82
4-Chlorotoluene
Concen: 54.744 ug/l
RT: 12.855 min Scan# 1810
Delta R.T. 0.006 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40


Tgt Ion: 91 Resp: 211628
Ion Ratio Lower Upper
91 100
126 33.9 18.0 54.0





#83
tert-Butylbenzene
Concen: 56.161 ug/l
RT: 13.075 min Scan# 1846
Delta R.T. 0.000 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40




Instrument :
MSVOA_Y
ClientSampleId :
ICVVY081222




Tgt Ion:119 Resp: 20995
Ion Ratio Lower Upper
119 100
91 65.8 32.0 96.0
134 26.2 13.2 39.6

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 08/15/2022
Supervised By :Mahesh Dadoda 08/16/2022




#84
1,2,4-Trimethylbenzene
Concen: 55.987 ug/l
RT: 13.117 min Scan# 1853
Delta R.T. 0.000 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40




Tgt Ion:105 Resp: 239114
Ion Ratio Lower Upper
105 100
120 45.3 23.4 70.3





#85
 sec-Butylbenzene
 Concen: 55.421 ug/l
 RT: 13.251 min Scan# 1875
 Delta R.T. 0.000 min
 Lab File: VY009990.D
 Acq: 12 Aug 2022 16:40




Instrument :
 MSVOA_Y
 ClientSampleId :
 ICVVY081222




Tgt Ion:105 Resp: 317424
 Ion Ratio Lower Upper
 105 100
 134 20.3 10.5 31.6

Manual Integrations
 APPROVED



Reviewed By :Krupa Patel 08/15/2022
 Supervised By :Mahesh Dadoda 08/16/2022




#86
 p-Isopropyltoluene
 Concen: 57.094 ug/l
 RT: 13.367 min Scan# 1894
 Delta R.T. 0.000 min
 Lab File: VY009990.D
 Acq: 12 Aug 2022 16:40



Tgt Ion:119 Resp: 258130
 Ion Ratio Lower Upper
 119 100
 134 26.4 13.5 40.4
 91 23.8 11.6 34.8









#87
1,3-Dichlorobenzene
Concen: 52.556 ug/l
RT: 13.367 min Scan# 1893
Delta R.T. 0.006 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40

Instrument :
MSVOA_Y
ClientSampleId :
ICVY081222

Tgt Ion:146 Resp: 130450
Ion Ratio Lower Upper
146 100
111 40.6 19.1 57.3
148 63.4 31.9 95.5




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 08/15/2022
Supervised By :Mahesh Dadoda 08/16/2022



#88
1,4-Dichlorobenzene
Concen: 50.880 ug/l
RT: 13.446 min Scan# 1907
Delta R.T. 0.006 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40

Tgt Ion:146 Resp: 131294
Ion Ratio Lower Upper
146 100
111 39.9 18.9 56.9
148 64.3 31.6 95.0





#89

n-Butylbenzene

Concen: 57.136 ug/l

RT: 13.696 min Scan# 1948

Delta R.T. 0.000 min

Lab File: VY009990.D

Acq: 12 Aug 2022 16:40

Instrument :

MSVOA_Y

ClientSampleId :

ICVY081222





Tgt Ion:	91	Resp:	249410
Ion	Ratio	Lower	Upper
91	100		
92	53.9	26.9	80.7
134	26.0	14.0	41.9

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 08/15/2022

Supervised By :Mahesh Dadoda 08/16/2022



#90

Hexachloroethane

Concen: 52.884 ug/l




RT: 13.959 min Scan# 1991


Delta R.T. 0.000 min

Lab File: VY009990.D

Acq: 12 Aug 2022 16:40

Tgt Ion:	117	Resp:	51273
Ion	Ratio	Lower	Upper
117	100		
201	71.0	38.8	116.3









#91
1,2-Dichlorobenzene
Concen: 53.506 ug/l
RT: 13.739 min Scan# 1955
Delta R.T. 0.000 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40

Instrument :
MSVOA_Y
ClientSampleId :
ICVY081222

Tgt Ion:146 Resp: 11946
Ion Ratio Lower Upper
146 100
111 41.5 19.7 59.1
148 63.6 32.1 96.5




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 08/15/2022
Supervised By :Mahesh Dadoda 08/16/2022



#92
1,2-Dibromo-3-Chloropropane
Concen: 51.429 ug/l
RT: 14.355 min Scan# 2056
Delta R.T. 0.000 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40

Tgt Ion: 75 Resp: 9924
Ion Ratio Lower Upper
75 100
155 87.3 46.4 139.1
157 113.6 57.6 172.7





#93

1,2,4-Trichlorobenzene

Concen: 54.010 ug/l

RT: 15.007 min Scan# 2163

Delta R.T. 0.000 min

Lab File: VY009990.D

Acq: 12 Aug 2022 16:40

Instrument :

MSVOA_Y

ClientSampleId :

ICVVY081222





Tgt Ion:180	Resp: 68614
Ion Ratio	Lower Upper
180 100	
182 95.9	47.3 141.9
145 33.0	15.1 45.3

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 08/15/2022

Supervised By :Mahesh Dadoda 08/16/2022



#94

Hexachlorobutadiene

Concen: 54.234 ug/l




RT: 15.111 min Scan# 2180


Delta R.T. 0.000 min

Lab File: VY009990.D

Acq: 12 Aug 2022 16:40


Tgt Ion:225	Resp: 38908
Ion Ratio	Lower Upper
225 100	
223 62.2	30.8 92.3
227 64.9	31.9 95.5





#95
Naphthalene
Concen: 49.507 ug/l
RT: 15.239 min Scan# 2200
Delta R.T. 0.006 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40




Instrument :
MSVOA_Y
ClientSampleId :
ICVY081222



Tgt Ion:128 Resp: 15156
Ion Ratio Lower Upper
128 100
127 12.9 10.6 15.8
129 10.9 8.8 13.2

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 08/15/2022
Supervised By :Mahesh Dadoda 08/16/2022



#96
1,2,3-Trichlorobenzene
Concen: 53.118 ug/l
RT: 15.422 min Scan# 2231
Delta R.T. 0.000 min
Lab File: VY009990.D
Acq: 12 Aug 2022 16:40

Tgt Ion:180 Resp: 60352
Ion Ratio Lower Upper
180 100
182 94.9 47.0 141.0
145 33.9 15.6 46.7

