

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082422\
 Data File : VY010188.D
 Acq On : 24 Aug 2022 11:00
 Operator : KP/MD
 Sample : VSTDIC010
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :

Quant Time: Aug 25 01:50:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y082422S.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 25 01:46:44 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	171773	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	274921	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	263879	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	140755	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	18327	12.353	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	24.700%#
35) Dibromofluoromethane	7.710	113	18792	11.813	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	23.620%#
50) Toluene-d8	10.179	98	66726	12.126	ug/l	0.00
Spiked Amount	50.000	Range	78 - 125	Recovery	=	24.260%#
62) 4-Bromofluorobenzene	12.483	95	24704	11.667	ug/l	0.00
Spiked Amount	50.000	Range	50 - 146	Recovery	=	23.340%#
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.900	85	12665	9.704	ug/l	96
3) Chloromethane	2.107	50	13422	10.390	ug/l	94
4) Vinyl Chloride	2.241	62	14408	10.203	ug/l	100
5) Bromomethane	2.644	94	11939	10.551	ug/l	98
6) Chloroethane	2.784	64	10709	10.666	ug/l	99
7) Trichlorofluoromethane	3.119	101	15614	9.908	ug/l	97
8) Diethyl Ether	3.522	74	9520	10.235	ug/l	99
9) 1,1,2-Trichlorotrifluo...	3.893	101	16872	10.573	ug/l	96
10) Methyl Iodide	4.082	142	17011	9.647	ug/l	88
11) Tert butyl alcohol	4.942	59	11482	58.263	ug/l #	84
12) 1,1-Dichloroethene	3.863	96	14441	10.106	ug/l	84
13) Acrolein	3.723	56	4336	66.943	ug/l	97
14) Allyl chloride	4.473	41	25611	10.178	ug/l	89
15) Acrylonitrile	5.155	53	26883	53.710	ug/l	98
16) Acetone	3.942	43	21734	52.753	ug/l #	87
17) Carbon Disulfide	4.186	76	34999	9.581	ug/l	98
18) Methyl Acetate	4.473	43	14119	9.999	ug/l	95
19) Methyl tert-butyl Ether	5.216	73	42682	10.119	ug/l	100
20) Methylene Chloride	4.704	84	33543	12.413	ug/l #	85
21) trans-1,2-Dichloroethene	5.210	96	16769	10.126	ug/l	85
22) Diisopropyl ether	6.113	45	59743	10.646	ug/l #	96
23) Vinyl Acetate	6.052	43	171478	51.879	ug/l #	94
24) 1,1-Dichloroethane	6.009	63	34399	10.786	ug/l #	97
25) 2-Butanone	6.984	43	38011	55.438	ug/l	94
26) 2,2-Dichloropropane	6.972	77	33299	11.369	ug/l	99
27) cis-1,2-Dichloroethene	6.978	96	21169	10.612	ug/l	95
28) Bromochloromethane	7.326	49	11917	11.148	ug/l	88
29) Tetrahydrofuran	7.338	42	21411	51.324	ug/l	94
30) Chloroform	7.496	83	35975	10.884	ug/l	98
31) Cyclohexane	7.777	56	26442	10.104	ug/l #	90
32) 1,1,1-Trichloroethane	7.698	97	29658	10.682	ug/l	96
36) 1,1-Dichloropropene	7.911	75	23756	10.015	ug/l	96
37) Ethyl Acetate	7.064	43	15795	10.492	ug/l #	94
38) Carbon Tetrachloride	7.893	117	25956	10.348	ug/l	98
39) Methylcyclohexane	9.179	83	25176	9.522	ug/l	92
40) Benzene	8.155	78	76024	10.493	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082422\
 Data File : VY010188.D
 Acq On : 24 Aug 2022 11:00
 Operator : KP/MD
 Sample : VSTDIC010
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :

Quant Time: Aug 25 01:50:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y082422S.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 25 01:46:44 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.344	41	20688	17.364	ug/l #	17
42) 1,2-Dichloroethane	8.234	62	21674	10.782	ug/l	90
43) Isopropyl Acetate	8.271	43	27394	10.420	ug/l #	91
44) Trichloroethene	8.935	130	20703	10.353	ug/l	95
45) 1,2-Dichloropropane	9.210	63	20661	10.863	ug/l	98
46) Dibromomethane	9.301	93	11164	10.644	ug/l	98
47) Bromodichloromethane	9.496	83	28007	10.854	ug/l	94
48) Methyl methacrylate	9.289	41	11377	10.040	ug/l #	84
49) 1,4-Dioxane	9.289	88	2540	208.218	ug/l	87
51) 4-Methyl-2-Pentanone	10.069	43	75846	52.681	ug/l	94
52) Toluene	10.240	92	47117	10.366	ug/l	97
53) t-1,3-Dichloropropene	10.465	75	26720	10.425	ug/l	99
54) cis-1,3-Dichloropropene	9.923	75	30851	10.395	ug/l #	87
55) 1,1,2-Trichloroethane	10.642	97	17693	11.136	ug/l	99
56) Ethyl methacrylate	10.508	69	18980	9.732	ug/l #	85
57) 1,3-Dichloropropane	10.789	76	27780	10.764	ug/l	97
58) 2-Chloroethyl Vinyl ether	9.776	63	47366	60.704	ug/l	99
59) 2-Hexanone	10.831	43	49712	51.006	ug/l	93
60) Dibromochloromethane	10.984	129	20374	10.867	ug/l	99
61) 1,2-Dibromoethane	11.087	107	15279	10.434	ug/l	99
64) Tetrachloroethene	10.715	164	19976	10.284	ug/l	94
65) Chlorobenzene	11.514	112	54341	10.582	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.587	131	21038	10.760	ug/l	95
67) Ethyl Benzene	11.593	91	89617	10.119	ug/l	99
68) m/p-Xylenes	11.703	106	69221	20.349	ug/l	94
69) o-Xylene	12.026	106	32367	9.977	ug/l	95
70) Styrene	12.044	104	56822	10.229	ug/l	95
71) Bromoform	12.209	173	13489	10.953	ug/l #	93
73) Isopropylbenzene	12.331	105	86376	9.925	ug/l	98
74) N-amyl acetate	12.142	43	24109	10.156	ug/l #	90
75) 1,1,2,2-Tetrachloroethane	12.581	83	20752	10.923	ug/l	99
76) 1,2,3-Trichloropropane	12.636	75	25528	11.957	ug/l #	100
77) Bromobenzene	12.611	156	21991	10.272	ug/l	97
78) n-propylbenzene	12.672	91	109463	10.117	ug/l	100
79) 2-Chlorotoluene	12.758	91	62491	10.269	ug/l	98
80) 1,3,5-Trimethylbenzene	12.813	105	75541	10.223	ug/l	98
81) trans-1,4-Dichloro-2-b...	12.374	75	6742	10.380	ug/l	85
82) 4-Chlorotoluene	12.855	91	66270	10.423	ug/l	99
83) tert-Butylbenzene	13.075	119	63032	9.919	ug/l	94
84) 1,2,4-Trimethylbenzene	13.117	105	73584	10.064	ug/l	97
85) sec-Butylbenzene	13.251	105	97632	10.089	ug/l	99
86) p-Isopropyltoluene	13.367	119	77700	9.837	ug/l	97
87) 1,3-Dichlorobenzene	13.367	146	45834	10.523	ug/l	99
88) 1,4-Dichlorobenzene	13.447	146	46403	10.575	ug/l	97
89) n-Butylbenzene	13.697	91	74270	9.803	ug/l	99
90) Hexachloroethane	13.959	117	17419	10.499	ug/l	90
91) 1,2-Dichlorobenzene	13.739	146	41652	10.697	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.355	75	3073	11.691	ug/l	79
93) 1,2,4-Trichlorobenzene	15.007	180	21588	11.728	ug/l	99
94) Hexachlorobutadiene	15.111	225	13660	12.303	ug/l	97
95) Naphthalene	15.233	128	39985	11.034	ug/l	98
96) 1,2,3-Trichlorobenzene	15.422	180	19213	11.769	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082422\
Data File : VY010188.D
Acq On : 24 Aug 2022 11:00
Operator : KP/MD
Sample : VSTDIC010
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :

Quant Time: Aug 25 01:50:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y082422S.M
Quant Title : SW846 8260
QLast Update : Thu Aug 25 01:46:44 2022
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082422\
 Data File : VY010188.D
 Acq On : 24 Aug 2022 11:00
 Operator : KP/MD
 Sample : VSTDIC010
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :

Quant Time: Aug 25 01:50:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y082422S.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 25 01:46:44 2022
 Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.783 min Scan# 91
Delta R.T. -0.000 min
Lab File: VY010188.D
Acq: 24 Aug 2022 11:00




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion:168 Resp: 171773
Ion Ratio Lower Upper
168 100
99 53.8 46.9 70.3



#2
Dichlorodifluoromethane
Concen: 9.704 ug/l
RT: 1.900 min Scan# 13
Delta R.T. 0.000 min
Lab File: VY010188.D
Acq: 24 Aug 2022 11:00

Tgt Ion: 85 Resp: 12665
Ion Ratio Lower Upper
85 100
87 34.2 16.1 48.2









#3
Chloromethane
Concen: 10.390 ug/l
RT: 2.107 min Scan# 41
Delta R.T. 0.000 min
Lab File: VY010188.D
Acq: 24 Aug 2022 11:00




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 50 Resp: 13422
Ion Ratio Lower Upper
50 100
52 34.1 24.6 37.0



#4
Vinyl Chloride
Concen: 10.203 ug/l
RT: 2.241 min Scan# 69
Delta R.T. 0.000 min
Lab File: VY010188.D
Acq: 24 Aug 2022 11:00


Tgt Ion: 62 Resp: 14408
Ion Ratio Lower Upper
62 100
64 31.1 25.0 37.6








#5
Bromomethane
Concen: 10.551 ug/l
RT: 2.644 min Scan# 11
Delta R.T. 0.006 min
Lab File: VY010188.D
Acq: 24 Aug 2022 11:00

Instrument :
MSVOA_Y
ClientSampleId :







Tgt Ion: 94 Resp: 11939
Ion Ratio Lower Upper
94 100
96 90.4 73.5 110.3



#6
Chloroethane
Concen: 10.666 ug/l
RT: 2.784 min Scan# 158
Delta R.T. 0.000 min
Lab File: VY010188.D
Acq: 24 Aug 2022 11:00

Tgt Ion: 64 Resp: 10709
Ion Ratio Lower Upper
64 100
66 31.0 25.4 38.0









#7
 Trichlorofluoromethane
 Concen: 9.908 ug/l
 RT: 3.119 min Scan# 211
 Delta R.T. 0.012 min
 Lab File: VY010188.D
 Acq: 24 Aug 2022 11:00




Instrument :
 MSVOA_Y
 ClientSampleId :


Tgt Ion:101 Resp: 15614
 Ion Ratio Lower Upper
 101 100
 103 57.4 48.0 72.0



#8
 Diethyl Ether
 Concen: 10.235 ug/l
 RT: 3.522 min Scan# 279
 Delta R.T. 0.000 min
 Lab File: VY010188.D
 Acq: 24 Aug 2022 11:00

Tgt Ion: 74 Resp: 9520
 Ion Ratio Lower Upper
 74 100
 45 111.5 55.1 165.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 10.573 ug/l
RT: 3.893 min Scan# 341
Delta R.T. 0.006 min
Lab File: VY010188.D
Acq: 24 Aug 2022 11:00




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion:101 Resp: 16872
Ion Ratio Lower Upper
101 100
85 44.3 40.6 60.8
151 80.4 64.9 97.3



#10
Methyl Iodide
Concen: 9.647 ug/l
RT: 4.082 min Scan# 371
Delta R.T. 0.006 min
Lab File: VY010188.D
Acq: 24 Aug 2022 11:00

Tgt Ion:142 Resp: 17011
Ion Ratio Lower Upper
142 100
127 41.8 41.5 62.3
141 14.1 11.9 17.9





#11

Tert butyl alcohol

Concen: 58.263 ug/l

RT: 4.942 min Scan# 513

Delta R.T. -0.006 min


Lab File: VY010188.D

Acq: 24 Aug 2022 11:00

Instrument :

MSVOA_Y

ClientSampleId :






Tgt Ion: 59 Resp: 11482

Ion Ratio Lower Upper

59 100

57 2.2 6.4 9.6#



#12

1,1-Dichloroethene

Concen: 10.106 ug/l

RT: 3.863 min Scan# 335

Delta R.T. 0.000 min

Lab File: VY010188.D

Acq: 24 Aug 2022 11:00




Tgt Ion: 96 Resp: 14441


Ion Ratio Lower Upper

96 100

61 151.5 145.8 218.6

98 64.6 52.5 78.7





#13

Acrolein

Concen: 66.943 ug/l

RT: 3.723 min Scan# 312

Delta R.T. -0.000 min

Lab File: VY010188.D

Acq: 24 Aug 2022 11:00

Instrument :

MSVOA_Y





ClientSampleId :

Tgt Ion: 56 Resp: 4336

Ion Ratio Lower Upper

56 100

55 74.6 57.6 86.4



#14

Allyl chloride

Concen: 10.178 ug/l

RT: 4.473 min Scan# 435

Delta R.T. 0.006 min

Lab File: VY010188.D

Acq: 24 Aug 2022 11:00




Tgt Ion: 41 Resp: 25611


Ion Ratio Lower Upper

41 100

39 62.6 58.8 88.2


76 35.0 25.1 37.7








#15
Acrylonitrile
Concen: 53.710 ug/l
RT: 5.155 min Scan# 54
Delta R.T. 0.006 min
Lab File: VY010188.D
Acq: 24 Aug 2022 11:00

Instrument :
MSVOA_Y
ClientSampleId :







Tgt Ion: 53 Resp: 26883
Ion Ratio Lower Upper
53 100
52 81.3 67.0 100.4
51 36.6 29.8 44.6



#16
Acetone
Concen: 52.753 ug/l
RT: 3.942 min Scan# 348
Delta R.T. 0.006 min
Lab File: VY010188.D
Acq: 24 Aug 2022 11:00

Tgt Ion: 43 Resp: 21734
Ion Ratio Lower Upper
43 100
58 34.4 22.0 33.0#









#17
Carbon Disulfide
Concen: 9.581 ug/l
RT: 4.186 min Scan# 387
Delta R.T. 0.006 min
Lab File: VY010188.D
Acq: 24 Aug 2022 11:00




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 76 Resp: 34999
Ion Ratio Lower Upper
76 100
78 9.9 7.4 11.0



#18
Methyl Acetate
Concen: 9.999 ug/l
RT: 4.473 min Scan# 435
Delta R.T. 0.006 min
Lab File: VY010188.D
Acq: 24 Aug 2022 11:00

Tgt Ion: 43 Resp: 14119
Ion Ratio Lower Upper
43 100
74 22.9 16.6 24.8





#19

Methyl tert-butyl Ether

Concen: 10.119 ug/l

RT: 5.216 min Scan# 51

Delta R.T. 0.012 min

Lab File: VY010188.D

Acq: 24 Aug 2022 11:00

Instrument :

MSVOA_Y





ClientSampleId :

Tgt Ion: 73 Resp: 42682

Ion Ratio Lower Upper

73 100

57 23.1 18.6 27.8



#20

Methylene Chloride

Concen: 12.413 ug/l

RT: 4.704 min Scan# 473

Delta R.T. 0.000 min

Lab File: VY010188.D

Acq: 24 Aug 2022 11:00

Tgt Ion: 84 Resp: 33543




Ion Ratio Lower Upper


84 100

49 120.9 114.4 171.6

51 34.4 35.3 52.9#


86 60.1 53.7 80.5








#21
trans-1,2-Dichloroethene
Concen: 10.126 ug/l
RT: 5.210 min Scan# 51
Delta R.T. 0.006 min
Lab File: VY010188.D
Acq: 24 Aug 2022 11:00

Instrument :
MSVOA_Y
ClientSampleId :







Tgt Ion: 96 Resp: 16769
Ion Ratio Lower Upper
96 100
61 125.6 120.5 180.7
98 59.9 49.8 74.8



#22
Diisopropyl ether
Concen: 10.646 ug/l
RT: 6.113 min Scan# 704
Delta R.T. 0.006 min
Lab File: VY010188.D
Acq: 24 Aug 2022 11:00

Tgt Ion: 45 Resp: 59743
Ion Ratio Lower Upper
45 100
43 60.2 49.3 73.9
87 28.8 19.2 28.8#
59 12.3 8.4 12.6









#23
Vinyl Acetate
Concen: 51.879 ug/l
RT: 6.052 min Scan# 693
Delta R.T. 0.007 min
Lab File: VY010188.D
Acq: 24 Aug 2022 11:00




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 43 Resp: 171478
Ion Ratio Lower Upper
43 100
86 11.0 7.1 10.7#



#24
1,1-Dichloroethane
Concen: 10.786 ug/l
RT: 6.009 min Scan# 687
Delta R.T. 0.000 min
Lab File: VY010188.D
Acq: 24 Aug 2022 11:00

Tgt Ion: 63 Resp: 34399
Ion Ratio Lower Upper
63 100
98 6.5 3.6 10.8
100 5.9 2.0 5.8#









#25
2-Butanone
Concen: 55.438 ug/l
RT: 6.984 min Scan# 846
Delta R.T. 0.006 min
Lab File: VY010188.D
Acq: 24 Aug 2022 11:00




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 43 Resp: 38011
Ion Ratio Lower Upper
43 100
72 24.9 17.7 26.5



#26
2,2-Dichloropropane
Concen: 11.369 ug/l
RT: 6.972 min Scan# 845
Delta R.T. 0.006 min
Lab File: VY010188.D
Acq: 24 Aug 2022 11:00

Tgt Ion: 77 Resp: 33299
Ion Ratio Lower Upper
77 100
97 20.4 10.4 31.4





#27

cis-1,2-Dichloroethene

Concen: 10.612 ug/l

RT: 6.978 min Scan# 845

Delta R.T. 0.006 min


Lab File: VY010188.D

Acq: 24 Aug 2022 11:00

Instrument :

MSVOA_Y

ClientSampleId :






Tgt Ion: 96 Resp: 21169

Ion Ratio Lower Upper

96 100

61 150.8 0.0 319.8

98 63.0 0.0 127.0



#28

Bromochloromethane

Concen: 11.148 ug/l

RT: 7.326 min Scan# 903

Delta R.T. 0.000 min

Lab File: VY010188.D

Acq: 24 Aug 2022 11:00




Tgt Ion: 49 Resp: 11917


Ion Ratio Lower Upper

49 100

129 1.5 0.0 2.2

130 78.8 55.4 83.0





#29

Tetrahydrofuran

Concen: 51.324 ug/l

RT: 7.338 min Scan# 905

Delta R.T. -0.000 min


Lab File: VY010188.D

Acq: 24 Aug 2022 11:00




Instrument :

MSVOA_Y

ClientSampleId :



Tgt Ion:	42	Resp:	21411
Ion	Ratio	Lower	Upper
42	100		
72	42.1	30.3	45.5
71	39.6	29.0	43.4



#30

Chloroform

Concen: 10.884 ug/l




RT: 7.496 min Scan# 931


Delta R.T. 0.000 min

Lab File: VY010188.D

Acq: 24 Aug 2022 11:00

Tgt Ion:	83	Resp:	35975
Ion	Ratio	Lower	Upper
83	100		
85	66.3	51.8	77.6









#31
Cyclohexane
Concen: 10.104 ug/l
RT: 7.777 min Scan# 91
Delta R.T. 0.006 min
Lab File: VY010188.D
Acq: 24 Aug 2022 11:00




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 56 Resp: 26442
Ion Ratio Lower Upper
56 100
69 43.3 26.7 40.1#
84 90.2 67.2 100.8



#32
1,1,1-Trichloroethane
Concen: 10.682 ug/l
RT: 7.698 min Scan# 964
Delta R.T. 0.006 min
Lab File: VY010188.D
Acq: 24 Aug 2022 11:00

Tgt Ion: 97 Resp: 29658
Ion Ratio Lower Upper
97 100
99 65.6 50.7 76.1
61 42.5 37.6 56.4





#33

1,2-Dichloroethane-d4

Concen: 12.353 ug/l

RT: 8.137 min Scan# 1035

Delta R.T. 0.006 min

Lab File: VY010188.D

Acq: 24 Aug 2022 11:00

Instrument :

MSVOA_Y





ClientSampleId :

Tgt Ion: 65 Resp: 18327

Ion Ratio Lower Upper

65 100

67 57.2 0.0 99.4



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.685 min Scan# 1126

Delta R.T. 0.000 min

Lab File: VY010188.D

Acq: 24 Aug 2022 11:00




Tgt Ion:114 Resp: 274921


Ion Ratio Lower Upper

114 100

63 19.3 0.0 43.2

88 14.7 0.0 31.8









#35
Dibromofluoromethane
Concen: 11.813 ug/l
RT: 7.710 min Scan# 965
Delta R.T. 0.006 min
Lab File: VY010188.D
Acq: 24 Aug 2022 11:00




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion:113 Resp: 18792
Ion Ratio Lower Upper
113 100
111 103.9 82.6 124.0
192 21.1 17.6 26.4



#36
1,1-Dichloropropene
Concen: 10.015 ug/l
RT: 7.911 min Scan# 999
Delta R.T. 0.000 min
Lab File: VY010188.D
Acq: 24 Aug 2022 11:00

Tgt Ion: 75 Resp: 23756
Ion Ratio Lower Upper
75 100
110 36.7 17.1 51.2
77 32.8 24.4 36.6





#37

Ethyl Acetate

Concen: 10.492 ug/l

RT: 7.064 min Scan# 860

Delta R.T. 0.000 min


Lab File: VY010188.D

Acq: 24 Aug 2022 11:00




Instrument :

MSVOA_Y

ClientSampleId :



Tgt Ion:	43	Resp:	15795
Ion	Ratio	Lower	Upper
43	100		
61	14.9	9.7	14.5#
70	10.5	7.1	10.7



#38

Carbon Tetrachloride

Concen: 10.348 ug/l




RT: 7.893 min Scan# 996


Delta R.T. 0.000 min

Lab File: VY010188.D

Acq: 24 Aug 2022 11:00

Tgt Ion:	117	Resp:	25956
Ion	Ratio	Lower	Upper
117	100		
119	99.9	77.9	116.9
121	29.9	24.8	37.2





#39

Methylcyclohexane

Concen: 9.522 ug/l

RT: 9.179 min Scan# 11

Delta R.T. 0.000 min

Lab File: VY010188.D

Acq: 24 Aug 2022 11:00

Instrument :

MSVOA_Y

ClientSampleId :





Tgt Ion: 83 Resp: 25176

Ion Ratio Lower Upper

83 100

55 74.6 66.5 99.7

98 44.2 37.9 56.9



#40

Benzene

Concen: 10.493 ug/l

RT: 8.155 min Scan# 1039

Delta R.T. 0.006 min

Lab File: VY010188.D




Acq: 24 Aug 2022 11:00


Tgt Ion: 78 Resp: 76024

Ion Ratio Lower Upper

78 100

77 25.7 20.4 30.6









#41
Methacrylonitrile
Concen: 17.364 ug/l
RT: 7.344 min Scan# 90
Delta R.T. 0.043 min
Lab File: VY010188.D
Acq: 24 Aug 2022 11:00




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 41 Resp: 20688
Ion Ratio Lower Upper
41 100
39 25.9 92.9 139.3#
67 0.0 0.0 0.0
52 0.0 0.0 0.0



#42
1,2-Dichloroethane
Concen: 10.782 ug/l
RT: 8.234 min Scan# 1052
Delta R.T. 0.006 min
Lab File: VY010188.D
Acq: 24 Aug 2022 11:00

Tgt Ion: 62 Resp: 21674
Ion Ratio Lower Upper
62 100
98 10.8 0.0 14.8









#43
Isopropyl Acetate
Concen: 10.420 ug/l
RT: 8.271 min Scan# 1058
Delta R.T. 0.006 min
Lab File: VY010188.D
Acq: 24 Aug 2022 11:00




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 43 Resp: 27394
Ion Ratio Lower Upper
43 100
61 30.1 20.2 30.2
87 13.8 8.9 13.3#



#44
Trichloroethene
Concen: 10.353 ug/l
RT: 8.935 min Scan# 1167
Delta R.T. 0.000 min
Lab File: VY010188.D
Acq: 24 Aug 2022 11:00

Tgt Ion:130 Resp: 20703
Ion Ratio Lower Upper
130 100
95 99.3 0.0 188.2





#45

1,2-Dichloropropane

Concen: 10.863 ug/l

RT: 9.210 min Scan# 1212

Delta R.T. 0.000 min

Lab File: VY010188.D

Acq: 24 Aug 2022 11:00

Instrument :

MSVOA_Y





ClientSampleId :

Tgt Ion: 63 Resp: 20661

Ion Ratio Lower Upper

63 100

65 30.1 23.0 34.6



#46

Dibromomethane

Concen: 10.644 ug/l

RT: 9.301 min Scan# 1227

Delta R.T. 0.000 min

Lab File: VY010188.D

Acq: 24 Aug 2022 11:00




Tgt Ion: 93 Resp: 11164


Ion Ratio Lower Upper

93 100

95 83.3 64.3 96.5

174 103.0 83.4 125.0





#47

Bromodichloromethane

Concen: 10.854 ug/l

RT: 9.496 min Scan# 1258

Delta R.T. 0.006 min

Lab File: VY010188.D





Acq: 24 Aug 2022 11:00

Instrument :

MSVOA_Y

ClientSampleId :

Tgt Ion: 83	Resp: 28007
Ion Ratio	Lower Upper
83 100	
85 59.5	51.9 77.9
127 8.4	6.6 10.0



#48

Methyl methacrylate

Concen: 10.040 ug/l




RT: 9.289 min Scan# 1225


Delta R.T. 0.000 min

Lab File: VY010188.D

Acq: 24 Aug 2022 11:00

Tgt Ion: 41	Resp: 11377
Ion Ratio	Lower Upper
41 100	
69 92.4	58.2 87.2
39 53.8	47.4 71.0





#49

1,4-Dioxane

Concen: 208.218 ug/l

RT: 9.289 min Scan# 1225

Delta R.T. -0.006 min


Lab File: VY010188.D

Acq: 24 Aug 2022 11:00

Instrument :

MSVOA_Y

ClientSampleId :






Tgt Ion: 88 Resp: 2540

Ion Ratio Lower Upper

88 100

43 33.7 32.8 49.2

58 63.5 59.8 89.6



#50

Toluene-d8

Concen: 12.126 ug/l

RT: 10.179 min Scan# 1371

Delta R.T. 0.000 min

Lab File: VY010188.D




Acq: 24 Aug 2022 11:00


Tgt Ion: 98 Resp: 66726

Ion Ratio Lower Upper

98 100

100 65.7 50.2 75.2









#51
4-Methyl-2-Pentanone
Concen: 52.681 ug/l
RT: 10.069 min Scan# 1353
Delta R.T. 0.000 min
Lab File: VY010188.D
Acq: 24 Aug 2022 11:00




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 43 Resp: 75846
Ion Ratio Lower Upper
43 100
58 38.5 28.1 42.1



#52
Toluene
Concen: 10.366 ug/l
RT: 10.240 min Scan# 1381
Delta R.T. 0.000 min
Lab File: VY010188.D
Acq: 24 Aug 2022 11:00

Tgt Ion: 92 Resp: 47117
Ion Ratio Lower Upper
92 100
91 174.6 142.6 213.8









#53
 t-1,3-Dichloropropene
 Concen: 10.425 ug/l
 RT: 10.465 min Scan# 1417
 Delta R.T. 0.006 min
 Lab File: VY010188.D
 Acq: 24 Aug 2022 11:00




Instrument :
 MSVOA_Y
 ClientSampleId :


Tgt Ion: 75 Resp: 26720
 Ion Ratio Lower Upper
 75 100
 77 31.1 25.4 38.2



#54
 cis-1,3-Dichloropropene
 Concen: 10.395 ug/l
 RT: 9.923 min Scan# 1329
 Delta R.T. 0.000 min
 Lab File: VY010188.D
 Acq: 24 Aug 2022 11:00

Tgt Ion: 75 Resp: 30851
 Ion Ratio Lower Upper
 75 100
 77 31.6 25.6 38.4
 39 45.1 47.6 71.4#









#55
1,1,2-Trichloroethane
Concen: 11.136 ug/l
RT: 10.642 min Scan# 1447
Delta R.T. 0.000 min
Lab File: VY010188.D
Acq: 24 Aug 2022 11:00




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 97	Resp: 17693
Ion Ratio	Lower Upper
97 100	
83 88.3	69.8 104.8
85 55.8	44.3 66.5
99 62.7	49.8 74.6



#56
Ethyl methacrylate
Concen: 9.732 ug/l
RT: 10.508 min Scan# 1425
Delta R.T. 0.000 min
Lab File: VY010188.D
Acq: 24 Aug 2022 11:00

Tgt Ion: 69	Resp: 18980
Ion Ratio	Lower Upper
69 100	
41 68.5	63.4 95.0
39 32.1	35.8 53.8#





#57

1,3-Dichloropropane

Concen: 10.764 ug/l

RT: 10.789 min Scan# 1471

Delta R.T. 0.000 min

Lab File: VY010188.D

Acq: 24 Aug 2022 11:00

Instrument :

MSVOA_Y





ClientSampleId :

Tgt Ion: 76 Resp: 27780

Ion Ratio Lower Upper

76 100

78 33.2 25.0 37.6



#58

2-Chloroethyl Vinyl ether

Concen: 60.704 ug/l

RT: 9.776 min Scan# 1305

Delta R.T. 0.000 min

Lab File: VY010188.D




Acq: 24 Aug 2022 11:00


Tgt Ion: 63 Resp: 47366

Ion Ratio Lower Upper

63 100

106 27.5 21.5 32.3





#59

2-Hexanone

Concen: 51.006 ug/l

RT: 10.831 min Scan# 1478

Delta R.T. 0.000 min

Lab File: VY010188.D

Acq: 24 Aug 2022 11:00

Instrument :

MSVOA_Y





ClientSampleId :

Tgt Ion: 43 Resp: 49712

Ion Ratio Lower Upper

43 100

58 53.8 24.4 73.4



#60

Dibromochloromethane

Concen: 10.867 ug/l

RT: 10.984 min Scan# 1503

Delta R.T. 0.000 min

Lab File: VY010188.D




Acq: 24 Aug 2022 11:00


Tgt Ion: 129 Resp: 20374

Ion Ratio Lower Upper

129 100

127 76.6 38.7 116.1





#61

1,2-Dibromoethane

Concen: 10.434 ug/l

RT: 11.087 min Scan# 1

Delta R.T. 0.000 min

Lab File: VY010188.D

Acq: 24 Aug 2022 11:00

Instrument :

MSVOA_Y





ClientSampleId :

Tgt Ion:107 Resp: 15279

Ion Ratio Lower Upper

107 100

109 94.9 75.4 113.2



#62

4-Bromofluorobenzene

Concen: 11.667 ug/l

RT: 12.483 min Scan# 1749

Delta R.T. 0.000 min

Lab File: VY010188.D

Acq: 24 Aug 2022 11:00




Tgt Ion: 95 Resp: 24704


Ion Ratio Lower Upper

95 100

174 85.7 0.0 177.0

176 82.9 0.0 175.8





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.490 min Scan# 1

Delta R.T. 0.000 min

Lab File: VY010188.D

Acq: 24 Aug 2022 11:00

Instrument :

MSVOA_Y

ClientSampleId :





Tgt Ion:117 Resp: 263879

Ion Ratio Lower Upper

117 100

82 52.1 43.0 64.4

119 32.6 25.4 38.0



#64

Tetrachloroethene

Concen: 10.284 ug/l

RT: 10.715 min Scan# 1459

Delta R.T. 0.000 min

Lab File: VY010188.D

Acq: 24 Aug 2022 11:00

Tgt Ion:164 Resp: 19976




Ion Ratio Lower Upper


164 100

166 121.2 100.4 150.6

129 84.8 71.4 107.0

131 80.2 70.8 106.2





#65

Chlorobenzene

Concen: 10.582 ug/l

RT: 11.514 min Scan# 11

Delta R.T. 0.000 min

Lab File: VY010188.D

Acq: 24 Aug 2022 11:00

Instrument :

MSVOA_Y





ClientSampleId :

Tgt Ion:112 Resp: 54341

Ion Ratio Lower Upper

112 100

114 33.1 25.8 38.6



#66

1,1,1,2-Tetrachloroethane

Concen: 10.760 ug/l

RT: 11.587 min Scan# 1602

Delta R.T. 0.000 min

Lab File: VY010188.D

Acq: 24 Aug 2022 11:00




Tgt Ion:131 Resp: 21038


Ion Ratio Lower Upper

131 100

133 94.6 47.9 143.6

119 69.9 30.8 92.3





#67

Ethyl Benzene

Concen: 10.119 ug/l

RT: 11.593 min Scan# 1

Delta R.T. 0.000 min

Lab File: VY010188.D

Acq: 24 Aug 2022 11:00

Instrument :

MSVOA_Y





ClientSampleId :

Tgt Ion: 91 Resp: 89617

Ion Ratio Lower Upper

91 100

106 31.1 24.5 36.7



#68

m/p-Xylenes

Concen: 20.349 ug/l

RT: 11.703 min Scan# 1621

Delta R.T. 0.006 min

Lab File: VY010188.D




Acq: 24 Aug 2022 11:00


Tgt Ion: 106 Resp: 69221

Ion Ratio Lower Upper

106 100

91 199.0 166.6 249.8





#69

o-Xylene

Concen: 9.977 ug/l

RT: 12.026 min Scan# 1

Delta R.T. 0.000 min

Lab File: VY010188.D

Acq: 24 Aug 2022 11:00

Instrument :

MSVOA_Y





ClientSampleId :

Tgt Ion:106 Resp: 32367

Ion Ratio Lower Upper

106 100

91 213.0 110.3 331.0



#70

Styrene

Concen: 10.229 ug/l

RT: 12.044 min Scan# 1677

Delta R.T. 0.000 min

Lab File: VY010188.D

Acq: 24 Aug 2022 11:00




Tgt Ion:104 Resp: 56822


Ion Ratio Lower Upper

104 100

78 46.3 42.4 63.6

103 54.3 44.4 66.6





#71

Bromoform

Concen: 10.953 ug/l

RT: 12.209 min Scan# 11

Delta R.T. 0.006 min

Lab File: VY010188.D





Acq: 24 Aug 2022 11:00

Instrument :

MSVOA_Y

ClientSampleId :

Tgt Ion:173	Resp:	13489
Ion Ratio	Lower	Upper
173	100	
175	44.4	24.6 74.0
254	0.0	0.0 0.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l




RT: 13.422 min Scan# 1903


Delta R.T. -0.006 min

Lab File: VY010188.D

Acq: 24 Aug 2022 11:00


Tgt Ion:152	Resp:	140755
Ion Ratio	Lower	Upper
152	100	
115	56.1	28.2 84.7
150	160.4	0.0 347.6








#73
Isopropylbenzene
Concen: 9.925 ug/l
RT: 12.331 min Scan# 11
Delta R.T. 0.000 min
Lab File: VY010188.D
Acq: 24 Aug 2022 11:00

Instrument :
MSVOA_Y
ClientSampleId :







Tgt Ion:105 Resp: 86376
Ion Ratio Lower Upper
105 100
120 27.6 13.4 40.2



#74
N-ethyl acetate
Concen: 10.156 ug/l
RT: 12.142 min Scan# 1693
Delta R.T. 0.000 min
Lab File: VY010188.D
Acq: 24 Aug 2022 11:00

Tgt Ion: 43 Resp: 24109
Ion Ratio Lower Upper
43 100
70 43.0 28.7 43.1
55 26.2 17.3 25.9#
61 23.7 17.0 25.6









#75
 1,1,2,2-Tetrachloroethane
 Concen: 10.923 ug/l
 RT: 12.581 min Scan# 1765
 Delta R.T. 0.000 min
 Lab File: VY010188.D
 Acq: 24 Aug 2022 11:00




Instrument :
 MSVOA_Y
 ClientSampleId :


Tgt Ion	Ratio	Lower	Upper
83	100		
131	10.2	5.6	16.8
85	63.7	32.3	96.8



#76
 1,2,3-Trichloropropane
 Concen: 11.957 ug/l
 RT: 12.636 min Scan# 1774
 Delta R.T. 0.006 min
 Lab File: VY010188.D
 Acq: 24 Aug 2022 11:00

Tgt Ion	Ratio	Lower	Upper
75	100		
77	0.0	0.0	0.0









#77
Bromobenzene
Concen: 10.272 ug/l
RT: 12.611 min Scan# 1769
Delta R.T. 0.006 min
Lab File: VY010188.D
Acq: 24 Aug 2022 11:00




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion	Ratio	Lower	Upper
156	100		
77	187.2	95.5	286.4
158	100.0	48.0	144.0



#78
n-propylbenzene
Concen: 10.117 ug/l
RT: 12.672 min Scan# 1780
Delta R.T. 0.000 min
Lab File: VY010188.D
Acq: 24 Aug 2022 11:00

Tgt Ion	Ratio	Lower	Upper
91	100		
120	22.7	11.3	33.9









#79
2-Chlorotoluene
Concen: 10.269 ug/l
RT: 12.758 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VY010188.D
Acq: 24 Aug 2022 11:00




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 91 Resp: 62491
Ion Ratio Lower Upper
91 100
126 34.7 16.9 50.6



#80
1,3,5-Trimethylbenzene
Concen: 10.223 ug/l
RT: 12.813 min Scan# 1803
Delta R.T. 0.000 min
Lab File: VY010188.D
Acq: 24 Aug 2022 11:00

Tgt Ion: 105 Resp: 75541
Ion Ratio Lower Upper
105 100
120 50.2 24.4 73.4









#81
trans-1,4-Dichloro-2-butene
Concen: 10.380 ug/l
RT: 12.374 min Scan# 1731
Delta R.T. -0.006 min
Lab File: VY010188.D
Acq: 24 Aug 2022 11:00




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 75 Resp: 6742
Ion Ratio Lower Upper
75 100
53 87.1 85.3 127.9
89 43.1 37.0 55.4



#82
4-Chlorotoluene
Concen: 10.423 ug/l
RT: 12.855 min Scan# 1810
Delta R.T. 0.000 min
Lab File: VY010188.D
Acq: 24 Aug 2022 11:00

Tgt Ion: 91 Resp: 66270
Ion Ratio Lower Upper
91 100
126 34.5 17.0 51.0





#83

tert-Butylbenzene

Concen: 9.919 ug/l

RT: 13.075 min Scan# 1846

Delta R.T. 0.000 min

Lab File: VY010188.D





Acq: 24 Aug 2022 11:00

Instrument :

MSVOA_Y

ClientSampleId :

Tgt Ion:	119	Resp:	63032
Ion	Ratio	Lower	Upper
119	100		
91	63.0	34.8	104.4
134	26.7	13.3	39.9



#84

1,2,4-Trimethylbenzene

Concen: 10.064 ug/l




RT: 13.117 min Scan# 1853


Delta R.T. 0.000 min

Lab File: VY010188.D

Acq: 24 Aug 2022 11:00

Tgt Ion:	105	Resp:	73584
Ion	Ratio	Lower	Upper
105	100		
120	46.5	22.3	66.9





#85

sec-Butylbenzene

Concen: 10.089 ug/l

RT: 13.251 min Scan# 1875

Delta R.T. 0.000 min

Lab File: VY010188.D

Acq: 24 Aug 2022 11:00

Instrument :

MSVOA_Y





ClientSampleId :

Tgt Ion:105 Resp: 97632

Ion Ratio Lower Upper

105 100

134 19.8 10.2 30.4



#86

p-Isopropyltoluene

Concen: 9.837 ug/l

RT: 13.367 min Scan# 1894

Delta R.T. 0.000 min

Lab File: VY010188.D

Acq: 24 Aug 2022 11:00




Tgt Ion:119 Resp: 77700


Ion Ratio Lower Upper

119 100

134 27.4 13.2 39.5

91 22.6 12.4 37.2









#87
1,3-Dichlorobenzene
Concen: 10.523 ug/l
RT: 13.367 min Scan# 1894
Delta R.T. 0.000 min
Lab File: VY010188.D
Acq: 24 Aug 2022 11:00




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion:146 Resp: 45834
Ion Ratio Lower Upper
146 100
111 37.5 19.5 58.5
148 62.9 31.6 95.0



#88
1,4-Dichlorobenzene
Concen: 10.575 ug/l
RT: 13.447 min Scan# 1907
Delta R.T. 0.000 min
Lab File: VY010188.D
Acq: 24 Aug 2022 11:00

Tgt Ion:146 Resp: 46403
Ion Ratio Lower Upper
146 100
111 41.2 19.3 57.8
148 63.5 32.6 97.8









#89
n-Butylbenzene
Concen: 9.803 ug/l
RT: 13.697 min Scan# 1948
Delta R.T. 0.000 min
Lab File: VY010188.D
Acq: 24 Aug 2022 11:00




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 91 Resp: 74270
Ion Ratio Lower Upper
91 100
92 52.1 26.3 78.9
134 24.7 13.0 38.9



#90
Hexachloroethane
Concen: 10.499 ug/l
RT: 13.959 min Scan# 1991
Delta R.T. 0.000 min
Lab File: VY010188.D
Acq: 24 Aug 2022 11:00

Tgt Ion: 117 Resp: 17419
Ion Ratio Lower Upper
117 100
201 76.4 43.0 128.8









#91
1,2-Dichlorobenzene
Concen: 10.697 ug/l
RT: 13.739 min Scan# 1955
Delta R.T. 0.000 min
Lab File: VY010188.D
Acq: 24 Aug 2022 11:00




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion	Ratio	Lower	Upper
146	100		
111	39.6	20.3	60.9
148	64.5	32.3	96.9



#92
1,2-Dibromo-3-Chloropropane
Concen: 11.691 ug/l
RT: 14.355 min Scan# 2056
Delta R.T. 0.000 min
Lab File: VY010188.D
Acq: 24 Aug 2022 11:00

Tgt Ion	Ratio	Lower	Upper
75	100		
155	95.0	38.0	113.9
157	115.6	47.8	143.3









#93
1,2,4-Trichlorobenzene
Concen: 11.728 ug/l
RT: 15.007 min Scan# 2163
Delta R.T. 0.000 min
Lab File: VY010188.D
Acq: 24 Aug 2022 11:00




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion	Ratio	Lower	Upper
180	100		
182	94.7	47.3	141.9
145	30.8	14.6	43.8



#94
Hexachlorobutadiene
Concen: 12.303 ug/l
RT: 15.111 min Scan# 2180
Delta R.T. 0.000 min
Lab File: VY010188.D
Acq: 24 Aug 2022 11:00

Tgt Ion	Ratio	Lower	Upper
225	100		
223	62.5	30.8	92.3
227	66.9	31.6	94.7





#95

Naphthalene

Concen: 11.034 ug/l

RT: 15.233 min Scan# 21

Delta R.T. -0.006 min


Lab File: VY010188.D

Acq: 24 Aug 2022 11:00




Instrument :

MSVOA_Y

ClientSampleId :



Tgt Ion:	128	Resp:	39985
Ion Ratio	100	Lower	Upper
128	100		
127	13.2	11.0	16.4
129	11.4	8.4	12.6



#96

1,2,3-Trichlorobenzene

Concen: 11.769 ug/l

RT: 15.422 min Scan# 2231

Delta R.T. 0.000 min

Lab File: VY010188.D

Acq: 24 Aug 2022 11:00

Tgt Ion:	180	Resp:	19213
Ion Ratio	100	Lower	Upper
180	100		
182	92.4	47.3	141.9
145	29.7	15.3	45.9

