

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092022\
 Data File : VY010576.D
 Acq On : 20 Sep 2022 15:12
 Operator : KP/MD
 Sample : VIBLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VIBLK

Quant Time: Sep 21 01:20:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091922S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 20 01:02:44 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	204776	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	353064	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	333850	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	153265	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	114064	58.237	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	116.480%
35) Dibromofluoromethane	7.716	113	108257	48.045	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.080%
50) Toluene-d8	10.179	98	415819	49.937	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	99.880%
62) 4-Bromofluorobenzene	12.477	95	146169	47.141	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	94.280%
Target Compounds						
25) 2-Butanone	6.978	43	4386	7.020	ug/l #	87
95) Naphthalene	15.233	128	78248	15.622	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092022\
Data File : VY010576.D
Acq On : 20 Sep 2022 15:12
Operator : KP/MD
Sample : VIBLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Quant Time: Sep 21 01:20:52 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091922S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 20 01:02:44 2022
Response via : Initial Calibration

