

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092022\
 Data File : VY010580.D
 Acq On : 20 Sep 2022 17:10
 Operator : KP/MD
 Sample : N4697-03
 Misc : 5.08g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-23-6.5-7.0

Quant Time: Sep 21 01:21:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091922S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 20 01:02:44 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	187918	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	327129	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	311392	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	144216	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	105385	58.632	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	117.260%
35) Dibromofluoromethane	7.709	113	101481	48.608	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.220%
50) Toluene-d8	10.179	98	389732	50.479	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	100.960%
62) 4-Bromofluorobenzene	12.477	95	133036	46.256	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	92.520%
Target Compounds						
95) Naphthalene	15.232	128	11768	2.497	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092022\
Data File : VY010580.D
Acq On : 20 Sep 2022 17:10
Operator : KP/MD
Sample : N4697-03
Misc : 5.08g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-23-6.5-7.0

Quant Time: Sep 21 01:21:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091922S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 20 01:02:44 2022
Response via : Initial Calibration

