

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092022\
 Data File : VY010589.D
 Acq On : 20 Sep 2022 20:40
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 21 01:22:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091922S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 20 01:02:44 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	75	0.00
2 T	Dichlorodifluoromethane	0.279	0.206	26.2#	72	0.00
3 P	Chloromethane	0.391	0.343	12.3	79	0.00
4 C	Vinyl Chloride	0.490	0.425	13.3#	75	0.00
5 T	Bromomethane	0.350	0.302	13.7	78	0.00
6 T	Chloroethane	0.352	0.323	8.2	78	0.00
7 T	Trichlorofluoromethane	0.657	0.591	10.0	76	0.00
8 T	Diethyl Ether	0.225	0.225	0.0	81	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.395	0.363	8.1	75	0.00
10 T	Methyl Iodide	0.457	0.399	12.7	70	0.00
11 T	Tert butyl alcohol	0.049	0.047	4.1	92	0.00
12 CM	1,1-Dichloroethene	0.353	0.314	11.0#	74	0.00
13 T	Acrolein	0.022	0.018	18.2	70	0.00
14 T	Allyl chloride	0.532	0.502	5.6	78	0.00
15 T	Acrylonitrile	0.117	0.126	-7.7	87	0.00
16 T	Acetone	0.098	0.094	4.1	73	0.00
17 T	Carbon Disulfide	0.788	0.635	19.4	74	0.00
18 T	Methyl Acetate	0.316	0.293	7.3	88	0.00
19 T	Methyl tert-butyl Ether	1.101	1.134	-3.0	82	0.00
20 T	Methylene Chloride	0.785	0.534	32.0#	77	0.00
21 T	trans-1,2-Dichloroethene	0.397	0.363	8.6	76	0.00
22 T	Diisopropyl ether	1.257	1.325	-5.4	84	0.00
23 T	Vinyl Acetate	0.742	0.814	-9.7	84	0.00
24 P	1,1-Dichloroethane	0.754	0.760	-0.8	81	0.00
25 T	2-Butanone	0.153	0.156	-2.0	82	0.00
26 T	2,2-Dichloropropane	0.725	0.654	9.8	73	0.00
27 T	cis-1,2-Dichloroethene	0.495	0.494	0.2	80	0.00
28 T	Bromochloromethane	0.301	0.408	-35.5#	84	0.00
29 T	Tetrahydrofuran	0.094	0.100	-6.4	86	0.00
30 C	Chloroform	0.816	0.842	-3.2#	82	0.00
31 T	Cyclohexane	0.617	0.512	17.0	75	0.00
32 T	1,1,1-Trichloroethane	0.726	0.709	2.3	78	0.00
33 S	1,2-Dichloroethane-d4	0.478	0.546	-14.2	84	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	78	0.00
35 S	Dibromofluoromethane	0.319	0.314	1.6	82	0.00
36 T	1,1-Dichloropropene	0.356	0.325	8.7	77	0.00
37 T	Ethyl Acetate	0.205	0.220	-7.3	88	0.00
38 T	Carbon Tetrachloride	0.409	0.382	6.6	77	0.00
39 T	Methylcyclohexane	0.416	0.358	13.9	75	0.00
40 TM	Benzene	1.060	1.014	4.3	80	0.00
41 T	Methacrylonitrile	0.105	0.108	-2.9	82	0.00
42 TM	1,2-Dichloroethane	0.326	0.331	-1.5	84	0.00
43 T	Isopropyl Acetate	0.376	0.389	-3.5	84	0.00
44 TM	Trichloroethene	0.303	0.283	6.6	78	0.00
45 C	1,2-Dichloropropane	0.274	0.279	-1.8#	84	0.00
46 T	Dibromomethane	0.161	0.165	-2.5	85	0.00
47 T	Bromodichloromethane	0.398	0.411	-3.3	83	0.00
48 T	Methyl methacrylate	0.166	0.174	-4.8	84	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092022\
 Data File : VY010589.D
 Acq On : 20 Sep 2022 20:40
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 21 01:22:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091922S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 20 01:02:44 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	83	0.00
50 S	Toluene-d8	1.087	1.230	-13.2	80	0.00
51 T	4-Methyl-2-Pentanone	0.213	0.227	-6.6	86	0.00
52 CM	Toluene	0.694	0.667	3.9#	80	0.00
53 T	t-1,3-Dichloropropene	0.402	0.404	-0.5	82	0.00
54 T	cis-1,3-Dichloropropene	0.448	0.449	-0.2	81	0.00
55 T	1,1,2-Trichloroethane	0.237	0.249	-5.1	86	0.00
56 T	Ethyl methacrylate	0.307	0.322	-4.9	83	0.00
57 T	1,3-Dichloropropane	0.392	0.403	-2.8	84	0.00
58 T	2-Chloroethyl Vinyl ether	0.155	0.178	-14.8	81	0.00
59 T	2-Hexanone	0.144	0.155	-7.6	85	0.00
60 T	Dibromochloromethane	0.290	0.305	-5.2	84	0.00
61 T	1,2-Dibromoethane	0.222	0.225	-1.4	83	0.00
62 S	4-Bromofluorobenzene	0.474	0.425	10.3	83	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	81	0.00
64 T	Tetrachloroethene	0.334	0.306	8.4	79	0.00
65 PM	Chlorobenzene	0.865	0.829	4.2	82	0.00
66 T	1,1,1,2-Tetrachloroethane	0.337	0.335	0.6	81	0.00
67 C	Ethyl Benzene	1.518	1.429	5.9#	80	0.00
68 T	m/p-Xylenes	0.588	0.558	5.1	81	0.00
69 T	o-Xylene	0.573	0.546	4.7	80	0.00
70 T	Styrene	0.974	0.956	1.8	82	0.00
71 P	Bromoform	0.209	0.214	-2.4	85	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	82	0.00
73 T	Isopropylbenzene	3.090	2.837	8.2	79	0.00
74 T	N-amyl acetate	0.749	0.747	0.3	83	0.00
75 P	1,1,2,2-Tetrachloroethane	0.630	0.617	2.1	85	0.00
76 T	1,2,3-Trichloropropane	0.477	0.451	5.5	88	0.00
77 T	Bromobenzene	0.735	0.695	5.4	82	0.00
78 T	n-propylbenzene	3.681	3.430	6.8	80	0.00
79 T	2-Chlorotoluene	2.097	1.946	7.2	80	0.00
80 T	1,3,5-Trimethylbenzene	2.611	2.438	6.6	81	0.00
81 T	trans-1,4-Dichloro-2-butene	0.215	0.195	9.3	79	0.00
82 T	4-Chlorotoluene	2.166	2.041	5.8	82	0.00
83 T	tert-Butylbenzene	2.335	2.208	5.4	81	0.00
84 T	1,2,4-Trimethylbenzene	2.594	2.455	5.4	82	0.00
85 T	sec-Butylbenzene	3.420	3.197	6.5	80	0.00
86 T	p-Isopropyltoluene	2.858	2.695	5.7	80	0.00
87 T	1,3-Dichlorobenzene	1.492	1.400	6.2	81	0.00
88 T	1,4-Dichlorobenzene	1.477	1.391	5.8	81	0.00
89 T	n-Butylbenzene	2.624	2.476	5.6	80	0.00
90 T	Hexachloroethane	0.549	0.532	3.1	83	0.00
91 T	1,2-Dichlorobenzene	1.329	1.294	2.6	83	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.105	0.103	1.9	85	0.00
93 T	1,2,4-Trichlorobenzene	0.811	0.806	0.6	81	0.00
94 T	Hexachlorobutadiene	0.493	0.464	5.9	78	0.00
95 T	Naphthalene	1.634	1.719	-5.2	85	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092022\
Data File : VY010589.D
Acq On : 20 Sep 2022 20:40
Operator : KP/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Sep 21 01:22:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091922S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 20 01:02:44 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.706	0.706	0.0	81	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6