

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092122\
 Data File : VY010605.D
 Acq On : 21 Sep 2022 17:41
 Operator : KP/MD
 Sample : N4695-02
 Misc : 5.09g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 INCINERATOR-ASH

Quant Time: Sep 22 06:11:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091922S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 20 01:02:44 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	90729	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	164867	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	165129	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	73206	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	47125	54.304	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	108.600%
35) Dibromofluoromethane	7.710	113	42885	40.758	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	81.520%
50) Toluene-d8	10.173	98	159617	41.606	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	83.220%
62) 4-Bromofluorobenzene	12.477	95	50156	33.883	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	67.760%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092122\
Data File : VY010605.D
Acq On : 21 Sep 2022 17:41
Operator : KP/MD
Sample : N4695-02
Misc : 5.09g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
INCINERATOR-ASH

Quant Time: Sep 22 06:11:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091922S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 20 01:02:44 2022
Response via : Initial Calibration

