

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092122\
 Data File : VY010606.D
 Acq On : 21 Sep 2022 18:05
 Operator : KP/MD
 Sample : N4714-01
 Misc : 5.07g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OK-01-091922

Quant Time: Sep 22 06:11:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091922S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 20 01:02:44 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	181510	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	340480	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	344844	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	161429	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	94146	54.228	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	108.460%
35) Dibromofluoromethane	7.710	113	85693	39.436	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	78.880%
50) Toluene-d8	10.179	98	328595	41.484	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	82.960%
62) 4-Bromofluorobenzene	12.477	95	110137	36.208	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	72.420%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092122\
Data File : VY010606.D
Acq On : 21 Sep 2022 18:05
Operator : KP/MD
Sample : N4714-01
Misc : 5.07g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OK-01-091922

Quant Time: Sep 22 06:11:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091922S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 20 01:02:44 2022
Response via : Initial Calibration

