

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092122\
 Data File : VY010607.D
 Acq On : 21 Sep 2022 18:28
 Operator : KP/MD
 Sample : N4713-01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OR-03-091922

Quant Time: Sep 22 06:11:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091922S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 20 01:02:44 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	8608	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	14933	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	14776	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	5777	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	5372	65.247	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	130.500%
35) Dibromofluoromethane	7.704	113	2009	21.080	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	42.160%#
50) Toluene-d8	10.173	98	14045	40.508	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	81.020%
62) 4-Bromofluorobenzene	12.483	95	4234	31.385	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	62.760%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092122\
Data File : VY010607.D
Acq On : 21 Sep 2022 18:28
Operator : KP/MD
Sample : N4713-01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OR-03-091922

Quant Time: Sep 22 06:11:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091922S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 20 01:02:44 2022
Response via : Initial Calibration

