

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092122\
 Data File : VY010605.D
 Acq On : 21 Sep 2022 17:41
 Operator : KP/MD
 Sample : N4695-02
 Misc : 5.09g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 INCINERATOR-ASH

Quant Time: Sep 22 06:11:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091922S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 20 01:02:44 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	90729	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	164867	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	165129	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	73206	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	47125	54.304	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	108.600%
35) Dibromofluoromethane	7.710	113	42885	40.758	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	81.520%
50) Toluene-d8	10.173	98	159617	41.606	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	83.220%
62) 4-Bromofluorobenzene	12.477	95	50156	33.883	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	67.760%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092122\
Data File : VY010605.D
Acq On : 21 Sep 2022 17:41
Operator : KP/MD
Sample : N4695-02
Misc : 5.09g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
INCINERATOR-ASH

Quant Time: Sep 22 06:11:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091922S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 20 01:02:44 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.783 min Scan# 91
Delta R.T. -0.000 min
Lab File: VY010605.D
Acq: 21 Sep 2022 17:41




Instrument :
MSVOA_Y
ClientSampleId :
INCINERATOR-ASH


Tgt Ion:168 Resp: 90729
Ion Ratio Lower Upper
168 100
99 54.3 46.9 70.3



#33
1,2-Dichloroethane-d4
Concen: 54.304 ug/l
RT: 8.136 min Scan# 1036
Delta R.T. -0.001 min
Lab File: VY010605.D
Acq: 21 Sep 2022 17:41

Tgt Ion: 65 Resp: 47125
Ion Ratio Lower Upper
65 100
67 51.0 0.0 99.4





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.685 min Scan# 1126

Delta R.T. 0.000 min

Lab File: VY010605.D

Acq: 21 Sep 2022 17:41

Instrument :

MSVOA_Y

ClientSampleId :

INCINERATOR-ASH





Tgt Ion:114 Resp: 164867

Ion Ratio Lower Upper

114 100

63 18.5 0.0 43.2

88 14.1 0.0 31.8



#35

Dibromofluoromethane

Concen: 40.758 ug/l

RT: 7.710 min Scan# 966

Delta R.T. -0.000 min

Lab File: VY010605.D

Acq: 21 Sep 2022 17:41




Tgt Ion:113 Resp: 42885


Ion Ratio Lower Upper

113 100

111 104.3 82.6 124.0

192 20.1 17.6 26.4









#50
Toluene-d8
Concen: 41.606 ug/l
RT: 10.173 min Scan# 11
Delta R.T. -0.000 min
Lab File: VY010605.D
Acq: 21 Sep 2022 17:41




Instrument :
MSVOA_Y
ClientSampleId :
INCINERATOR-ASH


Tgt Ion: 98 Resp: 159617
Ion Ratio Lower Upper
98 100
100 63.8 50.2 75.2



#62
4-Bromofluorobenzene
Concen: 33.883 ug/l
RT: 12.477 min Scan# 1748
Delta R.T. 0.000 min
Lab File: VY010605.D
Acq: 21 Sep 2022 17:41

Tgt Ion: 95 Resp: 50156
Ion Ratio Lower Upper
95 100
174 87.3 0.0 177.0
176 81.7 0.0 175.8





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.490 min Scan# 11

Delta R.T. -0.000 min

Lab File: VY010605.D

Acq: 21 Sep 2022 17:41

Instrument :

MSVOA_Y

ClientSampleId :

INCINERATOR-ASH





Tgt Ion:117 Resp: 165129

Ion Ratio Lower Upper

117 100

82 52.1 43.0 64.4

119 31.6 25.4 38.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.422 min Scan# 1903

Delta R.T. 0.000 min

Lab File: VY010605.D

Acq: 21 Sep 2022 17:41

Tgt Ion:152 Resp: 73206

Ion Ratio Lower Upper

152 100

115 58.4 28.2 84.7

150 158.5 0.0 347.6

